

**RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL
EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES
DISTRIBUTED IN MARCH 2000**

U.S. GEOLOGICAL SURVEY

Open-File Report 00-398



**RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL
EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES**

DISTRIBUTED IN MARCH 2000

By Jerry W. Farrar and Ashley M. Copen

U.S. GEOLOGICAL SURVEY

Open-File Report 00-398

**Lakewood, Colorado
2000**

DEPARTMENT OF THE INTERIOR

BRUCE BABBITT, Secretary

U.S. GEOLOGICAL SURVEY

Charles G. Groat, Director

For additional information
write to:
Chief, Branch of Quality Systems
U.S. Geological Survey
Water Resources Division
Box 25046, Mail Stop 401
Denver Federal Center
Denver, Colorado 80225-0046

Copies of this report can be
purchased from:
U.S. Geological Survey
Bldg 810, Information Services
DFC, Box 25286
Denver, Colorado 80225-0286
888-ASK-USGS

CONTENTS

	Page
Abstract	1
Introduction	1
Participating laboratories.....	3
Preparation of standard reference samples	6
Laboratory analyses	8
Laboratory performance ratings	10
Statistical presentation of data	10
Reference	11

FIGURE

Figure 1. Statistical parameters shown on reported-data graphs in tables 11 - 16	11
----------------------------------------------------------------------------------------	----

TABLES

Table 1. Laboratory participants in the analyses of standard reference samples distributed in March 2000	3
2. Constituents determined in standard reference samples distributed in March 2000.....	8
3. Analytical method codes	9
4. Overall laboratory performance ratings for standard reference samples distributed in March 2000	12
5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)	14
6. Laboratory performance ratings for standard reference sample M-154 (major constituents)	22
7. Laboratory performance ratings for standard reference sample N-65 (nutrient constituents)	28
8. Laboratory performance ratings for standard reference sample N-66 (nutrient constituents)	30
9. Laboratory performance ratings for standard reference sample P-34 (low ionic strength constituents)	32
10. Laboratory performance ratings for standard reference sample Hg-30 (mercury)	34
11. Statistical summary of reported data for standard reference sample T-161 (trace constituents)	35
12. Statistical summary of reported data for standard reference sample M-154 (major constituents)	64
13. Statistical summary of reported data for standard reference sample N-65 (nutrient constituents)	81
14. Statistical summary of reported data for standard reference sample N-66 (nutrient constituents)	87
15. Statistical summary of reported data for standard reference sample P-34 (low ionic strength constituents)	93
16. Statistical summary of reported data for standard reference sample Hg-30 (mercury)	105
17. Most probable values for constituents and properties in standard reference samples distributed in March 2000	107

RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES DISTRIBUTED IN MARCH 2000

By Jerry W. Farrar and Ashley M. Copen

ABSTRACT

This report presents the results of the U.S. Geological Survey's analytical evaluation program for six standard reference samples -- T-161 (trace constituents), M-154 (major constituents), N-65 (nutrient constituents), N-66 (nutrient constituents), P-34 (low ionic strength constituents), and Hg-30 (mercury) -- that were distributed in March 2000 to 144 laboratories enrolled in the U.S. Geological Survey sponsored interlaboratory testing program. Analytical data that were received from 132 of the laboratories were evaluated with respect to overall laboratory performance and relative laboratory performance for each analyte in the six reference samples. Results of these evaluations are presented in tabular form. Also presented are tables and graphs summarizing the analytical data provided by each laboratory for each analyte in the six standard reference samples. The most probable value for each analyte was determined using nonparametric statistics.

INTRODUCTION

The U.S. Geological Survey (USGS) conducts an interlaboratory analytical evaluation program semiannually. This program provides a variety of standard reference samples (SRSs) to accomplish quality assurance testing of laboratories and to provide an adequate supply of samples that contribute to quality control programs of participating laboratories. Natural-matrix reference materials are preferred for use in this interlaboratory evaluation program. A series of samples are prepared and distributed each spring and fall.

The program began in 1962 with a single sample containing major constituents that was prepared from distilled water and reagent grade chemicals. Twenty-three USGS laboratories participated in the first analytical evaluation program. Since that time, objectives of the program have been to:

- (1) evaluate and improve the performance of USGS and other participating laboratories;
- (2) provide a library of carefully prepared, homogeneous, stable reference materials for use in the quality control programs of laboratories;
- (3) identify analytical problem areas;
- (4) identify quality assurance needs with respect to environmental analyses and develop new reference materials to meet these needs; and
- (5) evaluate the accuracy and precision of analytical methods.

A total of 190 USGS and non-USGS laboratories are enrolled in the program, which can currently provide 9 different types of SRSs:

1. Trace constituents.
2. Major constituents.
3. Nutrient constituents.
4. Low ionic-strength constituents.
5. Mercury.
6. Whole water (water with suspended sediment).
7. Acid mine water constituents.
8. Ground-water trace constituents.
9. Ground-water major constituents.

Though this is not a laboratory certification program, participation in this continuing quality assurance program is mandatory for all laboratories providing water-quality data for USGS sponsored reports or storage in the USGS national databases. The results from this study can be used to alert participating laboratories of possible deficiencies in their analytical operations and provide reference materials for laboratory quality-control programs. Participating laboratories are identified only by a confidential laboratory code number.

A library of SRSs, from previous evaluations, is available. USGS offices and participating laboratories can purchase these SRSs for further testing, continuing quality assurance, and quality-control programs by contacting:

U.S. Geological Survey
Branch of Quality Systems
Denver Federal Center, Bldg. 53
P. O. Box 25046 MS 401
Denver, Colorado 80225-0046
(303) 236-1871

This report summarizes the analytical results submitted by 132 of the 144 laboratories that requested and were shipped SRSs for the April 2000 evaluation (table 1). Not all SRSs are requested or necessarily analyzed by all the laboratories; nor do all laboratories enrolled in the program participate in each evaluation. Analytical results for the following, which were mailed the week of March 27, 2000, are presented in this report.

T-161	Trace constituents	N-66	Nutrient constituents
M-154	Major constituents	P-34	Low ionic strength constituents
N-65	Nutrient constituents	Hg-30	Mercury

The USGS requested that analytical results be returned by June 5, 2000 for evaluation and preparation of this report. Laboratories that are providing analytical services to USGS offices are requested to analyze the appropriate SRSs for the same analytes requested by the USGS offices. All laboratories are requested to include the analytical methods used to determine the concentration of each analyte. When analytical method information was provided, it has been included in tables 11 - 16.

Table 1. Laboratory participants in the analyses of standard reference samples distributed in March 2000.

State	City	Participating Laboratory
Alabama	Tuscaloosa	Geological Survey of Alabama
Arkansas	Arkadelphia	Ouachita Baptist University, Department of Biology
	Fayetteville	University of Arkansas, Arkansas Water Resources Center
California	Little Rock	Arkansas Department of Pollution Control and Ecology
	Davis	University of California, Davis, Division of Environmental Studies
	La Verne	Metropolitan Water District, Water Quality Laboratory
	Martinez	Central Contra Costa Sanitary District
	Menlo Park	U.S. Geological Survey
	Oakland	East Bay Municipal Utility District
	San Diego	U.S. Geological Survey, Water Resources Division
	Santa Fe Springs	West Coast Analytical Service, Inc.
	Tahoe City	High Sierra Water Lab
	West Sacramento	California Department of Water Resources
Colorado	Alamosa	Bureau of Reclamation
	Arvada	Severn Trent Laboratories
	Aurora	Severn Trent Laboratories
	Boulder	U.S. Geological Survey, Surface Water Quality Research
	Boulder	U.S. Geological Survey, Webb Project
	Colorado Springs	City of Colorado Springs
	Denver	Denver Water Department
	Denver	Metro Wastewater Reclamation
	Denver	U.S. Geological Survey/WRD Acid Rain Global Climate
	Denver	U.S. Geological Survey, Colorado District Toxic Project
	Denver	U.S. Geological Survey, National Water Quality Laboratory
	Fort Collins	City of Fort Collins - Water Quality Laboratory
	Fort Collins	U.S. Department of Agriculture, Forest Service
	Fort Collins	CSU-Soil, Water, & Plant Testing Laboratory
	Grand Junction	Enviro-Chem Analytical, Inc.
	Greeley	Central Colorado Water Conservatory District
	Loveland	Northern Colorado Water Conservation District
	Northglenn	Northglenn Water Treatment Plant
Pueblo	City of Pueblo Waste Water Treatment Plant	
Westminster	City of Westminster, Semper Water Treatment Plant	
Delaware	Dover	Delaware Department of Natural Resources
Florida	Bradenton	Manatee County Environmental Management
	Brooksville	Southwest Florida Water Management District
	Ocala	U.S. Geological Survey, Water Resources Division, Quality Water Service Unit
	Orlando	Post, Buckley, Schuh, and Jernigan, Inc.
	Tallahassee	City of Tallahassee, Water Quality Division
	Tallahassee	Florida Department of Environmental Protection
	Tallahassee	Savannah Laboratories and Environmental Services
	Tampa	Hillsborough County Environmental Protection Commission
West Palm Beach	South Florida Water Management District	
Georgia	Athens	Micro-Macro International
	Athens	University of Georgia, Soil Testing and Plant Laboratory
	Decatur	Dekalb County WQC Laboratory
	Marietta	Cobb County Water System
Hawaii	Honolulu	University of Hawaii, Department of Oceanography
Idaho	Boise	U.S. Bureau of Reclamation
	Boise	Boise City Water Quality Laboratory
	Pocatello	Idaho State University Chemistry Department
Iowa	Des Moines	University of Iowa Hygienic Laboratory, Des Moines Branch

Table 1. Laboratory participants in the analyses of standard reference samples distributed in March 2000
 --continued

<u>State</u>	<u>City</u>	<u>Participating Laboratory</u>
Kansas	Lawrence	Kansas Geological Survey
	Wichita	City of Wichita, Water and Sewer Department
Kentucky	Frankfort	Division of Environmental Studies
	Lexington	Kentucky Geological Survey
Maine	Orono	University of Maine, Environmental Chemistry Laboratory
Maryland	Baltimore	Maryland Department of Health and Mental Hygiene
	Cambridge	University of Maryland, Horn Point Laboratory
	Solomons	University of Maryland, Chesapeake Biology Laboratory
Michigan	Detroit	Detroit Water and Sewerage Department, Analytical Laboratory
Minnesota	Minneapolis	University of Minnesota, Department of Geology and Geophysics
	St. Paul	Metropolitan Council Environmental Services
	St. Paul	University of Minnesota, Research Analytical Laboratory
Missouri	Jefferson City	Missouri Department of Health
Montana	Butte	Montana Bureau of Mines & Geology
	Helena	Department of Health and Environmental Services
	Jefferson City	Montana Tunnels Laboratory
Nebraska	McCook	Olsen Laboratory
Nevada	Las Vegas	University of Nevada, Las Vegas
	Reno	Desert Research Institute
	Reno	Truckee Meadows Water Reclamation
New Jersey	Trenton	New Jersey Department of Health
New York	Bolton Landing	Darrin Freshwater Institute
	Brewster	New York City Department of Environmental Protection, Brewster Laboratory
	Buffalo	Erie County Public Health Laboratory
	Grahamsville	New York City Department of Environmental Protection, Grahamsville Laboratory
	Hauppauge	Suffolk County Water Authority
	Hempstead	Nassau County Department of Health
	North Babylon	EcoTest Laboratories
	Rochester	Monroe County Department of Health
	Rochester	Columbia Analytical Services
	Shokan	New York City Department of Environmental Protection, Ben Nessin Laboratory
	Syracuse	Onondaga County DDS
	Syracuse	State University of New York, College of Environmental Science and Forestry
	Troy	U.S. Geological Survey, Water Resources Division
	Troy	Rensselaer Polytechnic Institute
	Valhalla	New York City Department of Environmental Protection, Kensico Laboratory
Wantagh	Cedar Creek Project Laboratory	
Yorktown	New York City Department of Environmental Protection, Croton Guttenhouse Laboratory	
North Carolina	Charlotte	Mecklenburg County
North Dakota	Bismarck	North Dakota Department of Health, East Laboratory
	Bismarck	North Dakota State Water Commission
	Bismarck	U.S. Bureau Of Reclamation
Ohio	Cincinnati	U.S. Environmental Protection Agency, NRMRL
	Cincinnati	U.S. Environmental Protection Agency, NERL
	Tiffin	Heidelberg College
	Wooster	Ohio State University, Ohio Agricultural Research and Developmental Center

Table 1. Laboratory participants in the analyses of standard reference samples distributed in March 2000
 --continued

<u>State</u>	<u>City</u>	<u>Participating Laboratory</u>
Oklahoma	Norman	Oklahoma Geological Survey
Oregon	Corvallis	U.S. Department of Agriculture, CCAL
	Hillsboro	Unified Sewerage Agency of Washington County
	Portland	Department of Environmental Quality
Pennsylvania	Harrisburg	Pennsylvania Department of Environmental Protection
	Mechanicsburg	Chemical Solution LTD
South Carolina	Charleston	College of Charleston-Department of Geology
	Columbia	Columbia Analytical Services
South Dakota	Brookings	Water Resources Institute-South Dakota State University
	Vermillion	University of South Dakota
Texas	Austin	Lower Colorado River Authority
	College Station	Albion Environmental
	Laredo	City of Laredo
Virginia	Chesapeake	City of Chesapeake
	Lorton	Fairfax Co. System Engineering and Monitoring Division
	Norfolk	Old Dominion University-Applied Marine Research
	Richmond	Division of Consolidated Laboratory Service
Vermont	Waterbury	Vermont Agency of Natural Resources
Washington	Richland	Battelle Pacific Northwest
	Seattle	Frontier Geoscience
Wisconsin	Madison	Wisconsin State Lab of Hygiene
	Madison	Madison Department of Public Health
	Middleton	U.S. Geological Survey
Wyoming	Laramie	Wyoming Department of Agriculture
Middle East Laboratories		
<u>Location</u>	<u>Participating Laboratory</u>	
Israel	Dan Sewage Treatment Plant Laboratory	
	Geological Survey of Israel Laboratory	
	Mekeroth Water Company, Ltd	
	Institute for Desert Research	
	Mekeroth Water Company, Rosh-Haayn Lab	
	Public Health Laboratory Beer Shiva	
	Public Health Lab, Tel-Aviv	
	Royal Scientific Society of Jordan, Environmental Research Centre	
Jordan West Bank	College of Science and Technology Al-Quds University	
	Birzeit University	

PREPARATION OF STANDARD REFERENCE SAMPLES

All of the SRSs used in this evaluation were prepared by USGS personnel located in Lakewood, Colorado, and were analyzed for analyte concentrations and physical property values before mailing. A library of these SRSs is maintained, and these SRSs can be purchased by participating laboratories and USGS offices for use in their quality-control programs.

Trace constituents sample T-161 was prepared using water collected from Chicago Creek near Idaho Springs, Colorado. The water was pumped through 0.45-, 0.2-, and 0.1-micrometer (μm) filters, in series, into a 1200-liter (L) polypropylene drum. The water was continuously circulated and passed through a 0.1- μm filter and ultraviolet sterilizer for 24 hours. Following this circulation, the water was acidified to pH 1.5 with nitric acid and chlorinated to 5 parts per million (ppm) free chlorine with sodium hypochlorite. The trace constituent concentrations were adjusted by adding reagent grade chemicals. The sample was circulated an additional 24 hours prior to bottling. During bottling, the sample was pumped through an ultraviolet sterilizer and a 0.1- μm filter. The polypropylene and fluorinated ethylene propylene bottles and caps used were acid leached with 0.2N HNO_3 , deionized-water rinsed, and autoclave sterilized.

Major constituents sample M-154 was prepared using water collected from the Platte River near Fort Lupton, Colorado. The water was pumped through 0.45- 0.2- and 0.1- μm filters, in series, into a 1200-L polypropylene drum. The water was chlorinated to 5-ppm free chlorine with sodium hypochlorite, continuously circulated, and passed through an ultraviolet sterilizer for 24 hours prior to bottling. The major constituent concentrations were adjusted by adding reagent grade chemicals. The sample was circulated an additional 24 hours prior to bottling. The sample is then allowed to sit for 48 hours. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- μm filter. The polypropylene bottles and caps used were acid leached with 0.2N HNO_3 , deionized-water rinsed, and autoclave sterilized.

Nutrient constituents sample N-65 was prepared using deionized water. These samples were prepared the week prior to the mailing for this SRS evaluation. The water was pumped through 0.45-, 0.2-, and 0.1- μm filters, in series, into a 50-L polypropylene drum. The desired nutrient concentrations were obtained by adding reagent-grade chemicals. The sample was continuously circulated through an ultraviolet sterilizer for 24 hours prior to being bottled. The 60-milliliter (mL) glass vials used were new, amber, acid leached with 0.2N HCl , deionized-water rinsed, and autoclave sterilized.

Nutrient constituents sample N-66 was prepared using water collected from the Fall River near Idaho Springs, Colorado. These samples were prepared the week prior to the mailing for this SRS evaluation. The water was pumped through 0.45- 0.2- and 0.1- μm filters, in series, into a 200-L polypropylene drum and continuously circulated and passed through an ultraviolet sterilizer for 24 hours. The desired nutrient concentrations were obtained by adding reagent-grade chemicals. The sample was passed through a 0.1- μm filter while bottling. The 250-mL polyethylene bottles used were new, amber, acid leached with 0.2N HCl , deionized-water rinsed, and autoclave sterilized.

Low ionic strength constituents sample P-34 was prepared in a 400-L polypropylene drum using snowmelt from Genessee Park, Colorado. The water was pumped into the drum through 0.45- 0.2- and 0.1- μm filters in series. The desired phosphate and fluoride concentrations were obtained by adding reagent-grade chemicals. Prior to bottling, the sample was continuously mixed for 24 hours while being circulated through a 0.1- μm filter and an ultraviolet sterilizer. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- μm filter. The 500-mL polypropylene bottles used were acid leached with 0.2N HNO_3 , deionized-water rinsed, and autoclave sterilized.

Mercury sample Hg-30 was prepared using water collected from Chicago Creek near Idaho Springs, Colorado. The sample was prepared in a 200-L polypropylene drum. The river water was pumped into this drum through 0.45-, 0.2-, and 0.1- μ m filters in series. The water was continuously circulated and passed through a 0.1- μ m filter and ultraviolet sterilizer for 48 hours. Nitric acid (5-percent, by volume) and dichromate compound (0.05-percent, by weight) were added to stabilize the sample. The desired mercury concentration was obtained by adding a mercury standard solution. Following an additional 24 hours of circulation, the sample was bottled. The 250-mL glass bottles and tetrafluoroethylene fluorocarbon resin caps used were new, acid leached, and deionized-water rinsed.

LABORATORY ANALYSES

The participating laboratories were asked to determine constituents that are summarized in table 2. The number of analytes varied from 28 in T-161 (trace constituents) to 1 in Hg-30 (mercury).

Table 2. *Analytes determined in standard reference samples distributed in March 2000*

[mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius]

Constituent or property		Units	T-161	M-154	N-65	N-66	P-34	Hg-30
Acidity	Acidity as CaCO ₃	mg/L					X	
Alk	Alkalinity as CaCO ₃	mg/L		X				
Ag	Silver	µg/L	X					
Al	Aluminum	µg/L	X					
As	Arsenic	µg/L	X					
B	Boron	µg/L	X	X			X	
Ba	Barium	µg/L	X					
Be	Beryllium	µg/L	X					
Ca	Calcium	mg/L	X	X			X	
Cd	Cadmium	µg/L	X					
Cl	Chloride	mg/L		X			X	
Co	Cobalt	µg/L	X					
Cr	Chromium	µg/L	X					
Cu	Copper	µg/L	X					
DSRD	Dissolved solids	mg/L		X				
F	Fluoride	mg/L		X			X	
Fe	Iron	µg/L	X					
Hg	Mercury	µg/L						X
K	Potassium	mg/L	X	X			X	
Li	Lithium	µg/L	X					
Mg	Magnesium	mg/L	X	X			X	
Mn	Manganese	µg/L	X					
Mo	Molybdenum	µg/L	X					
Na	Sodium	mg/L	X	X			X	
NH ₃ as N	Ammonia	mg/L				X	X	
NH ₃ +Org N as N	Ammonia + Organic N	mg/L				X	X	
Ni	Nickel	µg/L	X					
NO ₃ as N	Nitrate	mg/L				X	X	
Pb	Lead	µg/L	X					
pH		unit		X			X	
PO ₄ as P	Orthophosphate	mg/L				X	X	X
total P as P	Phosphorus	mg/L		X		X	X	
Sb	Antimony	µg/L	X					
Se	Selenium	µg/L	X					
SiO ₂	Silica	mg/L	X	X				
SO ₄	Sulfate	mg/L		X			X	
Sp Cond	Specific conductance	µS/cm		X			X	
Sr	Strontium	µg/L	X	X				
Tl	Thallium	µg/L	X					
U	Uranium	µg/L	X					
V	Vanadium	µg/L	X	X				
Zn	Zinc	µg/L	X					

Laboratories were requested to identify the method used for each constituent according to analytical method codes in table 3.

Table 3. Analytical method codes

Code	Method
0	Other
1	Atomic absorption: direct, air
2	Atomic absorption: direct, nitrous oxide
3	Atomic absorption: graphite furnace
4	Inductively coupled plasma
5	Direct current plasma
6	Inductively coupled plasma/mass spectrometry
7	Ion chromatography
8	Atomic absorption: cold vapor
9	Atomic fluorescence
10	Atomic absorption: extraction [<i>specify chelating agents</i>]
11	Atomic absorption: hydride [<i>specify reducing agent</i>]
12	Flame emission
20	Titration: colorimetric [<i>specify color reagent</i>]
21	Titration: electrometric [<i>specify reducing or oxidizing agent/color reagent</i>]
22	Colorimetric: [<i>specify reducing or oxidizing agent/color reagent</i>]
40	Ion selective electrode
41	Electrometric [<i>pH and specific conductance</i>]
50	Gravimetric: [<i>specify filtration, evaporation, and so forth</i>]
51	Turbidimetric

Participating laboratories were also asked to identify the method used, such as those references listed next, to further define the methods.

1. American Public Health Association, American Water Works Association, and Water Environment Federation, 1995, Standard methods for the examination of water and wastewater (19th ed.): Washington, D.C., American Public Health Association, variable pagination.
2. American Society for Testing and Materials, 1995, Annual book of ASTM standards: Philadelphia, v. 11.0, and v. 11.02.
3. Kopp, J.F., and McKee, G.F., 1979, Methods for chemical analysis of water and wastes: Cincinnati, U.S. Environmental Protection Agency, EPA 600/4-79-020, rev. 1983, 460 p.
4. Fishman, M.J., and Friedman, L.C., eds., 1989. Methods for determination of inorganic substances in water and fluvial sediments (3rd ed.): U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1, 545 p.
5. Miscellaneous manufacturer's instrument manuals or references.

LABORATORY PERFORMANCE RATINGS

To facilitate laboratory intercomparison, laboratory performance ratings that are based on the analyses reported for each SRS are included in tables 4 - 17 in this report. For each SRS, averages of all the analyte ratings and the number of analyte values reported are given for each participating laboratory. In some cases, laboratory reported values in tables 4 - 17 might have been reformatted in terms of significant figures to meet publication criteria. For example, a reported value of 15 may have been changed to 15.0 or a value of 102.86 may have been changed to 102.9 in these tables. However, the actual reported values by all the laboratories were used to calculate the statistical results and performance ratings presented in the report.

Laboratory determination of each analyte is rated on a scale 4 to 0, based on the absolute Z-value, as listed below:

Rating	Absolute Z-value
4 (Excellent)	0.00 to 0.50
3 (Good)	0.51 to 1.00
2 (Satisfactory)	1.01 to 1.50
1 (Marginal)	1.51 to 2.00
0 (Unsatisfactory)	Greater than 2.00

A weighted laboratory rating greater than 2.0 is considered satisfactory, and ratings less than 2.0 are considered unsatisfactory. Ratings are based on the relative performance of laboratories on specific samples and should be reviewed and evaluated on a case-by-case basis for each laboratory considering such factors as methods used and data needs of specific USGS projects using the laboratory data.

STATISTICAL PRESENTATION OF DATA


Data in this report have been evaluated using nonparametric statistics as described by Hoaglin and others (1983). This statistical approach is a resistant statistic because outliers do not influence the median, as they influence the mean in traditional parametric statistics. Analytical data for each analyte are presented in tabular and graphical forms in tables 11 - 16. Tabulated data for each analyte include the laboratory code number; reported values; analytical method; most probable value (MPV); number of reported analyses, excluding less than values, (N); data range; the F-pseudostandard deviation; and the Z-value. The Z-value is equivalent to the Z-score of traditional statistics. The F-pseudostandard deviation is equivalent to the standard deviation (σ) of traditional statistics when the data has a Gaussian distribution. If an analyte has a sufficient number of analyses by a given method, usually 7, the F-pseudostandard deviation for that analytical method is reported in the block of data listed for each analyte.

The median value calculated from the reported results is considered the MPV. The F-pseudostandard deviation is calculated by dividing the fourth-spread by 1.349; therefore the smaller the F-pseudostandard deviation the more precise the determinations. Laboratories reporting "less than" values are not performance rated unless their reported "less than" values are greater than two F-pseudostandard deviation from the median.

The graphical plot of the reported data is shown in figure 1. The upper and lower boundaries of the graphical plots generally are +3 and -3 F-pseudostandard deviation from the median. Computer-program scaling constraints do not permit these boundaries to always be graphed at exactly these values as shown in the graphical plot. Reported values are grouped by analytical method in ascending order of value.

In some cases, if the F-pseudostandard deviation is less than 5 percent of the MPV, the rating criterion is set to 5 percent of the MPV; as shown in Table 11, page 40.

The term “insufficient data” is included in some of the tables and is used when the number of analyses is less than seven or the calculated F-pseudosigma is greater than the MPV. An estimated MPV may be calculated from the available data for a single analytical method, this estimated concentration is denoted by MPV = Estimated. Estimated values are not used to rate laboratories.



NOTE: vertical scale is the concentration value of the individual analyte in appropriate units (see table 2). Horizontal scale is the laboratory reported values separated by method (different symbols) and plotted by increasing values. Numbers next to each symbol at the bottom of the figure are analytical method codes that are described in table 3. Laboratory-reported results greater than 3 F-pseudosigma from the median are not shown on the graphs.

Figure 1. Statistical parameters shown on reported-data graphs in tables 11 - 16

REFERENCE

Hoaglin, D.C., Mosteller, F., and Tukey, J.W., Eds. 1983, Understanding robust and exploratory data analysis: New York, NY, John Wiley, Inc., p. 38-41.

Table 4. Overall laboratory performance ratings for standard reference samples distributed in March 2000

[SRS, standard reference sample; Lab, laboratory number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for reported values of a sample type; V/64, number of reported values of 65 total possible values from all sample types; V/28, V/16, V/5, V/11 and V/1 are number of reported values possible for T-161, M-154, N-65, N-66, P-34 and Hg-30 respectively; NR, not rated; --, not reported.]

Lab	SRS =		T-161		M-154		N-65		N-66		P-34		Hg-30	
	OWR	V/64	OLR	V/28	OLR	V/16	OLR	V/5	OLR	V/5	OLR	V/11	OLR	V/1
1	3.6	63	3.8	28	3.9	15	3.8	5	3.8	5	2.6	9	3	1
2	2.8	8	--	--	--	--	--	--	--	--	2.8	8	--	--
3	2.3	61	2.8	27	2.8	16	0.0	3	0.0	4	1.9	10	4	1
4	4.0	2	--	--	4.0	2	--	--	--	--	--	--	--	--
5	2.6	48	2.7	24	2.7	15	--	--	--	--	2.3	9	--	--
10	3.3	27	3.0	9	3.7	12	3.4	5	--	--	--	--	2	1
11	2.1	63	2.1	27	1.9	16	0.6	5	3.8	5	2.6	9	2	1
12	1.7	27	1.4	7	2.5	11	3.5	2	0.0	5	1.0	1	2	1
13	2.9	42	2.8	21	3.2	13	1.7	3	2.5	4	--	--	4	1
19	2.9	29	2.7	18	3.4	9	--	--	2.5	2	--	--	--	--
21	3.7	6	2.0	1	--	--	4.0	5	--	--	--	--	--	--
23	2.6	52	3.5	21	2.3	13	1.3	4	1.0	5	2.2	9	--	--
24	2.7	31	2.1	18	3.5	13	--	--	--	--	--	--	--	--
25	2.1	60	1.8	25	2.9	16	1.6	5	1.4	5	2.4	9	--	--
26	3.0	29	3.1	19	3.0	10	--	--	--	--	--	--	--	--
28	0.8	4	--	--	--	--	0.0	2	1.5	2	--	--	--	--
31	4.0	5	--	--	--	--	4.0	5	--	--	--	--	--	--
32	3.6	28	3.6	27	--	--	--	--	--	--	--	--	2	1
38	3.5	27	--	--	3.4	10	3.6	5	3.4	5	3.6	7	--	--
39	2.2	5	--	--	--	--	--	--	2.2	5	--	--	--	--
42	3.1	43	3.2	27	3	12	1.5	2	3.0	2	--	--	--	--
43	3.5	6	--	--	3.5	6	--	--	--	--	--	--	--	--
45	3.5	6	--	--	--	--	4.0	3	3.0	3	--	--	--	--
46	3.4	55	3.2	23	3.4	14	3.4	5	3.8	5	3.9	7	4	1
47	4.0	1	--	--	--	--	--	--	--	--	--	--	4	1
48	2.3	45	2.5	24	1.8	13	--	--	--	--	2.6	8	--	--
50	3.5	46	3.3	27	3.6	13	--	--	3.6	5	--	--	4	1
51	3.2	5	--	--	--	--	3.2	5	--	--	--	--	--	--
53	2.5	4	--	--	--	--	3.5	2	1.5	2	--	--	--	--
54	4.0	1	--	--	--	--	--	--	--	--	--	--	4	1
57	2.8	43	2.8	27	2.8	16	--	--	--	--	--	--	--	--
59	2.4	58	2.3	23	3.4	14	2.2	5	2.4	5	1.5	10	4	1
69	3.0	26	2.6	16	3.7	10	--	--	--	--	--	--	--	--
70	2.6	51	2.5	26	3.1	14	2.0	5	2.4	5	--	--	2	1
72	2.4	10	--	--	--	--	1.4	5	3.4	5	--	--	--	--
76	3.6	22	3.7	12	3.4	9	4.0	1	--	--	--	--	--	--
81	2.6	58	2.1	23	3.2	14	2.6	5	2.4	5	2.8	10	3	1
83	3.5	30	3.5	13	3.4	8	4.0	2	3.0	3	3.8	4	--	--
84	2.6	14	2.0	4	3.6	8	--	--	0.0	2	--	--	--	--
85	2.7	30	--	--	3.3	12	2.3	4	1.5	4	2.8	10	--	--
86	2.9	45	2.7	21	3.3	11	3.5	2	2.3	3	3.6	7	0	1
87	2.5	41	2.2	18	3.1	12	2.4	5	2.2	5	--	--	3	1
89	2.9	59	2.3	23	3.1	14	3.6	5	3.8	5	3.5	11	2	1
90	3.3	6	--	--	--	--	3.7	3	3.0	3	--	--	--	--
93	2.9	38	2.5	16	3.1	11	3.0	1	3.0	1	3.3	9	--	--
96	3.0	30	2.6	14	3.1	7	3.0	3	4.0	5	--	--	2	1
97	3.0	46	2.9	23	3.5	12	2.6	5	3.2	5	--	--	0	1
102	2.2	21	--	--	2.0	11	2.6	5	2.2	5	--	--	--	--
105	3.0	59	3.3	26	3.1	16	2.7	3	2.4	5	2.9	8	0	1
107	3.1	16	2.5	6	3.3	4	--	--	3.8	4	3.5	2	--	--
109	2.7	24	2.2	12	3.2	11	--	--	--	--	--	--	4	1
110	3.3	15	3.0	6	--	--	3.0	1	--	--	3.6	8	--	--
113	3.1	52	3.2	19	3.5	14	3.8	4	3.4	5	1.9	10	--	--
114	1.8	22	1.0	10	2.4	7	2.5	2	2.7	3	--	--	--	--
121	3.3	17	3.1	12	3.8	5	--	--	--	--	--	--	--	--
126	2.6	16	2.5	15	--	--	--	--	4.0	1	--	--	--	--
127	2.7	65	2.5	27	3.2	16	2.0	5	3.0	5	2.9	11	2	1
129	2.6	10	--	--	--	--	2.0	5	3.2	5	--	--	--	--
131	1.9	34	2.0	23	1.7	11	--	--	--	--	--	--	--	--
134	3.7	65	3.9	28	3.5	16	4.0	5	2.8	5	3.7	10	3	1
138	3.5	60	3.7	24	3.4	16	3.6	5	3.4	5	3.2	10	--	--
140	2.3	47	2.5	15	2.5	12	1.8	5	2.2	5	2.0	10	--	--
142	2.4	57	2.5	26	2.3	16	3.0	5	2.8	5	--	4	4	1
143	3.3	15	--	--	2.8	5	3.8	5	3.4	5	2.8	--	--	--
144	2.9	10	2.8	9	--	--	--	--	--	--	--	--	4	1
145	2.9	53	3.1	18	3.1	15	2.6	5	2.2	5	3.0	9	0	1
146	2.9	56	2.6	22	3.4	13	3.0	5	2.6	5	3.3	10	1	1
147	3.6	8	3.6	7	--	--	--	--	--	--	--	--	4	1
149	2.2	19	1.8	10	2.5	8	--	--	--	--	--	--	4	1
151	3.3	26	3.4	20	3.0	6	--	--	--	--	--	--	--	--

Table 4. Overall laboratory performance ratings for standard reference samples distributed in March 2000 -- Continued

[SRS, standard reference sample; Lab, laboratory number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for reported values of a sample type; V/64, number of reported values of 65 total possible values from all sample types; V/28, V/16, V/5, V/5, V/11 and V/1 are number of reported values possible for T-161, M-154, N-65, N-66, P-34 and Hg-30 respectively; NR, not rated; --, not reported.]

Lab	SRS =		T-161		M-154		N-65		N-66		P-34		Hg-30	
	OWR	V/64	OLR	V/28	OLR	V/16	OLR	V/5	OLR	V/5	OLR	V/11	OLR	V/1
154	3.2	47	3.2	21	2.7	15	3.6	5	3.6	5	--	--	4	1
158	2.9	10	--	--	--	--	3.0	5	2.8	5	--	--	--	--
180	2.3	52	1.6	20	1.9	13	3.8	5	3.2	5	2.9	9	--	--
183	2.2	25	1.3	6	3.0	7	2.0	3	2.5	4	2.2	5	--	--
190	2.5	25	1.7	6	2.6	7	2.0	4	3.0	4	3.8	4	--	--
191	2.9	47	2.8	25	3.1	10	2.0	2	2.0	2	3.1	8	--	--
193	3.1	38	2.8	17	3.1	7	3.8	4	3.5	4	3.4	5	3	1
196	3.0	21	3.0	21	--	--	--	--	--	--	--	--	--	--
198	3.8	8	--	--	--	--	3.5	4	4.0	4	--	--	--	--
203	2.8	48	3.0	18	2.8	11	3.2	5	3.6	5	1.9	9	--	--
204	2.8	31	2.8	20	3.0	4	--	--	2.4	5	4.0	2	--	--
205	2.5	4	--	--	--	--	1.5	2	3.5	2	--	--	--	--
208	2.8	6	--	--	2.3	3	--	--	4.0	2	2.0	1	--	--
209	1.8	23	1.8	5	1.0	7	3.5	2	4.0	2	1.4	7	--	--
212	2.2	53	2.3	26	3.1	16	0.8	5	0.0	5	--	--	1	1
213	2.7	10	--	--	3.3	4	3.0	2	2.7	3	--	--	0	1
218	1.0	7	1.0	4	1.0	3	--	--	--	--	--	--	--	--
220	2.5	50	3.7	18	2.7	13	0.0	5	2.6	5	1.1	9	--	--
224	2.2	10	--	--	--	--	2.0	5	2.4	5	--	--	--	--
227	3.3	37	3.7	10	2.8	11	3.8	4	2.8	5	3.7	7	--	--
234	3.2	52	3.6	27	2.9	16	3.3	4	3.0	4	--	--	1	1
236	2.0	41	2.0	25	2.1	16	--	--	--	--	--	--	--	--
237	3.6	5	--	--	--	--	--	--	--	--	3.6	5	--	--
243	3.0	11	--	--	3.3	3	2.7	3	2.3	3	4.0	2	--	--
244	3.2	5	--	--	3.0	3	--	--	--	--	3.5	2	--	--
245	4.0	1	--	--	--	--	--	--	--	--	--	--	4	1
247	1.9	62	3.0	26	3.1	16	2.0	5	1.8	5	1.8	9	1	1
253	2.1	10	--	--	--	--	0.8	5	3.4	5	--	--	--	--
254	3.7	21	3.6	13	3.8	8	--	--	--	--	--	--	--	--
255	3.1	30	3.6	14	1.7	6	3.3	3	3.8	4	2.0	3	--	--
257	2.6	11	--	--	2.6	11	--	--	--	--	--	--	--	--
259	3.6	38	3.6	22	3.6	15	--	--	--	--	--	--	4	1
265	3.2	49	3.2	28	3.3	12	--	--	--	--	2.6	8	4	1
267	3.3	7	--	--	3.3	7	--	--	--	--	--	--	--	--
268	2.9	25	2.7	10	3.4	8	--	--	--	--	2.6	7	--	--
269	3.1	7	--	--	3.1	7	--	--	--	--	--	--	--	--
270	1.7	27	1.9	13	2.0	7	--	--	--	--	1.1	7	--	--
273	2.6	45	2.1	20	2.9	14	--	--	--	--	3.3	11	--	--
277	2.5	36	2.3	15	2.7	11	--	--	--	--	2.6	9	2	1
279	1.1	18	0.5	4	1.0	7	--	--	--	--	1.4	7	--	--
284	1.6	25	1.6	24	--	--	--	--	--	--	--	--	3	1
287	0.0	4	--	--	--	--	0.0	2	0.0	2	--	--	--	--
291	1.0	1	--	--	1.0	1	--	--	--	--	--	--	--	--
292	3.0	20	--	--	3.0	11	2.8	4	3.8	4	--	--	1	1
296	1.5	32	1.7	23	1.2	5	--	--	--	--	0.5	4	--	--
297	2.5	44	2.6	23	2.7	11	1.6	5	3.0	5	--	--	--	--
298	4.0	1	--	--	--	--	--	--	--	--	--	--	4	1
301	2.8	4	--	--	--	--	--	--	--	--	2.8	4	--	--
302	2.6	14	1.8	4	3.0	10	--	--	--	--	--	--	--	--
304	4.0	1	--	--	--	--	--	--	--	--	--	--	4	1
305	2.5	36	2.6	18	2.3	9	1.5	4	3.4	5	--	--	--	--
307	2.2	26	2.0	13	3.4	9	--	--	0.0	3	--	--	0	1
308	2.6	5	--	--	--	--	--	--	3.0	4	1.0	1	--	--
309	2.2	14	1.4	5	2.7	9	--	--	--	--	--	--	--	--
313	2.7	10	--	--	--	--	2.2	5	3.2	5	--	--	--	--
314	1.6	5	--	--	--	--	1.6	5	--	--	--	--	--	--
316	3.6	10	--	--	--	--	4.0	5	3.2	5	--	--	--	--
318	4.0	5	--	--	--	--	4.0	5	--	--	--	--	--	--
319	2.5	10	--	--	--	--	2.2	5	2.8	5	--	--	--	--
320	2.7	10	--	--	--	--	3.4	5	2.0	5	--	--	--	--
321	3.0	4	--	--	--	--	3.0	4	--	--	--	--	--	--
322	4.0	3	--	--	--	--	4.0	1	--	--	4.0	2	--	--
323	2.9	16	--	--	2.9	16	--	--	--	--	--	--	--	--
324	0.8	22	0.5	13	1.2	9	--	--	--	--	--	--	--	--
330	2.7	42	2.9	26	2.3	15	--	--	--	--	--	--	3	1
332	1.6	20	1.5	10	1.8	6	--	--	--	--	1.3	4	--	--

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Ag (Silver)				Al (Aluminum)		As (Arsenic)		B (Boron)		Ba (Barium)		Be (Beryllium)		Ca (Calcium)	
	MPV =	OLR	V/28	RV	mg/L	RV	mg/L	RV	mg/L	RV	mg/L	RV	mg/L	RV	mg/L	
				Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	
1	3.8	28	18.0	3	36.9	3	25.1	3	37.7	4	70.4	4	13.0	4	7.14	4
3	2.8	27	16.0	2	33.1	4	25.7	4	37.5	4	63.5	1	12.2	2	7.02	4
5	2.7	24	15.9	2	< 30	NR	26.4	4	35.4	4	66.9	3	12.4	3	6.73	1
10	3.0	9	--	--	--	--	28.0	2	--	--	--	--	--	--	--	--
11	2.1	27	15.0	1	70.0	0	26.0	4	30.0	1	63.0	0	12.0	1	7.10	4
12	1.4	7	15.8	2	--	--	27.0	3	--	--	--	--	--	--	--	--
13	2.8	21	20.2	0	--	--	24.6	2	--	--	70.3	4	12.6	3	7.61	2
19	2.7	18	19.6	1	43.5	1	25.5	4	--	--	66.5	2	13.1	4	6.94	3
21	2.0	1	--	--	--	--	--	--	--	--	--	--	--	--	--	--
23	3.5	21	16.6	3	37.8	3	25.1	3	--	--	68.2	3	13.2	3	6.96	3
24	2.1	18	--	--	--	--	--	--	58.0	0	76.0	1	--	--	7.10	4
25	1.8	25	16.0	2	< 52	NR	16.0	0	21.0	0	65.0	1	13.0	4	6.07	0
26	3.1	19	15.8	2	--	--	25.5	4	36.8	4	71.4	4	14.0	1	7.56	2
32	3.6	27	17.6	4	34.2	4	26.4	4	35.0	3	69.2	4	12.8	4	7.25	4
42	3.2	27	18.2	3	34.9	4	27.0	3	40.6	3	72.4	3	13.3	3	7.16	4
46	3.2	23	17.9	4	--	--	26.3	4	37.4	4	69.9	4	13.0	4	7.14	4
48	2.5	24	15.4	2	36.0	3	25.4	3	37.0	4	74.0	2	12.0	1	8.85	0
50	3.3	27	17.2	4	31.5	4	26.8	4	40.6	3	75.2	2	13.5	3	7.40	3
57	2.8	27	16.8	4	29.0	3	26.2	4	58.1	0	69.5	4	13.0	4	7.16	4
59	2.3	23	17.3	4	< 50	NR	26.8	4	--	--	72.4	3	14.6	0	6.43	0
69	2.6	16	17.1	4	28.9	3	24.6	2	--	--	86.6	0	11.9	1	6.67	2
70	2.5	26	18.8	2	31.3	4	29.8	0	< 100	NR	72.0	4	14.4	0	7.50	3
76	3.7	12	--	--	--	--	27.5	2	38.2	4	70.7	4	--	--	7.17	4
81	2.1	23	16.0	2	63.0	0	23.0	0	--	--	68.0	3	12.0	1	7.14	4
83	3.5	13	--	--	--	--	--	--	--	--	67.1	3	12.5	3	7.20	4
84	2.0	4	--	--	--	--	--	--	--	--	--	--	--	--	3.80	0
86	2.7	21	18.8	2	--	--	23.5	1	40.5	3	69.8	4	12.8	4	7.36	3
87	2.2	18	17.2	4	--	--	26.4	4	--	--	75.2	2	--	--	5.82	0
89	2.3	23	16.8	4	34.8	4	25.0	3	--	--	50.0	0	14.3	0	6.28	0
93	2.5	16	--	--	--	--	24.8	3	--	--	--	--	12.7	4	7.21	4
96	2.6	14	17.6	4	--	--	24.5	2	--	--	< 100	NR	12.3	2	--	--
97	2.9	23	17.5	4	34.2	4	27.2	3	--	--	71.8	4	12.6	3	7.02	4
105	3.3	26	17.3	4	27.7	3	25.3	3	< 200	NR	72.6	3	12.7	4	7.14	4
107	2.5	6	--	--	--	--	--	--	--	--	--	--	--	--	7.75	1
109	2.2	12	--	--	--	--	25.3	3	--	--	--	--	--	--	7.34	4
110	3.0	6	--	--	38.3	3	--	--	--	--	--	--	--	--	7.10	4
113	3.2	19	17.0	4	31.3	4	29.3	0	--	--	64.5	1	13.0	4	6.82	3
114	1.0	10	23.0	0	--	--	--	--	--	--	--	--	15.5	0	--	--
121	3.1	12	--	--	--	--	--	--	--	--	68.0	3	--	--	7.29	4
126	2.5	15	14.0	0	--	--	26.0	4	--	--	--	--	12.7	4	--	--
127	2.5	27	16.0	2	98.7	0	21.7	0	41.3	3	64.9	1	12.4	3	7.00	4
131	2.0	23	36.4	0	--	--	25.4	3	38.2	4	69.4	4	11.9	1	7.30	4
134	3.9	28	17.2	4	35.0	4	26.5	4	38.0	4	71.9	4	13.0	4	7.23	4
138	3.7	24	16.3	3	28.5	3	24.8	3	33.9	3	69.9	4	12.6	3	7.20	4
140	2.5	15	10.4	0	--	--	--	--	--	--	88.0	0	--	--	7.00	4
142	2.5	26	16.7	4	< 50	NR	27.4	3	< 50	NR	73.2	3	12.6	3	7.96	0
144	2.8	9	19.3	1	--	--	24.3	2	--	--	--	--	12.7	4	--	--
145	3.1	18	--	--	< 179	NR	< 39	NR	40.0	3	70.0	4	13.0	4	7.06	4
146	2.6	22	16.6	3	< 54.7	NR	29.5	0	--	--	74.5	2	13.5	3	7.22	4
147	3.6	7	--	--	36.7	3	26.0	4	--	--	--	--	--	--	--	--
149	1.8	10	--	--	34.0	4	27.0	3	--	--	80.0	0	11.0	0	--	--
151	3.4	20	17.0	4	29.0	3	26.0	4	--	--	70.0	4	13.0	4	--	--
154	3.2	21	18.0	3	--	--	28.2	1	--	--	71.7	4	13.2	3	7.30	4
180	1.6	20	16.0	2	37.3	3	--	--	44.2	2	64.7	1	11.8	1	6.62	1
183	1.3	6	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Ag (Silver)				Al (Aluminum)		As (Arsenic)		B (Boron)		Ba (Barium)		Be (Beryllium)		Ca (Calcium)	
	MPV =	17.3	mg/L		32.4	mg/L	26.1	mg/L	37.5	mg/L	70.4	mg/L	12.9	mg/L	7.17	mg/L
	F-pseudosigma =	1.3			6.3		1.3		4.9		3.3		0.5		0.28	
OLR	V/28	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
190	1.7	6	--	--	--	--	25.6	4	--	--	--	--	--	--	--	--
191	2.8	25	--	--	26.6	3	28.0	2	33.7	3	64.4	1	14.2	0	7.30	4
193	2.8	17	17.5	4	--	--	24.0	1	--	--	67.0	3	13.5	3	6.88	3
196	3.0	21	15.0	1	31.7	4	26.8	4	--	--	70.4	4	13.4	3	--	--
203	3.0	18	16.2	3	25.0	2	26.4	4	--	--	70.0	4	--	--	6.80	2
204	2.8	20	17.5	4	27.1	3	26.2	4	--	--	75.0	2	11.4	0	7.52	3
209	1.8	5	--	--	30.0	4	--	--	--	--	--	--	--	--	6.48	1
212	2.3	26	18.0	3	76.4	0	26.9	3	37.2	4	74.7	2	13.1	4	7.40	3
218	1.0	4	--	--	--	--	--	--	--	--	81.0	0	--	--	--	--
220	3.7	18	17.5	4	29.2	4	24.3	2	--	--	70.4	4	13.2	4	7.33	4
227	3.7	10	--	--	32.6	4	--	--	--	--	--	--	--	--	7.29	4
234	3.6	27	17.4	4	29.4	4	28.4	1	44.8	2	70.2	4	13.1	4	7.44	3
236	2.0	25	19.0	2	40.0	2	40.0	0	36.0	4	69.0	4	13.0	4	7.40	3
247	3.0	26	22.5	0	27.0	3	26.1	4	47.0	1	69.6	4	12.3	2	7.18	4
254	3.6	13	--	--	39.4	2	--	--	--	--	--	--	--	--	7.34	4
255	3.6	14	16.7	4	29.2	4	25.8	4	33.4	3	--	--	--	--	7.51	3
259	3.6	22	17.6	4	31.0	4	27.5	2	39.0	4	73.1	3	--	--	7.05	4
265	3.2	28	18.5	3	28.0	3	27.0	3	34.0	3	70.0	4	12.5	3	7.20	4
268	2.7	10	--	--	--	--	--	--	--	--	75.1	2	--	--	7.23	4
270	1.9	13	--	--	24.7	2	--	--	33.0	3	72.4	3	--	--	8.34	0
273	2.1	20	17.8	4	20.5	1	--	--	33.4	3	66.7	2	--	--	7.30	4
277	2.3	15	11.6	0	--	--	--	--	--	--	68.5	3	--	--	7.48	3
279	0.5	4	--	--	--	--	--	--	--	--	--	--	--	--	5.24	0
284	1.6	24	21.0	0	227.0	0	29.0	0	390.0	0	71.0	4	12.0	1	7.06	4
296	1.7	23	20.2	0	30.1	4	25.8	4	--	--	74.8	2	15.2	0	7.35	4
297	2.6	23	16.4	3	32.1	4	25.3	3	--	--	71.6	4	12.5	3	6.46	1
302	1.8	4	--	--	--	--	--	--	--	--	--	--	--	--	7.08	4
305	2.6	18	17.4	4	--	--	30.0	0	--	--	73.0	3	12.8	4	7.17	4
307	2.0	13	19.1	2	--	--	21.6	0	--	--	--	--	--	--	9.75	0
309	1.4	5	--	--	--	--	--	--	--	--	--	--	--	--	7.20	4
324	0.5	13	33.0	0	90.0	0	--	--	--	--	--	--	--	--	1.56	0
330	2.9	26	17.8	4	26.2	3	26.3	4	32.6	3	72.3	3	13.1	4	4.83	0
332	1.5	10	--	--	50.0	0	--	--	--	--	--	--	--	--	6.64	2

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte =	Cd (Cadmium)		Co (Cobalt)		Cr (Chromium)		Cu (Copper)		Fe (Iron)		K (Potassium)		Li (Lithium)	
MPV =	17.5	µg/L	12.5	mg/L	34.6	mg/L	22.0	mg/L	61.7	mg/L	1.26	mg/L	17.80	mg/L
F-pseudostigma =	0.9		0.6		1.7		1.7		10.2		0.11		1.60	
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	17.7	4	12.5	4	35.4	4	21.9	4	71.2	3	1.24	4	17.3	4
3	16.2	2	12.6	4	34.0	4	21.7	4	72.3	2	2.06	0	21.5	0
5	17.2	4	11.3	1	34.4	4	21.7	4	59.6	4	1.25	4	16.9	3
10	16.2	2	--	--	35.3	4	22.0	4	63.0	4	--	--	--	--
11	17.0	3	12.0	3	33.0	3	20.0	2	71.0	3	1.05	1	20.0	2
12	16.5	2	--	--	--	--	--	--	--	--	--	--	--	--
13	18.4	3	11.9	3	34.7	4	23.3	3	54.1	3	1.31	4	--	--
19	17.3	4	--	--	34.6	4	18.3	0	75.5	2	1.27	4	--	--
21	--	--	--	--	--	--	--	--	73.3	2	--	--	--	--
23	17.3	4	--	--	34.3	4	21.8	4	65.8	4	1.22	4	--	--
24	19.0	1	13.0	3	36.0	3	26.0	0	80.0	1	1.30	4	--	--
25	17.0	3	12.0	3	11.0	0	28.0	0	68.0	3	1.06	1	19.0	3
26	--	--	12.1	3	--	--	20.6	3	61.5	4	2.72	0	19.3	3
32	17.6	4	12.6	4	34.8	4	21.8	4	--	--	1.22	4	18.0	4
42	19.5	0	12.6	4	34.3	4	23.4	3	66.6	4	1.21	4	18.1	4
46	16.6	3	12.2	4	36.3	3	20.6	3	60.1	4	1.26	4	--	--
48	16.9	3	10.8	0	34.0	4	20.6	3	77.0	1	1.20	3	--	--
50	17.8	4	12.7	4	37.7	1	22.4	4	51.6	3	1.29	4	18.6	3
57	16.3	2	12.9	3	32.5	2	24.3	2	60.2	4	2.05	0	13.3	0
59	18.1	3	12.5	4	35.0	4	22.1	4	38.7	0	1.07	1	15.4	1
69	15.9	1	--	--	36.8	2	< 50	NR	55.0	3	1.24	4	--	--
70	17.5	4	11.0	0	31.5	1	20.0	2	59.1	4	1.21	4	--	--
76	--	--	--	--	35.3	4	21.9	4	--	--	--	--	--	--
81	15.0	0	13.0	3	34.0	4	19.0	1	70.0	3	1.36	3	--	--
83	17.0	3	--	--	34.7	4	21.2	4	57.4	4	--	--	--	--
84	--	--	--	--	--	--	--	--	--	--	--	--	--	--
86	18.2	3	9.9	0	38.7	0	21.1	3	--	--	1.28	4	--	--
87	17.7	4	--	--	33.4	3	24.0	2	62.2	4	1.05	1	--	--
89	17.9	4	10.0	0	33.6	3	17.7	0	52.8	3	1.26	4	--	--
93	16.7	3	--	--	32.8	2	25.7	0	58.6	4	0.86	0	--	--
96	18.1	3	13.2	2	32.7	2	22.1	4	70.0	3	--	--	--	--
97	17.2	4	11.4	1	37.0	2	22.4	4	53.9	3	--	--	--	--
105	16.3	2	11.9	3	34.4	4	21.1	3	64.0	4	1.23	4	42.0	0
107	--	--	--	--	--	--	--	--	80.0	1	1.29	4	--	--
109	--	--	--	--	--	--	--	--	54.6	3	1.39	2	18.3	4
110	--	--	--	--	--	--	--	--	--	--	1.36	3	--	--
113	17.7	4	--	--	--	--	23.6	3	61.6	4	1.15	3	--	--
114	31.5	0	--	--	62.5	0	19.0	1	62.5	4	--	--	--	--
121	19.0	1	10.0	0	--	--	--	--	60.0	4	--	--	--	--
126	15.0	0	--	--	35.0	4	23.0	3	70.0	3	--	--	--	--
127	16.3	2	12.4	4	38.5	0	27.1	0	56.1	3	1.26	4	17.9	4
131	15.8	1	15.4	0	34.2	4	153.0	0	57.0	4	1.30	4	16.4	3
134	17.2	4	12.5	4	34.7	4	21.9	4	70.0	3	1.20	3	18.5	4
138	17.1	4	--	--	34.0	4	21.7	4	60.8	4	1.22	4	--	--
140	17.8	4	--	--	39.0	0	23.1	3	61.0	4	1.25	4	--	--
142	18.4	3	12.5	4	34.5	4	19.8	2	68.0	3	1.21	4	19.6	2
144	15.9	1	--	--	35.3	4	--	--	--	--	--	--	--	--
145	18.0	3	15.0	0	32.0	1	< 26	NR	75.0	2	1.12	2	< 20	NR
146	18.5	2	12.7	4	34.9	4	20.0	2	76.3	2	1.21	4	--	--
147	17.3	4	--	--	--	--	21.9	4	74.0	2	--	--	--	--
149	17.0	3	--	--	36.0	3	18.0	0	--	--	--	--	--	--
151	18.0	3	--	--	33.0	3	20.0	2	69.8	3	--	--	17.0	4
154	18.6	2	13.0	3	35.4	4	21.7	4	70.0	3	--	--	--	--
180	< 19.7	NR	11.2	0	33.4	3	21.3	4	56.3	3	0.97	0	--	--
183	14.1	0	--	--	29.6	0	24.5	2	81.0	1	--	--	--	--

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value										
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00										
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00										
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)												
Analyte =		Cd (Cadmium)		Co (Cobalt)		Cr (Chromium)		Cu (Copper)		Fe (Iron)		K (Potassium)		Li (Lithium)		
MPV =		17.5 µg/L		12.5 mg/L		34.6 mg/L		22.0 mg/L		61.7 mg/L		1.26 mg/L		17.80 mg/L		
F-pseudosigma =		0.9		0.6		1.7		1.7		10.2		0.11		1.60		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
190	--	--	--	--	32.3	2	--	--	80.0	1	--	--	--	--	--	--
191	18.7	2	12.7	4	35.1	4	21.8	4	16.8	0	1.15	3	17.3	4	--	--
193	19.0	1	--	--	36.8	2	23.4	3	61.6	4	1.43	1	--	--	--	--
196	16.8	3	12.2	4	33.7	3	22.2	4	--	--	--	--	15.8	2	--	--
203	16.8	3	--	--	35.2	4	21.0	3	52.0	3	1.32	3	--	--	--	--
204	17.8	4	--	--	32.4	2	19.2	1	55.0	3	1.23	4	--	--	--	--
209	--	--	--	--	--	--	--	--	--	--	1.93	0	--	--	--	--
212	18.5	2	13.2	2	36.5	2	24.6	1	74.3	2	1.49	0	--	--	--	--
218	--	--	--	--	--	--	--	--	--	--	1.34	3	--	--	--	--
220	17.7	4	--	--	36.9	2	21.9	4	53.7	3	--	--	17.4	4	--	--
227	17.5	4	--	--	--	--	24.2	2	60.8	4	--	--	--	--	--	--
234	17.7	4	12.5	4	34.2	4	23.3	3	59.8	4	1.24	4	18.5	4	--	--
236	18.0	3	5.0	0	33.0	3	24.0	2	55.0	3	1.60	0	16.0	2	--	--
247	18.5	2	12.8	4	32.1	2	22.7	4	62.0	4	1.28	4	16.4	3	--	--
254	--	--	12.6	4	--	--	23.0	3	61.8	4	1.20	3	--	--	--	--
255	17.8	4	--	--	34.6	4	22.7	4	--	--	--	--	--	--	--	--
259	17.8	4	12.8	4	35.2	4	22.3	4	55.5	3	1.26	4	--	--	--	--
265	17.5	4	12.5	4	35.0	4	22.0	4	62.0	4	1.20	3	17.0	4	--	--
268	--	--	12.2	4	33.4	3	--	--	60.0	4	--	--	17.6	4	--	--
270	22.6	0	12.0	3	34.5	4	23.7	3	33.5	0	1.31	4	16.5	3	--	--
273	14.9	0	9.8	0	30.3	0	26.0	0	53.5	3	1.20	3	20.9	1	--	--
277	15.5	0	12.8	4	30.0	0	20.2	2	51.3	2	1.36	3	--	--	--	--
279	--	--	--	--	--	--	--	--	--	--	1.72	0	--	--	--	--
284	17.0	3	15.0	0	36.0	3	21.0	3	< 50	NR	1.15	3	--	--	--	--
296	19.4	0	13.7	1	37.5	1	23.7	3	63.0	4	1.19	3	--	--	--	--
297	17.2	4	12.3	4	34.0	4	22.8	4	70.9	3	0.80	0	--	--	--	--
302	--	--	--	--	--	--	--	--	--	--	1.77	0	--	--	--	--
305	18.4	3	--	--	36.3	3	22.0	4	64.6	4	1.64	0	--	--	--	--
307	17.0	3	--	--	33.0	3	22.4	4	61.0	4	--	--	--	--	--	--
309	--	--	--	--	--	--	--	--	--	--	2.00	0	--	--	--	--
324	22.0	0	--	--	51.0	0	52.0	0	< 100	NR	1.26	4	--	--	--	--
330	17.6	4	12.8	4	33.5	3	22.5	4	< 1000	NR	< 1	0	--	--	--	--
332	--	--	--	--	39.9	0	--	--	80.0	1	4.11	0	--	--	--	--

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte =	Mg (Magnesium)		Mn (Manganese)		Mo (Molybdenum)		Na (Sodium)		Ni (Nickel)		Pb (Lead)		Sb (Antimony)	
MPV =	1.51	mg/L	37.4	mg/L	18.9	mg/L	43.0	mg/L	29.0	mg/L	16.5	mg/L	14.0	mg/L
F-pseudosigma =	0.07		2.3		1.6		1.5		2.0		1.2		0.9	
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	1.49	4	37.8	4	19.4	4	42.7	4	28.9	4	17.2	3	14.3	4
3	1.43	2	34.1	2	18.6	4	43.0	4	26.1	2	17.6	3	14.7	3
5	1.41	2	36.0	3	16.5	1	42.3	4	32.1	1	22.9	0	< 20	NR
10	--	--	41.0	1	--	--	--	--	--	--	16.1	4	--	--
11	1.40	2	37.0	4	17.0	2	36.5	0	28.0	3	16.0	4	12.0	0
12	1.80	0	--	--	--	--	41.0	3	--	--	204.0	0	--	--
13	1.56	3	40.5	2	< 50	NR	43.6	4	32.3	1	17.4	3	14.9	2
19	1.48	4	38.1	4	--	--	43.0	4	29.2	4	12.1	0	--	--
21	--	--	--	--	--	--	--	--	--	--	--	--	--	--
23	1.50	4	37.6	4	19.2	4	40.4	2	28.7	4	16.2	4	14.4	4
24	1.40	2	40.0	2	20.0	3	43.0	4	34.0	0	--	--	--	--
25	1.37	1	35.0	2	--	--	39.3	1	27.0	2	16.0	4	<10	0
26	1.50	4	38.4	4	18.9	4	44.6	3	--	--	15.5	3	--	--
32	1.62	2	37.5	4	18.6	4	44.8	3	29.2	4	16.4	4	13.8	4
42	1.45	3	36.2	3	19.4	4	42.6	4	29.1	4	17.7	3	14.9	2
46	1.48	4	37.7	4	23.8	0	44.4	3	27.9	3	17.4	3	15.9	0
48	1.51	4	34.0	2	20.0	3	51.5	0	26.3	2	16.9	4	13.7	4
50	1.51	4	37.8	4	20.0	3	43.0	4	29.0	4	15.5	3	14.1	4
57	1.55	3	35.5	3	18.7	4	42.4	4	31.2	2	16.8	4	13.1	2
59	1.29	0	42.0	1	18.4	4	42.6	4	27.8	3	18.4	1	15.8	1
69	< 2	NR	39.0	3	--	--	43.1	4	< 50	NR	15.3	3	12.7	2
70	1.55	3	32.8	1	18.8	4	45.2	2	26.3	2	18.0	2	14.0	4
76	1.55	3	--	--	--	--	43.8	4	30.6	3	16.6	4	13.8	4
81	1.58	2	38.0	4	19.0	4	45.2	2	26.0	1	15.0	2	13.0	2
83	1.46	3	37.4	4	--	--	41.5	3	28.1	4	--	--	--	--
84	1.52	4	29.8	0	--	--	42.1	4	--	--	--	--	--	--
86	1.56	3	37.8	4	17.3	2	44.0	4	29.1	4	16.0	4	--	--
87	1.40	2	21.2	0	< 6	0	45.0	3	42.8	0	15.8	3	--	--
89	1.60	2	40.0	2	--	--	41.4	3	29.4	4	15.4	3	13.8	4
93	1.51	4	34.5	2	--	--	46.3	1	28.5	4	16.0	4	--	--
96	--	--	42.0	1	--	--	--	--	28.3	4	19.5	0	14.3	4
97	1.52	4	38.4	4	22.3	0	43.6	4	32.2	1	17.2	3	13.6	4
105	1.52	4	35.6	3	19.8	3	43.4	4	27.8	3	16.5	4	13.9	4
107	1.55	3	40.0	2	--	--	43.2	4	--	--	--	--	--	--
109	1.58	3	33.3	1	16.3	1	42.6	4	--	--	18.5	1	--	--
110	1.51	4	--	--	--	--	3.8	0	--	--	--	--	--	--
113	1.48	4	39.0	3	--	--	42.1	4	--	--	19.0	1	13.5	3
114	--	--	36.0	3	--	--	--	--	23.0	0	< 10	0	--	--
121	1.50	4	38.0	4	--	--	42.6	4	28.0	3	--	--	--	--
126	--	--	40.0	2	--	--	46.0	2	28.0	3	19.4	0	13.0	2
127	1.44	3	34.3	2	16.5	1	44.0	4	28.3	4	16.8	4	13.1	2
131	1.30	0	35.0	2	15.4	0	42.0	4	45.3	0	16.9	4	--	--
134	1.49	4	38.4	4	19.7	3	42.3	4	30.2	3	16.4	4	14.2	4
138	1.51	4	36.8	4	17.5	3	43.3	4	29.7	4	17.1	4	14.8	3
140	1.48	4	35.0	2	--	--	44.0	4	30.0	3	20.0	0	--	--
142	1.56	3	31.0	0	19.3	4	46.8	1	28.8	4	15.2	2	14.9	2
144	--	--	--	--	--	--	--	--	27.5	3	15.7	3	--	--
145	1.49	4	38.0	4	19.0	4	43.0	4	31.0	2	< 35	NR	--	--
146	1.52	4	38.0	4	19.3	4	42.3	4	28.7	4	19.3	0	12.3	1
147	--	--	--	--	--	--	--	--	--	--	16.8	4	--	--
149	--	--	--	--	16.0	1	--	--	30.0	3	14.0	1	--	--
151	--	--	35.0	2	20.0	3	--	--	27.0	2	17.0	4	14.0	4
154	--	--	38.0	4	16.8	2	--	--	29.0	4	16.5	4	14.0	4
180	1.37	1	35.1	2	< 34.1	NR	39.7	1	28.7	4	13.5	0	14.6	3
183	--	--	--	--	--	--	--	--	--	--	15.7	3	--	--

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value										
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00										
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00										
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)												
Analyte = Mg (Magnesium)		Mn (Manganese)		Mo (Molybdenum)		Na (Sodium)		Ni (Nickel)		Pb (Lead)		Sb (Antimony)				
MPV = 1.51 mg/L		37.4 mg/L		18.9 mg/L		43.0 mg/L		29.0 mg/L		16.5 mg/L		14.0 mg/L				
F-pseudosigma = 0.07		2.3		1.6		1.5		2.0		1.2		0.9				
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
190	--	--	--	--	--	--	--	--	25.7	1	14.0	1	--	--	--	--
191	1.53	4	37.4	4	16.9	2	42.0	4	28.5	4	16.0	4	12.3	1	--	--
193	1.52	4	--	--	--	--	41.6	3	31.4	2	17.2	3	15.0	2	--	--
196	--	--	36.4	4	19.4	4	--	--	27.0	2	15.8	3	14.5	3	--	--
203	1.49	4	36.0	3	--	--	42.7	4	32.0	1	15.8	3	--	--	--	--
204	1.56	3	35.6	3	--	--	43.9	4	25.1	1	16.6	4	14.9	2	--	--
209	1.35	0	--	--	--	--	42.3	4	--	--	--	--	--	--	--	--
212	1.56	3	39.9	2	20.6	2	45.2	2	30.7	3	19.0	1	15.3	2	--	--
218	1.83	0	--	--	--	--	--	--	--	--	--	--	--	--	--	--
220	1.49	4	37.2	4	19.3	4	41.8	3	--	--	--	--	--	--	--	--
227	1.47	4	37.6	4	--	--	--	--	--	--	16.2	4	--	--	--	--
234	1.60	2	37.6	4	18.5	4	44.1	3	29.2	4	17.5	3	14.0	4	--	--
236	1.40	2	35.0	2	18.0	3	42.0	4	33.0	1	< 10	0	< 10	0	--	--
247	1.54	4	34.2	2	19.4	4	43.3	4	29.4	4	17.1	4	14.5	3	--	--
254	1.52	4	37.9	4	--	--	43.4	4	29.2	4	--	--	--	--	--	--
255	1.58	3	37.0	4	--	--	--	--	28.2	4	17.6	3	--	--	--	--
259	1.46	3	38.9	3	17.8	3	43.2	4	30.0	3	16.3	4	--	--	--	--
265	1.55	3	38.0	4	20.0	3	43.0	4	29.5	4	16.0	4	13.0	2	--	--
268	1.43	2	36.5	4	--	--	38.4	0	--	--	--	--	--	--	--	--
270	3.05	0	--	--	--	--	55.6	0	--	--	--	--	--	--	--	--
273	1.50	4	35.4	3	--	--	41.6	3	30.0	3	16.1	4	--	--	--	--
277	1.52	4	30.0	0	--	--	42.2	4	28.0	3	16.0	4	--	--	--	--
279	1.70	0	--	--	--	--	40.5	2	--	--	--	--	--	--	--	--
284	1.49	4	13.0	0	12.0	0	43.2	4	10.0	0	8.0	0	< 200	NR	--	--
296	1.51	4	39.9	2	20.5	2	46.5	1	31.1	2	19.0	1	17.9	0	--	--
297	1.37	1	37.5	4	--	--	38.6	0	31.2	2	14.8	2	12.1	0	--	--
302	2.28	0	--	--	--	--	45.1	3	--	--	--	--	--	--	--	--
305	1.50	4	--	--	17.4	3	35.9	0	31.3	2	15.0	2	--	--	--	--
307	1.50	4	49.0	0	--	--	46.0	2	--	--	15.5	3	--	--	--	--
309	1.00	0	--	--	--	--	41.0	3	--	--	--	--	--	--	--	--
324	1.11	0	32.0	0	--	--	45.2	2	46.0	0	80.0	0	--	--	--	--
330	< 1	0	36.6	4	18.5	4	54.0	0	28.5	4	16.8	4	14.7	3	--	--
332	1.52	4	31.3	0	--	--	42.0	4	--	--	--	--	--	--	--	--

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte =	Se (Selenium)		SiO ₂ (Silica)		Sr (Strontium)		Tl (Thallium)		U (Uranium)		V (Vanadium)		Zn (Zinc)	
	MPV =	mg/L	RV	mg/L	RV	mg/L	RV	mg/L	RV	mg/L	RV	mg/L	RV	mg/L
F-pseudostigma =	1.11		1.0		2.4		3.8		0.44		1.2		2.4	
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	9.67	4	15.3	4	54.2	4	51.9	4	--	--	18.5	4	40.7	4
3	10.60	3	14.0	3	50.9	2	50.0	4	--	--	17.4	3	41.4	4
5	13.10	0	14.6	4	52.4	3	--	--	--	--	17.0	2	38.7	3
10	8.00	2	--	--	--	--	--	--	--	--	--	--	40.4	4
11	9.00	3	12.8	1	51.0	2	45.0	2	--	--	17.0	2	39.0	3
12	--	--	--	--	--	--	--	--	--	--	--	--	4.5	0
13	10.20	3	15.1	4	--	--	48.0	3	--	--	< 50	NR	42.9	3
19	13.30	0	--	--	--	--	--	--	--	--	--	--	40.9	4
21	--	--	--	--	--	--	--	--	--	--	--	--	--	--
23	8.04	2	--	--	--	--	46.5	3	--	--	--	--	40.8	4
24	--	--	16.1	2	56.0	3	--	--	--	--	20.0	2	44.0	2
25	9.60	4	13.8	3	54.0	4	48.0	3	--	--	< 13	0	9.0	0
26	9.18	4	15.6	3	--	--	--	--	--	--	--	--	38.5	3
32	11.00	2	16.0	2	54.8	4	49.8	4	7.72	3	18.5	4	39.0	3
42	10.80	2	14.8	4	54.0	4	52.0	4	--	--	18.7	4	46.1	0
46	9.73	4	--	--	--	--	52.8	3	--	--	16.7	2	39.3	3
48	8.20	2	--	--	--	--	49.0	4	--	--	16.5	1	41.0	4
50	9.60	4	15.6	3	58.2	2	46.0	2	--	--	19.2	3	40.6	4
57	10.40	3	15.2	4	54.0	4	49.8	4	--	--	17.0	2	50.0	0
59	10.70	2	--	--	--	--	53.8	3	--	--	18.8	4	43.2	2
69	8.50	3	--	--	--	--	51.6	4	--	--	--	--	< 50	NR
70	9.77	4	15.5	3	55.5	4	58.0	1	9.49	0	17.3	3	40.4	4
76	--	--	14.8	4	--	--	--	--	--	--	--	--	--	--
81	< 12	NR	--	--	56.0	3	41.0	0	--	--	18.0	4	45.0	1
83	--	--	14.5	4	--	--	--	--	--	--	--	--	39.0	3
84	--	--	--	--	--	--	--	--	--	--	--	--	--	--
86	7.35	1	--	--	55.7	3	--	--	--	--	21.3	0	40.9	4
87	2.10	0	15.0	4	--	--	--	--	--	--	--	--	40.8	4
89	6.16	0	15.0	4	--	--	55.6	2	--	--	19.9	2	36.2	1
93	12.50	0	--	--	--	--	--	--	--	--	17.4	3	37.1	2
96	9.50	4	--	--	--	--	--	--	--	--	< 40	NR	38.0	2
97	9.29	4	14.1	3	36.7	0	44.2	1	--	--	17.3	3	--	--
105	9.10	4	14.8	4	56.0	3	53.5	3	--	--	17.8	4	36.0	1
107	--	--	--	--	--	--	--	--	--	--	--	--	--	--
109	11.86	0	--	--	72.6	0	--	--	--	--	--	--	--	--
110	--	--	14.8	4	--	--	--	--	--	--	--	--	--	--
113	9.38	4	--	--	53.5	4	51.5	4	--	--	--	--	42.3	3
114	--	--	--	--	--	--	--	--	--	--	--	--	43.5	2
121	--	--	15.4	3	52.0	3	--	--	--	--	--	--	40.0	4
126	9.60	4	--	--	--	--	48.0	3	--	--	--	--	39.0	3
127	9.35	4	14.1	3	52.5	3	48.9	4	< 100	NR	13.2	0	40.9	4
131	5.50	0	12.8	1	50.3	2	--	--	--	--	--	--	37.4	2
134	9.70	4	14.8	4	55.0	4	50.3	4	8.10	4	18.3	4	41.0	4
138	9.18	4	--	--	54.2	4	53.3	3	--	--	18.5	4	40.8	4
140	--	--	15.9	3	--	--	--	--	--	--	--	--	42.0	3
142	12.80	0	16.8	1	58.9	1	52.0	4	8.30	3	18.2	4	32.5	0
144	9.00	3	--	--	--	--	--	--	--	--	--	--	40.0	4
145	--	--	14.5	4	53.0	4	--	--	--	--	19.0	3	41.0	4
146	7.02	0	--	--	--	--	47.7	3	--	--	18.4	4	34.7	0
147	--	--	--	--	--	--	--	--	--	--	--	--	41.6	4
149	--	--	--	--	--	--	--	--	--	--	--	--	--	--
151	10.00	4	14.6	4	56.0	3	52.0	4	--	--	--	--	41.0	4
154	11.50	1	15.6	3	54.7	4	50.2	4	--	--	18.6	4	42.0	3
180	< 21.4	NR	--	--	--	--	36.0	0	--	--	23.3	0	35.6	1
183	--	--	--	--	--	--	--	--	--	--	--	--	43.2	2

Table 5. Laboratory performance ratings for standard reference sample T-161 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value										
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00										
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00										
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)												
Analyte =	Se (Selenium)	SiO ₂ (Silica)	Sr (Strontium)	Tl (Thallium)	U (Uranium)	V (Vanadium)	Zn (Zinc)									
MPV =	9.58 mg/L	14.8 mg/L	54.2 mg/L	50.3 mg/L	7.97 mg/L	18.4 mg/L	40.6 mg/L									
F-pseudostigma =	1.11	1.0	2.4	3.8	0.44	1.2	2.4									
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
190	--	--	16.7	1	--	--	--	--	--	--	--	--	--	--	--	--
191	--	--	14.0	3	52.5	3	--	--	7.70	3	18.9	4	48.7	0	--	--
193	9.40	4	--	--	--	--	--	--	--	--	--	--	40.0	4	--	--
196	10.20	3	--	--	53.7	4	43.9	1	7.36	2	18.1	4	48.1	0	--	--
203	9.30	4	17.5	0	--	--	--	--	--	--	--	--	40.0	4	--	--
204	8.50	3	--	--	--	--	47.3	3	--	--	--	--	38.1	2	--	--
209	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--
212	9.56	4	15.5	3	57.9	2	52.8	3	--	--	16.7	2	42.9	3	--	--
218	--	--	16.9	1	--	--	--	--	--	--	--	--	--	--	--	--
220	9.50	4	--	--	--	--	--	--	--	--	17.9	4	39.6	4	--	--
227	--	--	14.2	3	--	--	--	--	--	--	--	--	40.0	4	--	--
234	9.88	4	15.0	4	55.4	4	52.5	3	--	--	18.2	4	41.4	4	--	--
236	< 100	NR	12.6	0	52.0	3	--	--	--	--	19.0	3	36.0	1	--	--
247	11.05	2	--	--	54.0	4	53.8	3	--	--	18.7	4	46.0	0	--	--
254	--	--	15.3	4	56.3	3	--	--	--	--	--	--	40.9	4	--	--
255	9.34	4	--	--	--	--	--	--	--	--	--	--	39.1	3	--	--
259	9.30	4	14.6	4	56.3	3	--	--	--	--	--	--	40.6	4	--	--
265	10.50	3	15.0	4	52.0	3	47.0	3	7.00	0	19.0	3	35.0	0	--	--
268	--	--	--	--	--	--	--	--	--	--	--	--	28.4	0	--	--
270	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--
273	--	--	--	--	53.8	4	31.0	0	--	--	--	--	28.7	0	--	--
277	--	--	--	--	--	--	--	--	--	--	--	--	39.1	3	--	--
279	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--
284	14.00	0	6.9	0	55.0	4	56.0	1	--	--	16.0	1	43.0	3	--	--
296	11.30	1	--	--	--	--	58.1	0	--	--	22.2	0	44.3	1	--	--
297	7.20	0	13.8	2	--	--	54.0	3	--	--	18.3	4	39.6	4	--	--
302	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--
305	--	--	--	--	--	--	56.9	1	--	--	19.0	3	43.4	2	--	--
307	7.89	1	--	--	--	--	--	--	--	--	--	--	47.0	0	--	--
309	--	--	6.7	0	--	--	--	--	--	--	--	--	--	--	--	--
324	--	--	--	--	--	--	--	--	--	--	--	--	35.0	0	--	--
330	9.60	4	15.3	4	57.1	2	52.2	3	8.50	2	18.4	4	44.0	2	--	--
332	--	--	13.4	2	57.2	2	--	--	--	--	--	--	--	--	--	--

Table 6. Laboratory performance ratings for standard reference sample M-154 (major constituents)

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

		Analyte = Alkalinity		B (Boron)		Ca (Calcium)		Cl (Chloride)		DSRD (Dissolved Solids)		
		MPV =	RV	RV	RV	RV	RV	RV	RV	RV	RV	
		112 mg/L	112 mg/L	284 µg/L	284 µg/L	67.5 mg/L	67.5 mg/L	141 mg/L	141 mg/L	701 mg/L	701 mg/L	
		F-pseudostigma =	RV	RV	RV	RV	RV	RV	RV	RV	RV	
Lab	OLR	V/16	Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	Rating	
1	3.9	15	113	4	276	4	67.3	4	137	3	710	4
3	2.8	16	114	4	280	4	65.6	3	155	1	710	4
4	4.0	2	--	--	--	--	--	--	140	4	--	--
5	2.7	15	101	1	272	3	63.9	2	139	4	699	4
10	3.7	12	113	4	--	--	68.0	4	139	4	706	4
11	1.9	16	121	1	270	3	66.0	4	145	3	640	1
12	2.5	11	105	2	--	--	68.0	4	144	4	678	3
13	3.2	13	120	2	--	--	71.9	2	138	4	696	4
19	3.4	9	110	4	--	--	65.5	3	135	3	711	4
23	2.3	13	109	3	--	--	76.8	0	146	3	684	4
24	3.5	13	113	4	288	4	68.0	4	140	4	--	--
25	2.9	16	120	2	225	0	69.8	3	139	4	692	4
26	3.0	10	121	1	269	3	71.6	2	--	--	691	4
38	3.4	10	112	4	--	--	67.4	4	--	--	690	4
42	3.3	12	11	0	305	2	66.7	4	--	--	--	--
43	3.5	6	123	1	--	--	--	--	142	4	716	4
46	3.4	14	109	3	273	3	66.9	4	144	4	706	4
48	1.8	13	106	2	276	4	75.4	0	149	2	720	3
50	3.6	13	114	4	274	4	66.2	4	136	3	707	4
57	2.8	16	34	0	370	0	67.7	4	140	4	712	4
59	3.4	14	116	3	--	--	69.5	3	141	4	702	4
69	3.7	10	112	4	--	--	64.2	3	139	4	694	4
70	3.1	14	111	4	286	4	70.0	3	127	1	701	4
76	3.4	9	--	--	--	--	68.4	4	142	4	747	2
81	3.2	14	113	4	--	--	66.7	4	142	4	696	4
83	3.4	8	110	4	--	--	67.0	4	--	--	--	--
84	3.6	8	112	4	--	--	63.0	2	--	--	--	--
85	3.3	12	107	3	--	--	6.6	0	139	4	694	4
86	3.3	11	--	--	284	4	69.2	4	--	--	--	--
87	3.1	12	113	4	--	--	66.7	4	143	4	710	4
89	3.1	14	113	4	--	--	60.9	1	139	4	705	4
93	3.1	11	111	4	--	--	66.8	4	135	3	--	--
96	3.1	7	110	4	--	--	--	--	138	4	723	3
97	3.5	12	113	4	--	--	68.4	4	141	4	711	4
102	2.0	11	--	--	--	--	69.0	4	142	4	--	--
105	3.1	16	112	4	264	2	67.3	4	146	3	694	4
107	3.3	4	109	3	--	--	--	--	132	2	--	--
109	3.2	11	121	1	--	--	67.9	4	146	3	725	3
113	3.5	14	111	4	--	--	66.9	4	137	3	694	4
114	2.4	7	108	3	--	--	--	--	140	4	450	0
121	3.8	5	--	--	--	--	67.5	4	--	--	--	--
127	3.2	16	108	3	295	3	67.5	4	143	4	713	4
131	1.7	11	--	--	266	3	62.0	1	129	1	--	--
134	3.5	16	113	4	290	4	67.1	4	141	4	703	4
138	3.4	16	112	4	282	4	68.0	4	141	4	678	3
140	2.5	12	--	--	--	--	69.0	4	141	4	712	4
142	2.3	16	100	0	301	3	74.0	1	142	4	698	4
143	2.8	5	--	--	--	--	--	--	173	0	711	4
145	3.1	15	98	0	292	4	66.7	4	142	4	--	--
146	3.4	13	96	0	--	--	67.5	4	140	4	671	3
149	2.5	8	112	4	--	--	--	--	142	4	698	4
151	3.0	6	111	4	--	--	--	--	--	--	710	4
154	2.7	15	96	0	291	4	71.5	2	141	4	650	2
180	1.9	13	122	1	270	3	62.3	1	148	3	--	--
183	3.0	7	113	4	--	--	--	--	136	3	--	--

Table 6. Laboratory performance ratings for standard reference sample M-154 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

		Analyte = Alkalinity		B (Boron)		Ca (Calcium)		Cl (Chloride)		DSRD (Dissolved Solids)		
		MPV = 112 mg/L		284 µg/L		67.5 mg/L		141 mg/L		701 mg/L		
		F-pseudosigma = 4		18		3.1		4		14		
Rating	Absolute Z-value	Rating	Absolute Z-value	Rating	Absolute Z-value	Rating	Absolute Z-value	Rating	Absolute Z-value	Rating	Absolute Z-value	
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00	0 (Unsatisfactory)	greater than 2.00	NR (Not Rated)						
3 (Good)	0.51 - 1.00											
2 (Satisfactory)	1.01 - 1.50											
Lab	OLR	V/16	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
190	2.6	7	109	3	--	--	--	--	--	--	696	4
191	3.1	10	--	--	--	--	70.8	3	143	4	--	--
193	3.1	7	112	4	--	--	63.2	2	--	--	--	--
203	2.8	11	105	2	--	--	61.7	1	143	4	--	--
204	3.0	4	107	3	--	--	--	--	138	4	--	--
208	2.3	3	--	--	--	--	--	--	136	3	--	--
209	1.0	7	--	--	--	--	63.0	2	74	0	--	--
212	3.1	16	109	3	273	3	71.2	2	135	3	717	4
213	3.3	4	112	4	--	--	--	--	139	4	--	--
218	1.0	3	--	--	--	--	73.2	1	--	--	--	--
220	2.7	13	103	1	297	3	68.6	4	153	1	--	--
227	2.8	11	115	4	--	--	74.5	0	130	1	716	4
234	2.9	16	117	3	297	3	67.4	4	142	4	690	4
236	2.1	16	112	4	267	3	68.2	4	158	0	682	3
243	3.3	3	--	--	--	--	--	--	--	--	--	--
244	3.0	3	117	3	--	--	--	--	--	--	--	--
247	3.1	16	111	4	283	4	68.4	4	140	4	623	0
254	3.8	8	--	--	--	--	68.8	4	139	4	--	--
255	1.7	6	--	--	298	3	71.6	2	--	--	--	--
257	2.6	11	116	3	--	--	69.0	4	144	4	694	4
259	3.6	15	112	4	311	2	69.3	3	141	4	694	4
265	3.3	12	--	--	300	3	69.8	3	141	4	--	--
267	3.3	7	112	4	--	--	66.4	4	140	4	--	--
268	3.4	8	109	3	--	--	67.5	4	134	3	--	--
269	3.1	7	114	4	--	--	70.0	3	140	4	--	--
270	2.0	7	--	--	--	--	70.0	3	167	0	--	--
273	2.9	14	114	4	260	2	67.1	4	138	4	708	4
277	2.7	11	219	0	--	--	74.2	1	139	4	757	1
279	1.0	7	--	--	--	--	64.3	3	--	--	--	--
291	1.0	1	--	--	--	--	--	--	--	--	--	--
292	3.0	11	114	4	--	--	60.5	0	144	4	--	--
296	1.2	5	--	--	--	--	72.3	2	--	--	--	--
297	2.7	11	113	4	--	--	61.8	1	136	3	--	--
302	3.0	10	107	3	--	--	64.3	3	138	4	--	--
305	2.3	9	115	3	--	--	64.9	3	--	--	710	4
307	3.4	9	109	3	--	--	64.0	2	142	4	--	--
309	2.7	9	115	3	--	--	64.0	2	136	3	--	--
323	2.9	16	101	1	286	4	64.5	3	142	4	680	3
324	1.2	9	129	0	--	--	39.3	0	184	0	--	--
330	2.3	15	111	4	435	0	73.1	1	148	3	--	--
332	1.8	6	--	--	--	--	60.2	0	--	--	--	--

Table 6. Laboratory performance ratings for standard reference sample M-154 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = F (Fluoride)	K (Potassium)		Mg (Magnesium)		Na (Sodium)		(total Phosphorus) as P			
MPV = 2.18 mg/L	9.08 mg/L		15.1 mg/L		139 mg/L		1.31 mg/L			
F-pseudostigma = 0.13	0.62		0.6		6		0.06			
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	2.19	4	8.92	4	14.9	4	135	3	--	--
3	1.87	0	9.46	3	14.6	3	138	4	1.43	1
4	--	--	--	--	--	--	--	--	--	--
5	0.18	0	8.68	3	14.0	2	134	3	1.28	4
10	2.20	4	8.90	4	15.6	3	139	4	--	--
11	4.00	0	10.10	1	14.3	3	109	0	1.31	4
12	1.80	0	--	--	17.2	0	135	3	1.22	1
13	1.99	2	9.14	4	15.7	3	142	3	1.23	2
19	--	--	9.47	3	14.8	4	143	3	--	--
23	2.10	3	7.90	1	13.8	1	142	3	1.22	1
24	--	--	9.10	4	15.0	4	143	3	--	--
25	2.28	3	8.92	4	15.7	3	134	3	1.34	3
26	--	--	9.14	4	15.4	4	143	3	--	--
38	--	--	8.96	4	16.4	1	141	4	1.38	2
42	--	--	9.22	4	15.0	4	142	3	1.31	4
43	--	--	--	--	--	--	--	--	--	--
46	2.27	3	8.94	4	14.8	4	143	3	1.35	3
48	--	--	9.73	2	18.3	0	171	0	1.20	1
50	2.22	4	8.40	2	14.9	4	138	4	--	--
57	2.10	3	15.00	0	15.2	4	130	2	1.40	1
59	2.06	3	8.71	3	14.3	3	137	4	1.31	4
69	2.28	3	9.12	4	14.6	3	136	4	--	--
70	2.20	4	9.10	4	15.4	4	145	2	1.34	3
76	2.13	4	--	--	15.3	4	140	4	--	--
81	2.38	2	11.30	0	15.2	4	141	4	1.34	3
83	--	--	8.60	3	14.4	3	132	2	1.31	4
84	2.13	4	--	--	15.3	4	136	4	--	--
85	2.10	3	9.52	3	15.4	4	142	3	1.32	4
86	--	--	9.36	4	15.8	3	144	3	1.32	4
87	--	--	8.61	3	14.8	4	142	3	1.34	3
89	2.30	3	8.53	3	14.5	3	139	4	1.30	4
93	2.10	3	8.37	2	13.8	1	139	4	--	--
96	2.05	3	--	--	--	--	--	--	--	--
97	2.18	4	--	--	15.0	4	140	4	1.35	3
102	2.10	3	8.00	1	16.6	0	110	0	1.30	4
105	2.00	2	8.15	2	15.4	4	141	4	1.32	4
107	2.18	4	--	--	--	--	--	--	--	--
109	2.25	3	9.53	3	15.1	4	139	4	--	--
113	2.26	3	8.56	3	15.1	4	138	4	1.29	4
114	--	--	--	--	--	--	--	--	1.26	3
121	--	--	--	--	15.2	4	138	4	--	--
127	2.43	1	10.10	1	14.8	4	141	4	1.36	3
131	2.11	3	10.40	0	14.3	3	142	3	1.05	0
134	2.11	3	8.70	3	14.5	3	134	3	1.39	2
138	2.40	1	8.75	3	15.3	4	139	4	1.39	2
140	2.38	2	8.45	2	15.0	4	132	2	1.23	2
142	2.28	3	9.60	3	16.6	0	150	1	1.29	4
143	--	--	--	--	--	--	--	--	1.30	4
145	2.31	3	8.55	3	14.9	4	138	4	1.41	1
146	2.17	4	9.81	2	15.3	4	142	3	1.29	4
149	1.90	0	--	--	--	--	--	--	1.15	0
151	--	--	--	--	--	--	--	--	--	--
154	2.26	3	--	--	15.9	2	133	3	1.47	0
180	2.50	0	7.91	1	13.6	1	124	0	1.27	3
183	2.10	3	--	--	--	--	--	--	1.38	2

Table 6. Laboratory performance ratings for standard reference sample M-154 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value				
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00				
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00				
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)						
Analyte = F (Fluoride)	K (Potassium)		Mg (Magnesium)		Na (Sodium)		(total Phosphorus) as P			
MPV =	2.18 mg/L	9.08 mg/L		15.1 mg/L		139 mg/L		1.31 mg/L		
F-pseudostigma =	0.13	0.62		0.6		6		0.06		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
190	21.40	0	--	--	--	--	--	--	1.28	3
191	1.88	0	9.08	4	15.8	3	135	3	1.27	3
193	--	--	9.86	2	14.9	4	134	3	--	--
203	--	--	9.10	4	15.0	4	132	2	1.30	4
204	--	--	--	--	--	--	--	--	--	--
208	--	--	--	--	--	--	--	--	0.80	0
209	--	--	11.65	0	13.8	1	159	0	--	--
212	2.15	4	9.01	4	16.0	2	143	3	1.22	1
213	--	--	--	--	--	--	--	--	1.40	1
218	--	--	--	--	18.4	0	--	--	--	--
220	2.26	3	3.83	0	15.0	4	137	4	1.68	0
227	2.18	4	--	--	16.0	2	--	--	1.26	3
234	2.29	3	8.57	3	15.3	4	137	4	1.00	0
236	1.56	0	8.30	2	13.6	1	133	2	1.30	4
243	--	--	--	--	--	--	--	--	1.37	2
244	--	--	--	--	--	--	--	--	--	--
247	2.12	4	9.03	4	15.3	4	140	4	1.31	4
254	--	--	9.17	4	15.5	3	140	4	--	--
255	2.40	1	--	--	16.2	1	--	--	1.11	0
257	--	--	13.00	0	15.0	4	131	2	1.07	0
259	2.15	4	9.00	4	15.4	4	142	3	1.34	3
265	2.90	0	8.85	4	15.5	3	141	4	1.30	4
267	--	--	--	--	17.5	0	--	--	1.33	4
268	--	--	--	--	15.2	4	142	3	--	--
269	2.20	4	--	--	19.4	0	--	--	--	--
270	2.17	4	9.50	3	15.0	4	--	--	--	--
273	1.94	1	8.80	4	15.0	4	131	2	1.15	0
277	2.20	4	9.30	4	16.0	2	138	4	--	--
279	1.90	0	11.30	0	14.6	3	117	0	--	--
291	--	--	--	--	--	--	--	--	--	--
292	2.06	3	8.80	4	14.6	3	133	2	1.37	2
296	--	--	9.33	4	16.9	0	159	0	--	--
297	--	--	8.73	3	14.9	4	138	4	0.73	0
302	2.08	3	8.84	4	16.3	1	135	3	--	--
305	--	--	11.88	0	13.8	1	107	0	1.32	4
307	--	--	--	--	15.0	4	137	4	1.25	2
309	--	--	9.60	3	14.3	3	147	2	--	--
323	2.26	3	9.81	2	14.9	4	131	2	1.34	3
324	--	--	8.35	2	13.9	2	144	3	--	--
330	2.47	0	9.20	4	16.0	2	179	0	1.31	4
332	--	--	9.94	2	14.8	4	128	1	--	--

Table 6. Laboratory performance ratings for standard reference sample M-154 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value						
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00						
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00						
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)								
Analyte = pH		SiO ₂ (Silica)		SO ₄ (Sulfate)		Sp Cond		Sr (Strontium)		V (Vanadium)		
MPV =	8.33	10.8 mg/L		164 mg/L		1124 µS/cm		596 µg/L		16.9 µg/L		
F-pseudostigma =	0.13	0.8		5		23		29		1.8		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	8.37	4	10.9	4	161	4	1139	4	602	4	16.9	4
3	8.40	3	9.8	2	168	4	1140	4	550	1	15.8	3
4	--	--	--	--	160	4	--	--	--	--	--	--
5	7.29	0	--	--	172	3	1146	4	595	4	16.6	4
10	8.42	3	10.5	4	153	2	1133	4	--	--	--	--
11	8.16	2	9.3	1	146	0	1090	3	571	3	15.0	2
12	8.20	2	--	--	162	4	1130	4	--	--	--	--
13	8.39	4	10.9	4	168	4	1144	4	--	--	< 50	NR
19	8.38	4	--	--	--	--	1170	3	--	--	--	--
23	8.40	3	11.4	3	178	1	1105	4	--	--	--	--
24	8.40	3	11.4	3	168	4	1140	4	632	2	16.0	3
25	8.41	3	9.6	2	165	4	1117	4	605	4	< 13	0
26	8.47	2	11.4	3	--	--	1112	4	--	--	--	--
38	8.40	3	10.9	4	--	--	1130	4	--	--	--	--
42	8.27	4	10.6	4	--	--	1102	4	612	3	18.2	3
43	8.29	4	--	--	167	4	1130	4	--	--	--	--
46	8.29	4	--	--	175	2	1100	4	--	--	14.4	2
48	8.20	2	--	--	173	2	1177	3	--	--	14.8	2
50	8.21	3	10.5	4	159	3	1130	4	--	--	--	--
57	8.30	4	10.9	4	170	3	1100	4	579	3	16.3	4
59	8.25	3	11.3	3	164	4	1128	4	--	--	14.6	2
69	8.30	4	--	--	167	4	--	--	--	--	--	--
70	8.46	2	12.1	1	161	4	1070	3	--	--	--	--
76	8.49	2	--	--	--	--	1113	4	612	3	< 50	NR
81	8.48	2	--	--	168	4	1110	4	595	4	15.0	2
83	--	--	10.2	3	162	4	--	--	--	--	--	--
84	8.40	3	--	--	164	4	1125	4	--	--	--	--
85	8.30	4	--	--	166	4	1068	3	--	--	--	--
86	8.34	4	--	--	162	4	1162	3	617	3	24.5	0
87	8.33	4	5.5	0	165	4	296	0	--	--	--	--
89	8.48	2	11.0	4	164	4	1160	3	--	--	25.7	0
93	8.33	4	9.7	2	169	3	1098	4	--	--	--	--
96	8.14	1	--	--	159	3	1120	4	--	--	--	--
97	8.48	2	10.2	3	152	2	1124	4	--	--	--	--
102	--	--	12.6	0	160	3	1169	3	1	0	--	--
105	8.30	4	9.9	2	162	4	1154	3	543	1	15.8	3
107	--	--	11.2	4	--	--	--	--	--	--	--	--
109	8.18	2	--	--	163	4	1105	4	--	--	--	--
113	8.23	3	10.8	4	160	3	1131	4	560	2	--	--
114	8.03	0	--	--	157	3	1109	4	--	--	--	--
121	--	--	10.7	4	--	--	--	--	580	3	--	--
127	8.40	3	9.8	2	171	3	1110	4	596	4	17.6	4
131	--	--	9.4	1	163	4	--	--	509	0	--	--
134	8.35	4	10.4	3	163	4	1123	4	593	4	15.6	3
138	8.48	2	10.8	4	168	4	1120	4	583	4	16.6	4
140	7.38	0	11.4	3	190	0	1080	3	--	--	--	--
142	8.41	3	11.9	2	177	1	1139	4	627	2	18.9	2
143	8.48	2	--	--	--	--	1122	4	--	--	--	--
145	8.30	4	10.5	4	161	4	1113	4	586	4	21.0	0
146	8.37	4	--	--	167	4	1100	4	--	--	17.1	4
149	8.60	0	--	--	163	4	1140	4	--	--	--	--
151	8.42	3	10.4	4	--	--	1077	3	792	0	--	--
154	8.43	3	11.2	4	175	2	1147	4	596	4	17.0	4
180	8.50	2	--	--	174	2	1120	4	--	--	16.3	4
183	8.31	4	--	--	150	1	1109	4	--	--	--	--

Table 6. Laboratory performance ratings for standard reference sample M-154 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value						
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00						
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00						
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)								
Analyte = pH		SiO ₂ (Silica)		SO ₄ (Sulfate)		Sp Cond		Sr (Strontium)		V (Vanadium)		
MPV =	8.33	10.8 mg/L		164 mg/L		1124 µS/cm		596 µg/L		16.9 µg/L		
F-pseudostigma =	0.13	0.8		5		23		29		1.8		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
190	8.20	2	11.9	2	--	--	1140	4	--	--	--	--
191	--	--	11.2	4	165	4	--	--	617	3	--	--
193	--	--	11.6	3	--	--	1150	4	--	--	--	--
203	7.97	0	11.9	2	164	4	1133	4	--	--	--	--
204	--	--	12.1	1	168	4	--	--	--	--	--	--
208	--	--	--	--	164	4	--	--	--	--	--	--
209	8.52	1	--	--	156	3	--	--	--	--	--	--
212	8.41	3	10.8	4	162	4	1120	4	646	1	16.7	4
213	8.29	4	--	--	--	--	--	--	--	--	--	--
218	--	--	11.8	2	--	--	--	--	--	--	--	--
220	8.33	4	--	--	164	4	1084	3	--	--	17.3	4
227	8.16	2	11.1	4	159	3	1128	4	--	--	--	--
234	8.59	0	10.3	3	165	4	735	0	576	3	17.3	4
236	8.25	3	8.8	0	164	4	1350	0	566	2	20.0	1
243	8.32	4	--	--	--	--	1130	4	--	--	--	--
244	8.48	2	--	--	--	--	1119	4	--	--	--	--
247	8.44	3	11.7	2	173	2	1142	4	626	2	19.7	1
254	--	--	10.8	4	164	4	--	--	612	3	--	--
255	--	--	--	--	156	3	--	--	--	--	--	--
257	8.25	3	--	--	158	3	1192	2	--	--	--	--
259	8.40	3	10.8	4	164	4	1140	4	610	4	--	--
265	--	--	10.5	4	166	4	--	--	622	3	17.5	4
267	8.40	3	--	--	--	--	1132	4	--	--	--	--
268	8.16	2	--	--	164	4	1120	4	--	--	--	--
269	8.40	3	--	--	--	--	1136	4	--	--	--	--
270	8.00	0	--	--	120	0	--	--	--	--	--	--
273	8.37	4	--	--	156	3	1140	4	496	0	--	--
277	8.20	2	--	--	167	4	1119	4	--	--	--	--
279	8.12	1	--	--	--	--	1	0	--	--	--	--
291	8.14	1	--	--	--	--	--	--	--	--	--	--
292	8.41	3	--	--	167	4	1114	4	--	--	--	--
296	--	--	--	--	--	--	--	--	--	--	21.1	0
297	--	--	9.3	1	172	2	1120	4	--	--	16.4	4
302	8.30	4	--	--	148	1	1140	4	--	--	--	--
305	8.40	3	--	--	--	--	--	--	--	--	15.7	3
307	8.32	4	--	--	167	4	1147	4	--	--	--	--
309	8.33	4	4.7	0	164	4	--	--	--	--	--	--
323	8.40	3	10.8	4	160	4	1113	4	684	0	18.8	2
324	7.54	0	--	--	164	4	493	0	--	--	--	--
330	8.17	2	10.6	4	162	4	1136	4	666	0	19.3	2
332	--	--	9.0	0	--	--	--	--	599	4	--	--

Table 7. Laboratory performance ratings for standard reference sample N-65 (nutrient constituents)

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value						
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00						
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00						
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)								
Analyte = NH ₃ as N (Ammonia)			NH ₃ + Org N as N (Ammonia + Organic N)			NO ₃ as N (Nitrate)		total P as P (total Phosphorus)			PO ₄ as P (Orthophosphate as P)	
MPV = 0.124 mg/L			0.159 mg/L			0.037 mg/L		0.119 mg/L			0.112 mg/L	
F-pseudosigma = 0.023			0.067			0.009		0.011			0.007	
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	3.8	5	0.119	4	0.148	4	0.030	3	0.123	4	0.114	4
3	0.0	3	0.070	0	< 1	NR	< 0.05	NR	0.530	0	0.600	0
10	3.4	5	0.120	4	0.160	4	0.044	3	0.123	4	0.122	2
11	0.6	5	0.060	0	0.400	0	0.060	0	0.100	1	0.120	2
12	3.5	2	--	--	--	--	0.030	3	--	--	0.110	4
13	1.7	3	0.175	0	--	--	< 0.05	NR	0.122	4	0.097	1
21	4.0	5	0.124	4	0.177	4	0.037	4	0.118	4	0.112	4
23	1.3	4	0.080	1	< 0.1	NR	0.040	4	0.060	0	0.050	0
25	1.6	5	0.110	3	0.230	2	0.181	0	0.110	3	0.313	0
28	0.0	2	0.340	0	--	--	0.150	0	--	--	--	--
31	4.0	5	0.124	4	0.177	4	0.037	4	0.118	4	0.112	4
38	3.6	5	0.128	4	0.120	3	0.032	3	0.116	4	0.110	4
42	1.5	2	--	--	--	--	0.048	2	--	--	0.125	1
45	4.0	3	--	--	--	--	0.037	4	0.115	4	0.115	4
46	3.4	5	0.111	3	0.235	2	0.035	4	0.117	4	0.114	4
51	3.2	5	0.130	4	0.160	4	0.020	1	0.113	3	0.111	4
53	3.5	2	--	--	--	--	0.030	3	--	--	0.110	4
59	2.2	5	0.100	2	0.200	3	0.050	2	0.200	0	0.110	4
70	2.0	5	0.102	3	0.121	3	0.019	1	0.156	0	0.107	3
72	1.4	5	0.080	1	0.155	4	0.180	0	0.090	0	0.120	2
76	4.0	1	0.117	4	--	--	--	--	--	--	--	--
81	2.6	5	0.109	3	0.059	2	0.042	3	0.097	1	0.115	4
83	4.0	2	0.120	4	--	--	--	--	--	--	0.114	4
85	2.3	4	0.111	3	--	--	0.037	4	0.144	0	0.121	2
86	3.5	2	0.140	3	--	--	--	--	0.115	4	--	--
87	2.4	5	0.250	0	0.270	1	0.038	4	0.121	4	0.106	3
89	3.6	5	0.115	4	0.124	3	0.044	3	0.120	4	0.111	4
90	3.7	3	0.112	3	0.143	4	0.040	4	--	--	--	--
93	3.0	1	0.110	3	--	--	--	--	--	--	--	--
96	3.0	3	0.135	4	< 0.15	NR	< 0.05	NR	0.109	3	0.102	2
97	2.6	5	0.131	4	0.150	4	0.023	1	0.100	1	0.105	3
102	2.6	5	0.150	2	0.220	3	0.000	0	0.118	4	0.111	4
105	2.7	3	0.060	0	< 1	NR	< 0.04	NR	0.120	4	0.114	4
110	3.0	1	0.142	3	--	--	--	--	--	--	--	--
113	3.8	4	0.130	4	< 0.5	NR	0.036	4	0.118	4	0.118	3
114	2.5	2	0.140	3	--	--	--	--	0.130	2	--	--
127	2.0	5	0.134	4	0.131	4	0.114	0	0.132	2	0.130	0
129	2.0	5	0.069	0	0.839	0	0.034	4	0.112	3	0.118	3
134	4.0	5	0.117	4	0.158	4	0.036	4	0.120	4	0.111	4
138	3.6	5	0.150	2	0.190	4	0.040	4	0.123	4	0.109	4
140	1.8	5	0.120	4	0.130	4	0.020	1	0.060	0	0.060	0
142	3.0	5	0.166	1	0.114	3	0.046	3	0.120	4	0.111	4
143	3.8	5	0.130	4	0.160	4	0.041	4	0.129	3	0.113	4
145	2.6	5	0.110	3	0.110	3	0.040	4	0.110	3	0.130	0
146	3.0	5	0.133	4	0.144	4	0.036	4	0.111	3	0.138	0
154	3.6	5	0.120	4	0.140	4	0.036	4	0.119	4	0.103	2
158	3.0	5	0.126	4	0.129	4	0.022	1	0.127	3	0.108	3
180	3.8	5	0.125	4	0.152	4	0.040	4	0.124	3	0.114	4
183	2.0	3	--	--	--	--	1.500	0	0.117	4	0.104	2
190	2.0	4	0.625	0	--	--	0.183	0	0.117	4	0.109	4

Table 7. Laboratory performance ratings for standard reference sample N-65 (nutrient constituents)--continued

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value													
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00													
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00													
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)															
Analyte = NH ₃ as N (Ammonia)				NH ₃ + Org N as N (Ammonia + Organic N)				NO ₃ as N (Nitrate)				total P as P (total Phosphorus)				PO ₄ as P (Orthophosphate as P)			
MPV = 0.124 mg/L				0.159 mg/L				0.037 mg/L				0.119 mg/L				0.112 mg/L			
F-pseudosigma = 0.023				0.067				0.009				0.011				0.007			
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating			
191	2.0	2	--	--	--	--	0.034	4	--	--	0.130	0							
193	3.8	4	0.118	4	0.106	3	0.037	4	0.120	4	--	--							
198	3.5	4	0.122	4	--	--	0.031	3	0.110	3	0.110	4							
203	3.2	5	0.130	4	0.130	4	0.040	4	0.130	2	0.120	2							
205	1.5	2	0.105	3	--	--	0.121	0	--	--	--	--							
209	3.5	2	--	--	0.220	3	0.038	4	--	--	--	--							
212	0.8	5	0.071	0	0.247	2	0.220	0	0.106	2	0.290	0							
213	3.0	2	< 1	NR	< 1	NR	--	--	0.130	2	0.110	4							
220	0.0	5	0.480	0	0.592	0	0.104	0	0.037	0	0.130	0							
224	2.0	5	0.183	0	0.517	0	0.045	3	0.109	3	0.112	4							
227	3.8	4	0.120	4	0.185	4	< 0.199	NR	0.110	3	0.112	4							
234	3.3	4	0.115	4	--	--	0.034	4	0.140	1	0.110	4							
243	2.7	3	0.110	3	--	--	0.020	1	0.120	4	--	--							
247	2.0	5	0.107	3	0.117	3	0.020	1	0.113	3	0.090	0							
253	0.8	5	0.211	0	0.107	3	0.000	0	0.168	0	0.125	1							
255	3.3	3	0.150	2	--	--	< 0.1	NR	0.120	4	0.110	4							
287	0.0	2	0.243	0	--	--	--	--	--	--	0.190	0							
292	2.8	4	0.140	3	--	--	0.130	0	0.120	4	0.114	4							
297	1.6	5	0.139	3	0.299	0	0.068	0	0.100	1	0.112	4							
305	1.5	4	0.158	2	< 1	NR	0.250	0	0.250	0	0.111	4							
313	2.2	5	0.158	2	0.104	3	0.032	3	0.109	3	0.094	0							
314	1.6	5	0.143	3	0.263	1	0.034	4	0.196	0	0.178	0							
316	4.0	5	0.131	4	0.132	4	0.036	4	0.118	4	0.110	4							
318	4.0	5	0.120	4	0.170	4	0.036	4	0.115	4	0.111	4							
319	2.2	5	0.180	0	0.440	0	0.040	4	0.127	3	0.110	4							
320	3.4	5	0.146	3	0.210	3	0.034	4	0.125	3	0.110	4							
321	3.0	4	0.120	4	--	--	0.030	3	0.125	3	0.122	2							
322	4.0	1	--	--	--	--	--	--	--	--	0.109	4							

Table 8. Laboratory performance ratings for standard reference sample N-66 (nutrient constituents)

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value									
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00									
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00									
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)											
Lab	Analyte = NH ₃ as N (Ammonia)			NH ₃ + Org N as N (Ammonia + Organic N)			NO ₃ as N (Nitrate)			total P as P (total Phosphorus)			PO ₄ as P (Orthophosphate as P)		
	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
1	3.8	5	0.766	4	0.923	4	0.950	4	0.872	4	0.787	3			
3	0.0	4	0.320	0	< 1	NR	0.182	0	0.591	0	0.589	0			
11	3.8	5	0.790	4	1.000	3	0.910	4	0.870	4	0.800	4			
12	0.0	5	0.630	0	0.600	0	0.190	0	0.580	0	0.566	0			
13	2.5	4	1.000	0	--	--	0.970	3	0.894	3	0.826	4			
19	2.5	2	0.700	2	--	--	0.970	3	--	--	--	--			
23	1.0	5	0.690	1	0.960	4	0.200	0	0.770	0	0.550	0			
25	1.4	5	0.680	1	1.360	0	0.927	4	0.780	1	0.877	1			
28	1.5	2	0.980	0			0.960	3	--	--	--	--			
38	3.4	5	0.817	3	0.900	4	0.924	4	0.902	2	0.826	4			
39	2.2	5	0.809	3	1.120	1	0.932	4	0.832	3	0.942	0			
42	3.0	2	--	--	--	--	0.918	4	--	--	0.861	2			
45	3.0	3	--	--	--	--	0.962	3	0.914	2	0.830	4			
46	3.8	5	0.754	4	0.961	4	0.948	4	0.856	4	0.844	3			
50	3.6	5	0.806	3	1.020	3	0.945	4	0.844	4	0.796	4			
53	1.5	2	--	--	--	--	0.870	2	--	--	0.885	1			
59	2.4	5	0.740	3	1.000	3	1.000	2	1.190	0	0.820	4			
70	2.4	5	0.646	0	0.826	3	0.900	3	0.846	4	0.767	2			
72	3.4	5	0.760	4	0.952	4	0.880	2	0.890	3	0.800	4			
81	2.4	5	0.855	1	0.695	0	0.940	4	0.833	3	0.815	4			
83	3.0	3	0.720	2	--	--	0.930	4	--	--	0.841	3			
84	0.0	2	1.000	0	--	--	0.694	0	--	--	--	--			
85	1.5	4	0.685	1	--	--	0.954	4	1.040	0	0.887	1			
86	2.3	3	0.743	3	--	--	0.870	2	0.897	2	--	--			
87	2.2	5	0.860	1	0.990	3	0.811	0	0.820	3	0.807	4			
89	3.8	5	0.745	3	0.950	4	0.916	4	0.837	4	0.802	4			
90	3.0	3	0.765	4	0.847	3	0.978	2	--	--	--	--			
93	3.0	1	0.810	3	--	--	--	--	--	--	--	--			
96	4.0	5	0.749	4	0.899	4	0.920	4	0.876	4	0.795	4			
97	3.2	5	0.812	3	0.970	4	0.983	2	0.870	4	0.790	3			
102	2.2	5	0.780	4	0.850	3	1.160	0	0.860	4	0.540	0			
105	2.4	5	0.700	2	4.000	0	0.900	3	0.830	3	0.810	4			
107	3.8	4	0.795	3			0.924	4	0.851	4	0.811	4			
113	3.4	5	0.763	4	1.046	2	0.961	3	0.856	4	0.830	4			
114	2.7	3	0.790	4	--	--	1.070	0	0.840	4	--	--			
126	4.0	1	--	--	--	--	0.940	4	--	--	--	--			
127	3.0	5	0.710	2	0.834	3	0.918	4	0.836	4	0.862	2			
129	3.2	5	0.805	3	1.020	3	0.974	3	0.879	3	0.808	4			
134	2.8	5	0.862	1	0.846	3	0.952	4	0.883	3	0.840	3			
138	3.4	5	0.759	4	0.951	4	0.930	4	0.832	3	0.764	2			
140	2.2	5	1.010	0	1.030	3	0.979	2	0.820	3	0.790	3			
142	2.8	5	0.962	0	0.879	4	0.907	3	0.816	3	0.825	4			
143	3.4	5	0.770	4	0.970	4	0.925	4	0.882	3	0.862	2			
145	2.2	5	0.750	4	0.860	3	0.880	2	0.940	0	0.860	2			
146	2.6	5	0.847	1	0.927	4	0.985	2	0.892	3	0.846	3			
154	3.6	5	0.750	4	1.010	3	0.970	3	0.844	4	0.805	4			
158	2.8	5	0.756	4	0.857	3	0.957	3	0.957	0	0.828	4			
180	3.2	5	0.782	4	0.974	4	1.010	1	0.880	3	0.805	4			
183	2.5	4	0.770	4	--	--	1.100	0	0.818	3	0.788	3			
190	3.0	4	0.017	0	--	--	0.930	4	0.856	4	0.800	4			
191	2.0	2	--	--	--	--	0.930	4	--	--	0.920	0			
193	3.5	4	0.748	4	0.825	3	0.936	4	0.819	3	--	--			
198	4.0	4	0.788	4	--	--	0.934	4	0.844	4	0.808	4			
203	3.6	5	0.830	2	0.910	4	0.920	4	0.870	4	0.830	4			
204	2.4	5	0.824	2	0.922	4	0.966	3	0.781	1	0.853	2			

Table 9. Laboratory performance ratings for standard reference sample P-34 (low ionic strength constituents)

[MPV, most probable value; mg/L, milligrams per liter; μ S/cm, microseimens per centimeter at 25 °C; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Acidity as CaCO ₃			Ca (Calcium)		Cl (Chloride)		F (Fluoride)		K (Potassium)		Mg (Magnesium)			
Lab	OLR	V/11	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
1	2.6	9			1.62	4	4.58	1	0.140	3	0.232	4	0.588	4
2	2.8	8			1.77	1	4.63	0	0.164	4	0.212	3	0.624	3
3	1.9	10	10.0	1	1.61	4	3.51	0	0.167	4	1.160	0	0.594	4
5	2.3	9			1.53	2	4.18	4	0.180	3	< 1	NR	0.552	2
11	2.6	9			1.60	4	4.00	3	0.100	0	0.150	0	0.540	2
12	1.0	1												
23	2.2	9			2.40	0	4.23	4	0.018	0	0.200	2	0.600	4
25	2.4	9	< 8	NR	1.28	0	4.20	4	0.170	4	0.189	2	0.519	1
38	3.6	7			1.68	3					0.250	4	0.625	3
46	3.9	7			1.66	4	4.20	4					0.592	4
48	2.6	8			1.65	4	4.00	3			0.258	3	0.660	1
59	1.5	10			0.65	0	1.32	0	0.160	4	0.350	0	1.010	0
81	2.8	10	3.35	4	1.63	4	4.05	3	0.212	0	0.202	3	0.617	3
83	3.8	4			1.64	4							0.580	4
85	2.8	10			1.60	4	4.20	4	0.160	4	0.306	1	0.661	1
86	3.6	7			1.66	4	4.15	4			0.261	3	0.623	3
89	3.5	11	4.6	4	1.62	4	4.12	4	0.160	4	0.230	4	0.620	3
93	3.3	9			1.57	3	4.35	3	0.170	4	0.230	4	0.570	3
105	2.9	8	10.8	1	1.65	4	4.00	3	< 0.2	NR	< 1	NR	0.610	4
107	3.5	2												
110	3.6	8			1.58	3	4.22	4			0.230	4	0.577	4
113	1.9	10			1.34	0	4.48	2	0.162	4	0.182	1	0.532	1
127	2.9	11	7.09	3	1.75	2	4.35	3	0.141	3	0.228	4	0.662	1
134	3.7	10			1.66	4	4.19	4	0.160	4	0.240	4	0.582	4
138	3.2	10			1.67	3	3.99	3	0.214	0	0.235	4	0.592	4
140	2.0	10			1.59	4	3.96	2	0.210	0	0.273	3	0.570	3
143	2.8	4					3.56	0						
145	3.0	9			1.63	4	4.19	4	0.140	3	< 0.71	NR	0.590	4
146	3.3	10	7.42	3	1.66	4	4.17	4	0.174	3	0.199	2	0.599	4
180	2.9	9			1.55	3	4.25	4	0.193	2	< 0.62	NR	0.538	2
183	2.2	5					5.16	0	0.170	4				
190	3.8	4							0.156	4				
191	3.1	8			1.67	3	4.09	4	0.160	4	0.220	4	0.630	2
193	3.4	5			1.59	4					0.257	3	0.578	4
203	1.9	9			1.47	1	3.90	2			0.220	4	0.510	0
204	4.0	2												
208	2.0	1					3.96	2						
209	1.4	7			1.35	0	4.32	3			0.880	0	0.530	1
220	1.1	9	1.75	3	1.58	3	4.66	0	0.209	0			0.570	3
227	3.7	7			1.69	3	4.22	4	0.160	4			0.600	4
237	3.6	5	2.50	3	1.55	3					0.243	4		
243	4.0	2												
244	3.5	2												
247	1.8	9			1.80	0	3.76	0	0.150	4	0.294	1	0.595	4
255	2.0	3			1.70	3			< 0.5	NR			0.626	3
265	2.6	8			1.65	4	3.76	0	< 0.1	0	0.240	4	0.600	4
268	2.6	7			1.81	0	3.90	2					0.580	4
270	1.1	7			0.90	0	4.20	4	0.100	0	0.250	4		
273	3.3	11	3.24	4	1.68	3	4.15	4	0.150	4	0.240	4	0.620	3
277	2.6	9			1.60	4	4.31	3	0.190	2	0.200	2	0.610	4
279	1.4	7			1.63	4			0.300	0	0.330	0	0.580	4
296	0.5	4			1.74	2					0.140	0	0.500	0
301	2.8	4					4.26	4						
308	1.0	1												
322	4.0	2					4.15	4						
332	1.3	4			1.43	0					0.940	0	0.590	4

Table 9. Laboratory performance ratings for standard reference sample P-34 (low ionic strength constituents)--Continued

[MPV, most probable value; mg/L, milligrams per liter; μ S/cm, microseimens per centimeter at 25 °C; Lab, laboratory number; ULK, overall laboratory rating

for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Na (Sodium)		pH		PO ₄ as P		SO ₄ (Sulfate)		Specific Conductance		
MPV =	0.450 mg/L		4.32		0.108	mg/L	0.437 mg/L	41.8	μ S/cm	
F-pseudostigma =	0.037		0.10		0.011		0.141	1.7		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	0.441	4	6.74	0			0.359	3	31.2	0
2	0.474	3	4.36	4			0.459	4		
3	0.750	0	4.27	3	0.150	0	< 10	NR	43.0	3
5	0.417	3	3.31	0	0.098	3	0.910	0	42.4	4
11	0.430	3	4.32	4	0.100	3	< 1	NR	41.0	4
12			4.13	1						
23	0.420	3	4.31	4	0.090	1	< 1	NR	39.1	2
25	0.432	4	4.41	3	0.113	4	< 5	NR	51.0	0
38	0.440	4	4.40	3	0.107	4			42.8	4
46	0.452	4	4.34	4	0.110	4			43.0	3
48	0.483	3	4.00	0	0.114	3	< 1	NR	42.5	4
59	0.480	3	4.23	3	0.096	2	14.700	0	43.1	3
81	0.564	0	4.25	3	0.109	4	< 5	NR	41.4	4
83	0.430	3					0.370	4		
85	0.477	3	4.32	4	0.113	4	0.290	2	38.0	1
86	0.431	3	4.32	4					42.2	4
89	0.470	3	4.07	0	0.107	4	0.478	4	42.3	4
93	0.420	3	4.40	3			0.400	4	40.6	3
105	0.450	4	4.50	1	0.100	3	< 1	NR	40.5	3
107			4.38	3					42.3	4
110	0.439	4	4.32	4			0.437	4	39.4	2
113	< 0.2	0	4.50	1	0.110	4	0.320	3	40.5	3
127	0.453	4	4.36	4	0.125	1	0.462	4	43.3	3
134	0.385	1	4.35	4	0.107	4	0.450	4	41.8	4
138	0.442	4	4.32	4	0.097	3	0.373	4	39.7	3
140	0.424	3	4.04	0	0.090	1	1.000	0	41.0	4
143			4.36	4	0.114	3			41.0	4
145	0.370	0	4.25	3	0.110	4	0.630	2	39.9	3
146	0.465	4	4.36	4	0.085	1	< 5	NR	42.0	4
180	0.483	3	4.30	4	0.110	4	0.411	4	46.0	0
183			4.40	3	0.110	4			47.8	0
190			4.26	3	0.108	4			41.4	4
191	0.440	4			0.220	0	0.490	4		
193	0.450	4							38.8	2
203	0.400	2	3.59	0	0.102	4	0.800	0	41.9	4
204			4.35	4					42.3	4
208										
209	0.410	2	3.98	0			0.430	4		
220	0.530	0			0.150	0	0.654	1	36.0	0
227			4.24	3	0.107	4	< 0.780	NR	40.8	4
237	0.455	4	4.28	4						
243			4.29	4					41.6	4
244			4.32	4					42.9	3
247	0.561	0	4.38	3	0.070	0	< 1	NR	42.4	4
255					0.132	0	< 15	NR		
265	0.450	4			0.095	2	0.350	3		
268	0.440	4	3.80	0			0.420	4	42.0	4
270	1.400	0	4.06	0			1.700	0		
273	0.500	2	4.31	4	0.080	0	0.370	4	42.6	4
277	0.400	2	4.10	0			0.370	4	39.0	2
279	0.410	2	3.78	0					0.0	0
296	0.700	0								
301			4.41	3			0.078	0	42.5	4
308									38.6	1
322							0.440	4		
332	0.520	1								

Table 8. Laboratory performance ratings for standard reference sample N-66 (nutrient constituents)

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value				
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00				
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00				
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)						
Analyte = NH ₃ as N (Ammonia)		NH ₃ + Org N as N (Ammonia + Organic N)		NO ₃ as N (Nitrate)		total P as P (total Phosphorus)		PO ₄ as P (Orthophosphate as P)		
MPV = 0.770 mg/L		0.925 mg/L		0.931 mg/L		0.856 mg/L		0.811 mg/L		
F-pseudosigma = 0.048		0.105		0.033		0.040		0.030		
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating
205	3.5	2	0.763	4	--	--	0.975	3	--	--
208	4.0	2	--	--	--	--	0.930	4	--	--
209	4.0	2	--	--	0.901	4	0.954	4	--	--
212	0.0	5	0.442	0	0.564	0	0.290	0	0.565	0
213	2.7	3	< 1	NR	1.800	0	--	--	0.850	4
220	2.6	5	0.770	4	0.917	4	0.971	3	0.766	0
224	2.4	5	0.789	4	1.170	0	0.993	2	0.882	3
227	2.8	5	0.720	2	0.858	3	0.900	3	0.830	3
234	3.0	4	0.760	4	--	--	0.921	4	0.820	3
243	2.3	3	0.810	3	--	--	0.930	4	0.750	0
247	1.8	5	0.839	2	0.867	3	0.800	0	0.814	2
253	3.4	5	0.819	2	0.889	4	0.906	3	0.862	4
255	3.8	4	0.800	3	--	--	0.920	4	0.870	4
287	0.0	2	0.971	0	--	--	--	--	--	--
292	3.8	4	0.800	3	--	--	0.920	4	0.860	4
297	3.0	5	0.799	3	0.726	1	0.901	3	0.839	4
305	3.4	5	0.794	4	1.000	3	0.960	3	0.857	4
307	0.0	3	0.060	0	--	--	< 0.014	0	0.069	0
308	3.0	4	0.710	2	--	--	0.940	4	0.870	4
313	3.2	5	0.769	4	0.764	1	0.930	4	0.857	4
316	3.2	5	0.810	3	0.883	4	0.955	3	0.890	3
319	2.8	5	0.750	4	1.210	0	0.940	4	0.903	2
320	2.0	5	0.920	0	1.924	0	0.955	3	0.863	4

Table 9. Laboratory performance ratings for standard reference sample P-34 (low ionic strength constituents)

[MPV, most probable value; mg/L, milligrams per liter; μ S/cm, microseimens per centimeter at 25 °C; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Acidity as CaCO ₃			Ca (Calcium)		Cl (Chloride)		F (Fluoride)		K (Potassium)		Mg (Magnesium)			
Lab	OLR	V/11	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
1	2.6	9			1.62	4	4.58	1	0.140	3	0.232	4	0.588	4
2	2.8	8			1.77	1	4.63	0	0.164	4	0.212	3	0.624	3
3	1.9	10	10.0	1	1.61	4	3.51	0	0.167	4	1.160	0	0.594	4
5	2.3	9			1.53	2	4.18	4	0.180	3	< 1	NR	0.552	2
11	2.6	9			1.60	4	4.00	3	0.100	0	0.150	0	0.540	2
12	1.0	1												
23	2.2	9			2.40	0	4.23	4	0.018	0	0.200	2	0.600	4
25	2.4	9	< 8	NR	1.28	0	4.20	4	0.170	4	0.189	2	0.519	1
38	3.6	7			1.68	3					0.250	4	0.625	3
46	3.9	7			1.66	4	4.20	4					0.592	4
48	2.6	8			1.65	4	4.00	3			0.258	3	0.660	1
59	1.5	10			0.65	0	1.32	0	0.160	4	0.350	0	1.010	0
81	2.8	10	3.35	4	1.63	4	4.05	3	0.212	0	0.202	3	0.617	3
83	3.8	4			1.64	4							0.580	4
85	2.8	10			1.60	4	4.20	4	0.160	4	0.306	1	0.661	1
86	3.6	7			1.66	4	4.15	4			0.261	3	0.623	3
89	3.5	11	4.6	4	1.62	4	4.12	4	0.160	4	0.230	4	0.620	3
93	3.3	9			1.57	3	4.35	3	0.170	4	0.230	4	0.570	3
105	2.9	8	10.8	1	1.65	4	4.00	3	< 0.2	NR	< 1	NR	0.610	4
107	3.5	2												
110	3.6	8			1.58	3	4.22	4			0.230	4	0.577	4
113	1.9	10			1.34	0	4.48	2	0.162	4	0.182	1	0.532	1
127	2.9	11	7.09	3	1.75	2	4.35	3	0.141	3	0.228	4	0.662	1
134	3.7	10			1.66	4	4.19	4	0.160	4	0.240	4	0.582	4
138	3.2	10			1.67	3	3.99	3	0.214	0	0.235	4	0.592	4
140	2.0	10			1.59	4	3.96	2	0.210	0	0.273	3	0.570	3
143	2.8	4					3.56	0						
145	3.0	9			1.63	4	4.19	4	0.140	3	< 0.71	NR	0.590	4
146	3.3	10	7.42	3	1.66	4	4.17	4	0.174	3	0.199	2	0.599	4
180	2.9	9			1.55	3	4.25	4	0.193	2	< 0.62	NR	0.538	2
183	2.2	5					5.16	0	0.170	4				
190	3.8	4							0.156	4				
191	3.1	8			1.67	3	4.09	4	0.160	4	0.220	4	0.630	2
193	3.4	5			1.59	4					0.257	3	0.578	4
203	1.9	9			1.47	1	3.90	2			0.220	4	0.510	0
204	4.0	2												
208	2.0	1					3.96	2						
209	1.4	7			1.35	0	4.32	3			0.880	0	0.530	1
220	1.1	9	1.75	3	1.58	3	4.66	0	0.209	0			0.570	3
227	3.7	7			1.69	3	4.22	4	0.160	4			0.600	4
237	3.6	5	2.50	3	1.55	3					0.243	4		
243	4.0	2												
244	3.5	2												
247	1.8	9			1.80	0	3.76	0	0.150	4	0.294	1	0.595	4
255	2.0	3			1.70	3			< 0.5	NR			0.626	3
265	2.6	8			1.65	4	3.76	0	< 0.1	0	0.240	4	0.600	4
268	2.6	7			1.81	0	3.90	2					0.580	4
270	1.1	7			0.90	0	4.20	4	0.100	0	0.250	4		
273	3.3	11	3.24	4	1.68	3	4.15	4	0.150	4	0.240	4	0.620	3
277	2.6	9			1.60	4	4.31	3	0.190	2	0.200	2	0.610	4
279	1.4	7			1.63	4			0.300	0	0.330	0	0.580	4
296	0.5	4			1.74	2					0.140	0	0.500	0
301	2.8	4					4.26	4						
308	1.0	1												
322	4.0	2					4.15	4						
332	1.3	4			1.43	0					0.940	0	0.590	4

Table 9. Laboratory performance ratings for standard reference sample P-34 (low ionic strength constituents)--Continued

[MPV, most probable value; mg/L, milligrams per liter; μ S/cm, microseimens per centimeter at 25 °C; Lab, laboratory number; ULK, overall laboratory rating

for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Na (Sodium)		pH		PO ₄ as P		SO ₄ (Sulfate)		Specific Conductance		
MPV =	0.450 mg/L		4.32		0.108	mg/L	0.437 mg/L	41.8	μ S/cm	
F-pseudostigma =	0.037		0.10		0.011		0.141	1.7		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	0.441	4	6.74	0			0.359	3	31.2	0
2	0.474	3	4.36	4			0.459	4		
3	0.750	0	4.27	3	0.150	0	< 10	NR	43.0	3
5	0.417	3	3.31	0	0.098	3	0.910	0	42.4	4
11	0.430	3	4.32	4	0.100	3	< 1	NR	41.0	4
12			4.13	1						
23	0.420	3	4.31	4	0.090	1	< 1	NR	39.1	2
25	0.432	4	4.41	3	0.113	4	< 5	NR	51.0	0
38	0.440	4	4.40	3	0.107	4			42.8	4
46	0.452	4	4.34	4	0.110	4			43.0	3
48	0.483	3	4.00	0	0.114	3	< 1	NR	42.5	4
59	0.480	3	4.23	3	0.096	2	14.700	0	43.1	3
81	0.564	0	4.25	3	0.109	4	< 5	NR	41.4	4
83	0.430	3					0.370	4		
85	0.477	3	4.32	4	0.113	4	0.290	2	38.0	1
86	0.431	3	4.32	4					42.2	4
89	0.470	3	4.07	0	0.107	4	0.478	4	42.3	4
93	0.420	3	4.40	3			0.400	4	40.6	3
105	0.450	4	4.50	1	0.100	3	< 1	NR	40.5	3
107			4.38	3					42.3	4
110	0.439	4	4.32	4			0.437	4	39.4	2
113	< 0.2	0	4.50	1	0.110	4	0.320	3	40.5	3
127	0.453	4	4.36	4	0.125	1	0.462	4	43.3	3
134	0.385	1	4.35	4	0.107	4	0.450	4	41.8	4
138	0.442	4	4.32	4	0.097	3	0.373	4	39.7	3
140	0.424	3	4.04	0	0.090	1	1.000	0	41.0	4
143			4.36	4	0.114	3			41.0	4
145	0.370	0	4.25	3	0.110	4	0.630	2	39.9	3
146	0.465	4	4.36	4	0.085	1	< 5	NR	42.0	4
180	0.483	3	4.30	4	0.110	4	0.411	4	46.0	0
183			4.40	3	0.110	4			47.8	0
190			4.26	3	0.108	4			41.4	4
191	0.440	4			0.220	0	0.490	4		
193	0.450	4							38.8	2
203	0.400	2	3.59	0	0.102	4	0.800	0	41.9	4
204			4.35	4					42.3	4
208										
209	0.410	2	3.98	0			0.430	4		
220	0.530	0			0.150	0	0.654	1	36.0	0
227			4.24	3	0.107	4	< 0.780	NR	40.8	4
237	0.455	4	4.28	4						
243			4.29	4					41.6	4
244			4.32	4					42.9	3
247	0.561	0	4.38	3	0.070	0	< 1	NR	42.4	4
255					0.132	0	< 15	NR		
265	0.450	4			0.095	2	0.350	3		
268	0.440	4	3.80	0			0.420	4	42.0	4
270	1.400	0	4.06	0			1.700	0		
273	0.500	2	4.31	4	0.080	0	0.370	4	42.6	4
277	0.400	2	4.10	0			0.370	4	39.0	2
279	0.410	2	3.78	0					0.0	0
296	0.700	0								
301			4.41	3			0.078	0	42.5	4
308									38.6	1
322							0.440	4		
332	0.520	1								

Table 10. Laboratory performance ratings for standard reference sample Hg-30 (Mercury)

[MPV, most probable value; ug/L, micrograms per liter; Lab, laboratory number; V/1, number of reported values of 1 value; RV, reported value; <, less than; NR, not rated.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Hg (Mercury)
 MPV = 1.92 µg/L
 F-pseudostigma = 0.16

Lab	V/1	RV	Rating
1	1	2.00	3
3	1	1.90	4
10	1	2.10	2
11	1	1.75	2
12	1	2.10	2
13	1	1.91	4
32	1	2.10	2
46	1	1.87	4
47	1	1.95	4
50	1	1.91	4
54	1	1.90	4
59	1	1.93	4
70	1	2.13	2
81	1	1.79	3
86	1	1.34	0
87	1	2.00	3
89	1	1.76	2
96	1	2.09	2
97	1	2.51	0
105	1	2.28	0
109	1	1.96	4
127	1	2.09	2
134	1	2.01	3
142	1	1.92	4
144	1	1.87	4
145	1	1.07	0
146	1	1.67	1
147	1	1.95	4
149	1	1.90	4
154	1	1.90	4
193	1	1.84	3
212	1	1.67	1
213	1	1.50	0
234	1	1.65	1
245	1	1.89	4
247	1	2.20	1
259	1	1.92	4
265	1	1.95	4
277	1	1.73	2
284	1	2.07	3
292	1	1.60	1
298	1	1.95	4
304	1	1.96	4
307	1	2.33	0
330	1	1.80	3

Table 11. *Statistical summary of reported data for standard reference sample T-161 (trace constituents)*

Definition of analytical methods, abbreviations, and symbols

Analytical methods and codes

0. Other/Not reported	=	atomic absorption: direct air
1. AA: direct, air	=	atomic absorption: direct, nitrous oxide
2. AA: direct, N ₂ O	=	atomic absorption: graphite furnace
3. AA: graphite furnace	=	inductively coupled plasma
4. ICP	=	direct current plasma
5. DCP	=	inductively coupled plasma / mass spectrometry
6. ICP/MS	=	ion chromatography
7. IC	=	atomic absorption: cold vapor
8. AA: cold vapor	=	atomic absorption: hydride (reducing agent specified)
11. AA: hydride	=	atomic absorption: flame emission
12. AA: flame emission	=	Titration colorimetric (specify color reagent)
20. Titration color	=	colorimetric (color reagent specified)
22. Color	=	


Abbreviations and figure symbols

N	=	number of analyses--(excluding less than values)
MPV	=	most probable value
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit
Lwl	=	lower warning limit
Ucl	=	upper control limit
Lcl	=	lower control limit
µg/L	=	micrograms per liter
mg/L	=	milligrams per liter
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported

<u>Constituent</u>	<u>page</u>	<u>Constituent</u>	<u>page</u>
Ag Silver	36	Mg Magnesium	50
Al Aluminum	37	Mn Manganese	51
As Arsenic	38	Mo Molybdenum	52
B Boron	39	Na Sodium	53
Ba Barium	40	Ni Nickel	54
Be Beryllium	41	Pb Lead	55
Ca Calcium	42	Sb Antimony	56
Cd Cadmium	43	Se Selenium	57
Co Cobalt	44	SiO ₂ Silica	58
Cr Chromium	45	Sr Strontium	59
Cu Copper	46	Tl Thallium	60
Fe Iron	47	U Uranium	61
K Potassium	48	V Vanadium	62
Li Lithium	49	Zn Zinc	63

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Ag (Silver) Concentration Unit : mg/L



SUMMARY

N =	3	16	22	20
Minimum =	10.4	14.0	11.6	15.0
Maximum =	33.0	19.3	36.4	22.5
Median =		17.1	17.2	17.5
F-pseudosigma =		1.2	2.1	0.9


- 1. AA: direct, air
- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	17.3
F-pseudosigma =	1.3
N =	61
Uh =	18.0
Lh =	16.3


Lab	Rating	Z-value	1	3	4	6
1	3	0.56	--	--	--	18.0
3	2	-1.02	--	--	16.0	--
5	2	-1.10	--	--	15.9	--
11	1	-1.80	--	--	15.0	--
12	2	-1.18	--	15.8	--	--
13	0	2.27	--	--	20.2	--
19	1	1.80	--	--	19.6	--
23	3	-0.55	--	16.6	--	--
25	2	-1.02	--	--	16.0	--
26	2	-1.18	--	15.8	--	--
32	4	0.24	--	--	--	17.6
42	3	0.71	--	--	--	18.2
46	4	0.47	--	17.9	--	--
48	2	-1.49	--	--	--	15.4
50	4	-0.08	--	--	--	17.2
57	4	-0.39	--	--	16.8	--
59	4	0.00	--	--	--	17.3
69	4	-0.16	--	17.1	--	--
70	2	1.18	--	18.8	--	--
81	2	-1.02	--	--	16.0	--
86	2	1.18	--	--	18.8	--
87	4	-0.08	--	--	17.2	--
89	4	-0.39	--	16.8	--	--
96	4	0.24	--	--	17.6	--
97	4	0.16	--	--	17.5	--
105	4	0.00	--	--	--	17.3
113	4	-0.24	--	17.0	--	--
114	0	4.47	23.0	--	--	--
126	0	-2.59	--	14.0	--	--
127	2	-1.02	--	16.0	--	--
131	0	14.98	--	--	36.4	--
134	4	-0.08	--	--	17.2	--
138	3	-0.78	--	--	--	16.3
140	0	-5.41	10.4	--	--	--
142	4	-0.47	--	--	--	16.7
144	1	1.57	--	19.3	--	--
146	3	-0.55	--	--	16.6	--
151	4	-0.24	--	--	--	17.0
154	3	0.55	--	--	18.0	--
180	2	-1.02	--	--	16.0	--
193	4	0.16	--	17.5	--	--
196	1	-1.80	--	--	--	15.0
203	3	-0.86	--	16.2	--	--
204	4	0.16	--	--	--	17.5
212	3	0.55	--	--	--	18.0
220	4	0.16	--	--	17.5	--
234	4	0.08	--	--	17.4	--
236	2	1.33	--	--	19.0	--
247	0	4.08	--	--	--	22.5
255	4	-0.47	--	--	--	16.7

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Al (Aluminum) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	4	23	19	1	2. AA: direct, nitrous oxide	MPV =	32.4	
Minimum =	90.0	28.9	20.5	26.2	30.0	3. AA: graphite furnace	F-pseudosigma =	6.3	
Maximum =		34.8	227.0	50.0		4. ICP	N =	48	
Median =			35.0	31.0		6. ICP/MS	Uh =	37.6	
F-pseudosigma =			9.2	5.0		7. Ion chromatography	Lh =	29.0	


Lab	Rating	Z-value	2	3	4	6	7	Lab	Rating	Z-value	2	3	4	6	7
1	3	0.72	--	--	--	36.9	--	297	4	-0.04	--	--	32.1	--	--
3	4	0.12	--	--	33.1	--	--	324	0	9.10	90.0	--	--	--	--
5	NR	--	--	--	< 30	--	--	330	3	-0.97	--	--	--	26.2	--
11	0	5.94	--	--	70.0	--	--	332	0	2.78	--	--	--	50.0	--
19	1	1.76	--	--	43.5	--	--								
23	3	0.86	--	--	37.8	--	--								
25	NR	--	--	--	< 52	--	--								
32	4	0.29	--	--	--	34.2	--								
42	4	0.40	--	--	--	34.9	--								
48	3	0.58	--	--	--	36.0	--								
50	4	-0.13	--	--	--	31.5	--								
57	3	-0.53	--	--	29.0	--	--								
59	NR	--	--	--	--	< 50	--								
69	3	-0.55	--	28.9	--	--	--								
70	4	-0.17	--	--	--	31.3	--								
81	0	4.84	--	--	63.0	--	--								
89	4	0.39	--	34.8	--	--	--								
97	4	0.29	--	34.2	--	--	--								
105	3	-0.73	--	--	--	27.7	--								
110	3	0.94	--	--	38.3	--	--								
113	4	-0.17	--	--	31.3	--	--								
127	0	10.47	--	--	98.7	--	--								
134	4	0.42	--	--	35.0	--	--								
138	3	-0.61	--	--	28.5	--	--								
142	NR	--	--	--	< 50	--	--								
145	NR	--	--	--	< 179	--	--								
146	NR	--	--	--	< 54.7	--	--								
147	3	0.69	--	--	--	36.7	--								
149	4	0.26	--	34.0	--	--	--								
151	3	-0.53	--	--	--	29.0	--								
180	3	0.78	--	--	37.3	--	--								
191	3	-0.91	--	--	--	26.6	--								
196	4	-0.10	--	--	--	31.7	--								
203	2	-1.16	--	--	25.0	--	--								
204	3	-0.83	--	--	--	27.1	--								
209	4	-0.37	--	--	--	--	30.0								
212	0	6.95	--	--	76.4	--	--								
220	4	-0.50	--	--	29.2	--	--								
227	4	0.04	--	--	32.6	--	--								
234	4	-0.47	--	--	29.4	--	--								
236	2	1.21	--	--	40.0	--	--								
247	3	-0.84	--	--	--	27.0	--								
254	2	1.11	--	--	39.4	--	--								
255	4	-0.50	--	--	--	29.2	--								
259	4	-0.21	--	--	--	31.0	--								
265	3	-0.69	--	--	--	28.0	--								
270	2	-1.21	--	--	24.7	--	--								
273	1	-1.87	--	--	20.5	--	--								
284	0	30.71	--	--	227.0	--	--								
296	4	-0.36	--	--	--	30.1	--								

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: As (Arsenic) Concentration Unit : mg/L



ANALYTICAL METHOD CODE





SUMMARY									
N =	26	11	22	4	3. AA: graphite furnace	MPV =	26.1		
Minimum =	16.0	24.8	24.8	23.5	4. ICP	F-pseudostigma =	1.3		
Maximum =	29.8	40.0	30.0	28.0	6. ICP/MS	N =	63		
Median =	25.4	25.7	26.6		11. AA: hydride NaBH4	Uh =	27.0		
F-pseudostigma =	2.0	1.7	0.7			Lh =	25.2		

Lab	Rating	Z-value	3	4	6	11na	Lab	Rating	Z-value	3	4	6	11na
1	3	-0.77	25.1	--	--	--	212	3	0.56	--	--	26.9	--
3	4	-0.33	--	25.7	--	--	220	2	-1.33	24.3	--	--	--
5	4	0.19	26.4	--	--	--	234	1	1.68	28.4	--	--	--
10	2	1.38	--	--	--	28.0	236	0	10.27	--	40.0	--	--
11	4	-0.10	--	26.0	--	--	247	4	0.00	--	--	26.1	--
12	3	0.64	27.0	--	--	--	255	4	-0.25	--	--	25.8	--
13	2	-1.14	24.6	--	--	--	259	2	1.01	--	--	27.5	--
19	4	-0.47	--	25.5	--	--	265	3	0.64	--	--	27.0	--
23	3	-0.77	--	25.1	--	--	284	0	2.12	--	29.0	--	--
25	0	-7.52	16.0	--	--	--	296	4	-0.25	--	--	25.8	--
26	4	-0.47	--	--	--	25.5	297	3	-0.61	--	25.3	--	--
32	4	0.19	--	--	26.4	--	305	0	2.86	--	--	30.0	--
42	3	0.64	--	--	27.0	--	307	0	-3.37	21.6	--	--	--
46	4	0.12	26.3	--	--	--	330	4	0.12	--	--	26.3	--
48	3	-0.55	--	--	25.4	--							
50	4	0.49	--	--	26.8	--							
57	4	0.04	--	26.2	--	--							
59	4	0.49	--	--	26.8	--							
69	2	-1.14	24.6	--	--	--							
70	0	2.71	29.8	--	--	--							
76	2	1.04	--	--	27.5	--							
81	0	-2.33	23.0	--	--	--							
86	1	-1.96	--	--	--	23.5							
87	4	0.19	--	--	--	26.4							
89	3	-0.84	25.0	--	--	--							
93	3	-0.99	--	24.8	--	--							
96	2	-1.22	24.5	--	--	--							
97	3	0.79	27.2	--	--	--							
105	3	-0.62	--	--	25.3	--							
109	3	-0.65	25.3	--	--	--							
113	0	2.34	29.3	--	--	--							
126	4	-0.10	26.0	--	--	--							
127	0	-3.29	21.7	--	--	--							
131	3	-0.55	--	25.4	--	--							
134	4	0.27	26.5	--	--	--							
138	3	-0.99	--	--	24.8	--							
142	3	0.93	--	--	27.4	--							
144	2	-1.36	24.3	--	--	--							
145	NR	--	--	< 39	--	--							
146	0	2.49	--	29.5	--	--							
147	4	-0.10	--	--	26.0	--							
149	3	0.64	27.0	--	--	--							
151	4	-0.10	--	--	26.0	--							
154	1	1.53	28.2	--	--	--							
190	4	-0.40	25.6	--	--	--							
191	2	1.38	--	--	28.0	--							
193	1	-1.59	24.0	--	--	--							
196	4	0.49	--	--	26.8	--							
203	4	0.19	26.4	--	--	--							
204	4	0.04	--	--	26.2	--							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: B (Boron) Concentration Unit : mg/L




SUMMARY					
N =	24	9	4. ICP	MPV =	37.5
Minimum =	21.0	32.6	6. ICP/MS	F-pseudosigma =	4.9
Maximum =	390.0	40.6		N =	33
Median =	37.6	37.0		Uh =	40.6
F-pseudosigma =	6.0	3.7		Lh =	34.0


Lab	Rating	Z-value	4	6
1	4	0.04	37.7	--
3	4	0.00	37.5	--
5	4	-0.43	35.4	--
11	1	-1.53	30.0	--
24	0	4.19	58.0	--
25	0	-3.37	21.0	--
26	4	-0.14	36.8	--
32	3	-0.51	--	35.0
42	3	0.63	--	40.6
46	4	-0.02	37.4	--
48	4	-0.10	--	37.0
50	3	0.63	--	40.6
57	0	4.21	58.1	--
70	NR	--	< 100	--
76	4	0.14	--	38.2
86	3	0.61	40.5	--
105	NR	--	< 200	--
127	3	0.78	41.3	--
131	4	0.14	38.2	--
134	4	0.10	38.0	--
138	3	-0.74	33.9	--
142	NR	--	< 50	--
145	3	0.51	40.0	--
180	2	1.37	44.2	--
191	3	-0.78	--	33.7
212	4	-0.06	37.2	--
234	2	1.49	44.8	--
236	4	-0.31	36.0	--
247	1	1.94	47.0	--
255	3	-0.84	33.4	--
259	4	0.31	--	39.0
265	3	-0.72	--	34.0
270	3	-0.92	33.0	--
273	3	-0.84	33.4	--
284	0	72.05	390.0	--
330	3	-1.00	--	32.6

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Ba (Barium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																
N =	1	0	5	36	21	1. AA: direct, air	MPV =	70.4								
Minimum =	88.0		50.0	63.0	64.4	2. AA: direct, nitrous oxide	F-pseudosigma =	3.3								
Maximum =			86.6	81.0	75.2	3. AA: graphite furnace	Rating criterion =	3.5								
Median =				69.7	72.4	4. ICP	N =	63								
F-pseudosigma =				2.8	2.7	6. ICP/MS	Uh =	72.8								
							Lh =	68.4								
Lab	Rating	Z-value	1	2	3	4	6	Lab	Rating	Z-value	1	2	3	4	6	
1	4	0.01	--	--	--	70.4	--	234	4	-0.05	--	--	--	70.2	--	
3	1	-1.95	--	--	--	63.5	--	236	4	-0.39	--	--	--	69.0	--	
5	3	-0.99	--	--	--	66.9	--	247	4	-0.22	--	--	--	--	69.6	
11	0	-2.09	--	--	--	63.0	--	259	3	0.78	--	--	--	--	73.1	
13	4	-0.02	--	--	--	70.3	--	265	4	-0.11	--	--	--	--	70.0	
19	2	-1.10	--	--	--	66.5	--	268	2	1.34	--	--	--	--	75.1	
23	3	-0.62	--	--	--	68.2	--	270	3	0.58	--	--	--	72.4	--	
24	1	1.60	--	--	--	76.0	--	273	2	-1.04	--	--	--	66.7	--	
25	1	-1.53	--	--	--	65.0	--	277	3	-0.53	--	--	--	68.5	--	
26	4	0.29	--	--	--	71.4	--	284	4	0.18	--	--	--	71.0	--	
32	4	-0.33	--	--	--	--	69.2	296	2	1.26	--	--	--	--	74.8	
42	3	0.58	--	--	--	--	72.4	297	4	0.35	--	--	--	71.6	--	
46	4	-0.13	--	--	--	69.9	--	305	3	0.74	--	--	--	--	73.0	
48	2	1.03	--	--	--	--	74.0	330	3	0.55	--	--	--	--	72.3	
50	2	1.37	--	--	--	--	75.2									
57	4	-0.25	--	--	--	69.5	--									
59	3	0.58	--	--	--	--	72.4									
69	0	4.61	--	--	86.6	--	--									
70	4	0.46	--	--	--	--	72.0									
76	4	0.09	--	--	--	--	70.7									
81	3	-0.67	--	--	--	68.0	--									
83	3	-0.93	--	--	--	67.1	--									
86	4	-0.16	--	--	--	69.8	--									
87	2	1.37	--	--	75.2	--	--									
89	0	-5.79	--	--	50.0	--	--									
96	NR	--	--	< 100	--	--	--									
97	4	0.41	--	--	71.8	--	--									
105	3	0.63	--	--	--	--	72.6									
113	1	-1.67	--	--	--	64.5	--									
121	3	-0.67	--	--	--	68.0	--									
127	1	-1.55	--	--	--	64.9	--									
131	4	-0.28	--	--	--	69.4	--									
134	4	0.43	--	--	--	71.9	--									
138	4	-0.13	--	--	--	69.9	--									
140	0	5.01	88.0	--	--	--	--									
142	3	0.80	--	--	--	--	73.2									
145	4	-0.11	--	--	--	70.0	--									
146	2	1.17	--	--	--	74.5	--									
149	0	2.74	--	--	80.0	--	--									
151	4	-0.11	--	--	--	--	70.0									
154	4	0.38	--	--	--	71.7	--									
180	1	-1.61	--	--	--	64.7	--									
191	1	-1.70	--	--	--	--	64.4									
193	3	-0.96	--	--	--	67.0	--									
196	4	0.01	--	--	--	--	70.4									
203	4	-0.11	--	--	--	70.0	--									
204	2	1.32	--	--	--	--	75.0									
212	2	1.23	--	--	--	--	74.7									
218	0	3.02	--	--	--	81.0	--									
220	4	0.00	--	--	--	70.4	--									

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Be (Beryllium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	8	28	19	2. AA: direct, nitrous oxide	MPV =	12.9		
Minimum =	15.5	11.0	11.8	11.4	3. AA: graphite furnace	F-pseudostigma =	0.5		
Maximum =		14.3	14.0	15.2	4. ICP	Rating criterion =	0.6		
Median =		12.7	12.9	13.0	6. ICP/MS	N =	56		
F-pseudostigma =		0.7	0.5	0.6		Uh =	13.2		
						Lh =	12.5		


Lab	Rating	Z-value	2	3	4	6	Lab	Rating	Z-value	2	3	4	6
1	4	0.12	--	--	--	13.0	265	3	-0.68	--	--	--	12.5
3	2	-1.18	--	--	12.2	--	284	1	-1.52	--	--	12.0	--
5	3	-0.85	--	--	12.4	--	296	0	3.82	--	--	--	15.2
11	1	-1.52	--	--	12.0	--	297	3	-0.68	--	--	12.5	--
13	3	-0.52	--	--	12.6	--	305	4	-0.12	--	--	--	12.8
19	4	0.32	--	--	13.1	--	330	4	0.32	--	--	--	13.1
23	3	0.48	--	--	13.2	--							
25	4	0.15	--	--	13.0	--							
26	1	1.82	--	--	14.0	--							
32	4	-0.18	--	--	--	12.8							
42	3	0.65	--	--	--	13.3							
46	4	0.15	--	--	13.0	--							
48	1	-1.52	--	--	--	12.0							
50	3	0.98	--	--	--	13.5							
57	4	0.15	--	--	13.0	--							
59	0	2.82	--	--	--	14.6							
69	1	-1.68	--	11.9	--	--							
70	0	2.48	--	--	--	14.4							
81	1	-1.52	--	--	12.0	--							
83	3	-0.68	--	--	12.5	--							
86	4	-0.18	--	--	12.8	--							
89	0	2.32	--	14.3	--	--							
93	4	-0.35	--	--	12.7	--							
96	2	-1.02	--	12.3	--	--							
97	3	-0.52	--	12.6	--	--							
105	4	-0.35	--	--	--	12.7							
113	4	0.15	--	--	13.0	--							
114	0	4.32	15.5	--	--	--							
126	4	-0.35	--	12.7	--	--							
127	3	-0.85	--	--	12.4	--							
131	1	-1.68	--	--	11.9	--							
134	4	0.15	--	--	13.0	--							
138	3	-0.52	--	--	12.6	--							
142	3	-0.52	--	--	--	12.6							
144	4	-0.35	--	12.7	--	--							
145	4	0.15	--	--	13.0	--							
146	3	0.98	--	--	13.5	--							
149	0	-3.18	--	11.0	--	--							
151	4	0.15	--	--	--	13.0							
154	3	0.48	--	--	13.2	--							
180	1	-1.85	--	--	11.8	--							
191	0	2.15	--	--	--	14.2							
193	3	0.98	--	13.5	--	--							
196	3	0.82	--	--	--	13.4							
204	0	-2.52	--	--	--	11.4							
212	4	0.32	--	--	--	13.1							
220	4	0.45	--	--	13.2	--							
234	4	0.32	--	--	13.1	--							
236	4	0.15	--	--	13.0	--							
247	2	-1.02	--	--	--	12.3							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Cd (Cadmium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	3	18	27	23	1. AA: direct, air	MPV =	17.5		
Minimum =	17.8	14.1	14.9	16.3	3. AA: graphite furnace	F-pseudosigma =	0.9		
Maximum =	31.5	19.0	22.6	19.5	4. ICP	N =	71		
Median =		16.9	17.3	17.8	6. ICP/MS	Uh =	18.1		
F-pseudosigma =		1.3	0.9	0.7		Lh =	16.9		

Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	0.24	--	--	--	17.7	196	3	-0.76	--	--	--	16.8
3	2	-1.40	--	--	16.2	--	203	3	-0.76	--	16.8	--	--
5	4	-0.32	--	17.2	--	--	204	4	0.32	--	--	--	17.8
10	2	-1.40	--	16.2	--	--	212	2	1.08	--	--	--	18.5
11	3	-0.54	--	--	17.0	--	220	4	0.22	--	17.7	--	--
12	2	-1.08	--	16.5	--	--	227	4	0.00	--	--	17.5	--
13	3	0.97	--	--	18.4	--	234	4	0.22	--	--	17.7	--
19	4	-0.22	--	--	17.3	--	236	3	0.54	--	--	18.0	--
23	4	-0.22	--	--	17.3	--	247	2	1.08	--	--	--	18.5
24	1	1.62	--	--	19.0	--	255	4	0.32	--	--	--	17.8
25	3	-0.54	--	--	17.0	--	259	4	0.32	--	--	--	17.8
32	4	0.11	--	--	--	17.6	265	4	0.00	--	--	--	17.5
42	0	2.16	--	--	--	19.5	270	0	5.50	--	--	22.6	--
46	3	-0.97	--	16.6	--	--	273	0	-2.81	--	--	14.9	--
48	3	-0.65	--	--	--	16.9	277	0	-2.16	--	--	15.5	--
50	4	0.32	--	--	--	17.8	284	3	-0.54	--	--	17.0	--
57	2	-1.30	--	--	16.3	--	296	0	2.05	--	--	--	19.4
59	3	0.65	--	--	--	18.1	297	4	-0.33	--	--	17.2	--
69	1	-1.73	--	15.9	--	--	305	3	0.98	--	--	--	18.4
70	4	0.00	--	--	--	17.5	307	3	-0.54	--	17.0	--	--
81	0	-2.70	--	15.0	--	--	324	0	4.86	22.0	--	--	--
83	3	-0.54	--	--	17.0	--	330	4	0.11	--	--	--	17.6
86	3	0.76	--	--	18.2	--							
87	4	0.22	--	--	17.7	--							
89	4	0.43	--	17.9	--	--							
93	3	-0.86	--	--	16.7	--							
96	3	0.65	--	18.1	--	--							
97	4	-0.32	--	17.2	--	--							
105	2	-1.30	--	--	--	16.3							
113	4	0.22	--	--	17.7	--							
114	0	15.11	31.5	--	--	--							
121	1	1.62	--	--	19.0	--							
126	0	-2.70	--	15.0	--	--							
127	2	-1.30	--	--	16.3	--							
131	1	-1.83	--	--	15.8	--							
134	4	-0.32	--	--	17.2	--							
138	4	-0.43	--	--	--	17.1							
140	4	0.32	17.8	--	--	--							
142	3	0.97	--	--	--	18.4							
144	1	-1.73	--	15.9	--	--							
145	3	0.54	--	--	18.0	--							
146	2	1.08	--	--	18.5	--							
147	4	-0.22	--	--	--	17.3							
149	3	-0.54	--	17.0	--	--							
151	3	0.54	--	--	--	18.0							
154	2	1.19	--	18.6	--	--							
180	NR	--	--	--	< 19.7	--							
183	0	-3.67	--	14.1	--	--							
191	2	1.30	--	--	--	18.7							
193	1	1.62	--	19.0	--	--							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Co (Cobalt) Concentration Unit : mg/L




SUMMARY					
N =	5	25	18	3. AA: graphite furnace	MPV = 12.5
Minimum =	10.0	5.0	10.8	4. ICP	F-pseudosigma = 0.6
Maximum =	13.2	15.4	13.7	6. ICP/MS	N = 48
Median =		12.5	12.6		Uh = 12.8
F-pseudosigma =		0.7	0.4		Lh = 12.0


Lab	Rating	Z-value	3	4	6
1	4	0.05	--	--	12.5
3	4	0.16	--	12.6	--
5	1	-1.90	11.3	--	--
11	3	-0.79	--	12.0	--
13	3	-0.95	--	11.9	--
24	3	0.79	--	13.0	--
25	3	-0.79	--	12.0	--
26	3	-0.63	--	12.1	--
32	4	0.16	--	--	12.6
42	4	0.16	--	--	12.6
46	4	-0.48	--	12.2	--
48	0	-2.70	--	--	10.8
50	4	0.32	--	--	12.7
57	3	0.63	--	12.9	--
59	4	0.00	--	--	12.5
70	0	-2.38	--	--	11.0
81	3	0.79	--	13.0	--
86	0	-4.13	--	9.9	--
89	0	-3.97	10.0	--	--
96	2	1.11	13.2	--	--
97	1	-1.75	11.4	--	--
105	3	-0.95	--	--	11.9
121	0	-3.97	--	10.0	--
127	4	-0.16	12.4	--	--
131	0	4.60	--	15.4	--
134	4	-0.08	--	12.5	--
142	4	0.00	--	--	12.5
145	0	3.97	--	15.0	--
146	4	0.32	--	12.7	--
154	3	0.79	--	13.0	--
180	0	-2.06	--	11.2	--
191	4	0.32	--	--	12.7
196	4	-0.48	--	--	12.2
212	2	1.11	--	--	13.2
234	4	0.00	--	12.5	--
236	0	-11.90	--	5.0	--
247	4	0.48	--	--	12.8
254	4	0.16	--	12.6	--
259	4	0.48	--	--	12.8
265	4	0.00	--	--	12.5
268	4	-0.48	--	--	12.2
270	3	-0.73	--	12.0	--
273	0	-4.29	--	9.8	--
277	4	0.48	--	12.8	--
284	0	3.97	--	15.0	--
296	1	1.90	--	--	13.7
297	4	-0.32	--	12.3	--
330	4	0.48	--	--	12.8

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Cr (Chromium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	2	18	26	24	1. AA: direct, air	MPV =	34.6	
Minimum =	51.0	39.0	11.0	30.0	31.5	2. AA: direct, nitrous oxide	F-pseudosigma =	1.7	
Maximum =		62.5	37.0	38.7	39.9	3. AA: graphite furnace	N =	71	
Median =			35.1	34.2	34.7	4. ICP	Uh =	35.7	
F-pseudosigma =			2.4	1.3	1.3	6. ICP/MS	Lh =	33.4	


Lab	Rating	Z-value	1	2	3	4	6	Lab	Rating	Z-value	1	2	3	4	6
1	4	0.47	--	--	--	--	35.4	204	2	-1.29	--	--	--	--	32.4
3	4	-0.35	--	--	--	34.0	--	212	2	1.11	--	--	--	--	36.5
5	4	-0.12	--	--	34.4	--	--	220	2	1.34	--	--	36.9	--	--
10	4	0.41	--	--	35.3	--	--	234	4	-0.23	--	--	--	34.2	--
11	3	-0.94	--	--	--	33.0	--	236	3	-0.94	--	--	--	33.0	--
13	4	0.06	--	--	--	34.7	--	247	2	-1.45	--	--	--	--	32.1
19	4	0.00	--	--	--	34.6	--	255	4	0.00	--	--	--	--	34.6
23	4	-0.18	--	--	--	34.3	--	259	4	0.35	--	--	--	--	35.2
24	3	0.82	--	--	--	36.0	--	265	4	0.23	--	--	--	--	35.0
25	0	-13.84	--	--	11.0	--	--	268	3	-0.70	--	--	--	--	33.4
32	4	0.12	--	--	--	--	34.8	270	4	-0.04	--	--	--	34.5	--
42	4	-0.18	--	--	--	--	34.3	273	0	-2.52	--	--	--	30.3	--
46	3	1.00	--	--	36.3	--	--	277	0	-2.70	--	--	--	30.0	--
48	4	-0.35	--	--	--	--	34.0	284	3	0.82	--	--	--	36.0	--
50	1	1.82	--	--	--	--	37.7	296	1	1.70	--	--	--	--	37.5
57	2	-1.23	--	--	--	32.5	--	297	4	-0.34	--	--	--	34.0	--
59	4	0.23	--	--	--	--	35.0	305	3	1.00	--	--	--	--	36.3
69	2	1.29	--	--	36.8	--	--	307	3	-0.94	--	--	33.0	--	--
70	1	-1.82	--	--	--	--	31.5	324	0	9.62	51.0	--	--	--	--
76	4	0.39	--	--	--	--	35.3	330	3	-0.65	--	--	--	--	33.5
81	4	-0.35	--	--	34.0	--	--	332	0	3.11	--	--	--	--	39.9
83	4	0.06	--	--	--	34.7	--								
86	0	2.40	--	--	--	38.7	--								
87	3	-0.70	--	--	--	33.4	--								
89	3	-0.59	--	--	33.6	--	--								
93	2	-1.06	--	--	--	32.8	--								
96	2	-1.11	--	--	32.7	--	--								
97	2	1.41	--	--	37.0	--	--								
105	4	-0.12	--	--	--	--	34.4								
114	0	16.36	--	62.5	--	--	--								
126	4	0.23	--	--	35.0	--	--								
127	0	2.29	--	--	--	38.5	--								
131	4	-0.23	--	--	--	34.2	--								
134	4	0.06	--	--	--	34.7	--								
138	4	-0.35	--	--	--	34.0	--								
140	0	2.58	--	39.0	--	--	--								
142	4	-0.06	--	--	--	--	34.5								
144	4	0.41	--	--	35.3	--	--								
145	1	-1.52	--	--	--	32.0	--								
146	4	0.18	--	--	--	34.9	--								
149	3	0.82	--	--	36.0	--	--								
151	3	-0.94	--	--	--	--	33.0								
154	4	0.47	--	--	--	35.4	--								
180	3	-0.70	--	--	--	33.4	--								
183	0	-2.93	--	--	29.6	--	--								
190	2	-1.35	--	--	32.3	--	--								
191	4	0.29	--	--	--	--	35.1								
193	2	1.29	--	--	36.8	--	--								
196	3	-0.53	--	--	--	--	33.7								
203	4	0.35	--	--	35.2	--	--								

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Cu (Copper) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	5	10	32	23	1. AA: direct, air	MPV =	22.0		
Minimum =	18.0	17.7	18.3	19.2	3. AA: graphite furnace	F-pseudostigma =	1.7		
Maximum =	52.0	24.5	153.0	24.6	4. ICP	N =	70		
Median =		22.1	22.3	22.0	6. ICP/MS	Uh =	23.4		
F-pseudostigma =		1.3	2.2	0.7		Lh =	21.1		


Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	-0.07	--	--	--	21.9	203	3	-0.59	--	--	21.0	--
3	4	-0.18	--	--	21.7	--	204	1	-1.64	--	--	--	19.2
5	4	-0.18	--	--	21.7	--	212	1	1.52	--	--	--	24.6
10	4	0.00	--	22.0	--	--	220	4	-0.08	--	21.9	--	--
11	2	-1.17	--	--	20.0	--	227	2	1.29	--	--	24.2	--
13	3	0.76	--	--	23.3	--	234	3	0.76	--	--	23.3	--
19	0	-2.17	--	--	18.3	--	236	2	1.17	--	--	24.0	--
23	4	-0.12	--	--	21.8	--	247	4	0.38	--	--	--	22.7
24	0	2.35	--	--	26.0	--	254	3	0.59	--	--	23.0	--
25	0	3.52	--	--	28.0	--	255	4	0.41	--	--	--	22.7
26	3	-0.82	--	--	20.6	--	259	4	0.18	--	--	--	22.3
32	4	-0.12	--	--	--	21.8	265	4	0.00	--	--	--	22.0
42	3	0.82	--	--	--	23.4	270	3	0.99	--	--	23.7	--
46	3	-0.82	--	20.6	--	--	273	0	2.35	--	--	26.0	--
48	3	-0.82	--	--	--	20.6	277	2	-1.06	--	--	20.2	--
50	4	0.23	--	--	--	22.4	284	3	-0.59	--	--	21.0	--
57	2	1.35	--	--	24.3	--	296	3	1.00	--	--	--	23.7
59	4	0.06	--	--	--	22.1	297	4	0.44	--	--	22.8	--
69	NR	--	< 50	--	--	--	305	4	0.00	--	--	--	22.0
70	2	-1.17	--	--	--	20.0	307	4	0.23	--	22.4	--	--
76	4	-0.06	--	--	--	21.9	324	0	17.60	52.0	--	--	--
81	1	-1.76	--	19.0	--	--	330	4	0.29	--	--	--	22.5
83	4	-0.47	--	--	21.2	--							
86	3	-0.53	--	--	21.1	--							
87	2	1.17	--	--	24.0	--							
89	0	-2.52	--	17.7	--	--							
93	0	2.17	--	--	25.7	--							
96	4	0.06	--	22.1	--	--							
97	4	0.23	--	22.4	--	--							
105	3	-0.53	--	--	--	21.1							
113	3	0.94	--	--	23.6	--							
114	1	-1.76	19.0	--	--	--							
126	3	0.59	23.0	--	--	--							
127	0	2.99	--	--	27.1	--							
131	0	76.83	--	--	153.0	--							
134	4	-0.06	--	--	21.9	--							
138	4	-0.18	--	--	21.7	--							
140	3	0.65	23.1	--	--	--							
142	2	-1.29	--	--	--	19.8							
145	NR	--	--	--	< 26	--							
146	2	-1.17	--	--	20.0	--							
147	4	-0.06	--	--	--	21.9							
149	0	-2.35	18.0	--	--	--							
151	2	-1.17	--	--	--	20.0							
154	4	-0.18	--	--	21.7	--							
180	4	-0.41	--	--	21.3	--							
183	2	1.47	--	24.5	--	--							
191	4	-0.12	--	--	--	21.8							
193	3	0.82	--	23.4	--	--							
196	4	0.12	--	--	--	22.2							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Fe (Iron) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	14	2	46	7	1	1. AA: direct, air	MPV =	61.7	
Minimum =	54.6	52.8	33.5	16.8	73.3	3. AA: graphite furnace	F-pseudosigma =	10.2	
Maximum =	81.0	53.9	80.0	80.0		4. ICP	N =	70	
Median =	65.2		61.6	55.5		6. ICP/MS	Uh =	70.0	
F-pseudosigma =	6.7		9.3	18.4		22fe: Color. ferricyanide	Lh =	56.3	


Lab	Rating	Z-value	1	3	4	6	22fe	Lab	Rating	Z-value	1	3	4	6	22fe
1	3	0.94	--	--	71.2	--	--	203	3	-0.96	--	--	52.0	--	--
3	2	1.04	--	--	72.3	--	--	204	3	-0.66	--	--	55.0	--	--
5	4	-0.21	--	--	59.6	--	--	212	2	1.24	--	--	74.3	--	--
10	4	0.13	63.0	--	--	--	--	220	3	-0.78	--	--	53.7	--	--
11	3	0.92	--	--	71.0	--	--	227	4	-0.09	--	--	60.8	--	--
13	3	-0.75	--	--	54.1	--	--	234	4	-0.19	--	--	59.8	--	--
19	2	1.36	--	--	75.5	--	--	236	3	-0.66	--	--	55.0	--	--
21	2	1.14	--	--	--	73.3	--	247	4	0.03	--	--	62.0	--	--
23	4	0.40	65.8	--	--	--	--	254	4	0.01	--	--	61.8	--	--
24	1	1.80	--	--	80.0	--	--	259	3	-0.61	--	--	55.5	--	--
25	3	0.62	--	--	68.0	--	--	265	4	0.03	--	--	62.0	--	--
26	4	-0.02	--	--	61.5	--	--	268	4	-0.17	--	--	60.0	--	--
42	4	0.48	--	--	66.6	--	--	270	0	-2.77	--	--	33.5	--	--
46	4	-0.16	--	--	60.1	--	--	273	3	-0.81	--	--	53.5	--	--
48	1	1.51	--	--	--	77.0	--	277	2	-1.02	--	--	51.3	--	--
50	3	-0.99	--	--	--	51.6	--	284	NR	--	--	--	< 50	--	--
57	4	-0.15	--	--	60.2	--	--	296	4	0.13	--	--	--	63.0	--
59	0	-2.26	--	--	--	38.7	--	297	3	0.91	--	--	70.9	--	--
69	3	-0.66	55.0	--	--	--	--	305	4	0.29	64.6	--	--	--	--
70	4	-0.26	--	--	59.1	--	--	307	4	-0.07	61.0	--	--	--	--
81	3	0.82	--	--	70.0	--	--	324	NR	--	< 100	--	--	--	--
83	4	-0.42	--	--	57.4	--	--	330	NR	--	--	--	< 1000	--	--
87	4	0.05	--	--	62.2	--	--	332	1	1.80	--	--	--	80.0	--
89	3	-0.88	--	52.8	--	--	--								
93	4	-0.31	--	--	58.6	--	--								
96	3	0.82	70.0	--	--	--	--								
97	3	-0.77	--	53.9	--	--	--								
105	4	0.23	--	--	64.0	--	--								
107	1	1.80	80.0	--	--	--	--								
109	3	-0.70	54.6	--	--	--	--								
113	4	-0.01	--	--	61.6	--	--								
114	4	0.08	62.5	--	--	--	--								
121	4	-0.17	--	--	60.0	--	--								
126	3	0.82	70.0	--	--	--	--								
127	3	-0.55	--	--	56.1	--	--								
131	4	-0.46	--	--	57.0	--	--								
134	3	0.82	--	--	70.0	--	--								
138	4	-0.09	--	--	60.8	--	--								
140	4	-0.07	61.0	--	--	--	--								
142	3	0.62	--	--	68.0	--	--								
145	2	1.31	--	--	75.0	--	--								
146	2	1.44	--	--	76.3	--	--								
147	2	1.21	--	--	74.0	--	--								
151	3	0.80	69.8	--	--	--	--								
154	3	0.82	--	--	70.0	--	--								
180	3	-0.53	--	--	56.3	--	--								
183	1	1.90	81.0	--	--	--	--								
190	1	1.80	80.0	--	--	--	--								
191	0	-4.42	--	--	--	16.8	--								
193	4	-0.01	--	--	61.6	--	--								

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: K (Potassium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	11	39	9	1	3	1. AA: direct, air	MPV =	1.26	
Minimum =	1.05	0.80	1.07	1.77	1.24	4. ICP	F-pseudosigma =	0.11	
Maximum =	1.72	2.72	4.11		1.32	6. ICP/MS	N =	63	
Median =	1.26	1.24	1.22			7. Ion chromatography	Uh =	1.35	
F-pseudosigma =	0.10	0.08	0.33			12. Flame emission	Lh =	1.20	

Lab	Rating	Z-value	1	4	6	7	12	Lab	Rating	Z-value	1	4	6	7	12
1	4	-0.15	1.24	--	--	--	--	265	3	-0.54	--	1.20	--	--	--
3	0	7.22	--	2.06	--	--	--	270	4	0.45	--	1.31	--	--	--
5	4	-0.09	--	1.25	--	--	--	273	3	-0.54	--	1.20	--	--	--
11	1	-1.89	--	1.05	--	--	--	277	3	0.90	--	1.36	--	--	--
13	4	0.45	--	1.31	--	--	--	279	0	4.15	1.72	--	--	--	--
19	4	0.09	--	1.27	--	--	--	284	3	-0.99	--	1.15	--	--	--
23	4	-0.36	--	1.22	--	--	--	296	3	-0.63	--	--	1.19	--	--
24	4	0.36	--	1.30	--	--	--	297	0	-4.16	--	0.80	--	--	--
25	1	-1.80	--	1.06	--	--	--	302	0	4.60	--	--	--	1.77	--
26	0	13.17	--	2.72	--	--	--	305	0	3.43	--	--	1.64	--	--
32	4	-0.36	--	--	1.22	--	--	309	0	6.68	--	--	2.00	--	--
42	4	-0.45	--	1.21	--	--	--	324	4	0.00	1.26	--	--	--	--
46	4	0.00	--	1.26	--	--	--	330	0	-2.35	--	< 1	--	--	--
48	3	-0.54	--	--	1.20	--	--	332	0	25.72	--	--	4.11	--	--
50	4	0.27	--	1.29	--	--	--								
57	0	7.13	--	2.05	--	--	--								
59	1	-1.71	--	--	1.07	--	--								
69	4	-0.18	--	--	--	--	1.24								
70	4	-0.45	--	1.21	--	--	--								
81	3	0.90	--	1.36	--	--	--								
86	4	0.18	--	1.28	--	--	--								
87	1	-1.89	1.05	--	--	--	--								
89	4	0.00	--	--	--	--	1.26								
93	0	-3.61	--	0.86	--	--	--								
105	4	-0.27	--	1.23	--	--	--								
107	4	0.27	1.29	--	--	--	--								
109	2	1.17	1.39	--	--	--	--								
110	3	0.92	1.36	--	--	--	--								
113	3	-0.99	--	1.15	--	--	--								
127	4	0.00	1.26	--	--	--	--								
131	4	0.36	--	1.30	--	--	--								
134	3	-0.54	1.20	--	--	--	--								
138	4	-0.36	--	1.22	--	--	--								
140	4	-0.09	1.25	--	--	--	--								
142	4	-0.45	--	1.21	--	--	--								
145	2	-1.26	--	1.12	--	--	--								
146	4	-0.45	--	1.21	--	--	--								
180	0	-2.62	--	0.97	--	--	--								
191	3	-0.99	--	--	1.15	--	--								
193	1	1.53	1.43	--	--	--	--								
203	3	0.54	--	--	--	--	1.32								
204	4	-0.27	--	1.23	--	--	--								
209	0	6.05	--	1.93	--	--	--								
212	0	2.08	--	1.49	--	--	--								
218	3	0.71	--	1.34	--	--	--								
234	4	-0.18	--	1.24	--	--	--								
236	0	3.07	--	1.60	--	--	--								
247	4	0.18	--	1.28	--	--	--								
254	3	-0.54	--	1.20	--	--	--								
259	4	0.00	--	--	1.26	--	--								

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Li (Lithium) Concentration Unit : mg/L



SUMMARY

N =	1	16	11
Minimum =	18.3	13.3	15.4
Maximum =		42.0	18.6
Median =		18.5	17.0
F-pseudosigma =		2.0	0.7


- 1. AA: direct, air
- 4. ICP
- 6. ICP/MS

MPV =	17.8
F-pseudosigma =	1.6
N =	28
Uh =	18.8
Lh =	16.7


Lab	Rating	Z-value	1	4	6
1	4	-0.29	--	--	17.3
3	0	2.41	--	21.5	--
5	3	-0.55	--	16.9	--
11	2	1.45	--	20.0	--
25	3	0.80	--	19.0	--
26	3	1.00	--	19.3	--
32	4	0.16	--	--	18.0
42	4	0.22	--	18.1	--
50	3	0.55	--	--	18.6
57	0	-2.86	--	13.3	--
59	1	-1.51	--	--	15.4
105	0	15.58	--	42.0	--
109	4	0.38	18.3	--	--
127	4	0.10	--	17.9	--
131	3	-0.87	--	16.4	--
134	4	0.46	--	18.5	--
142	2	1.19	--	19.6	--
145	NR	--	--	< 20	--
151	4	-0.48	--	--	17.0
191	4	-0.29	--	--	17.3
196	2	-1.25	--	--	15.8
220	4	-0.22	--	17.4	--
234	4	0.48	--	18.5	--
236	2	-1.12	--	--	--
247	3	-0.87	--	--	16.4
265	4	-0.48	--	--	17.0
268	4	-0.10	--	--	17.6
270	3	-0.80	--	--	16.5
273	1	2.02	--	20.9	--

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Mg (Magnesium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	11	51	10	1	1. AA: direct, air	MPV =	1.51		
Minimum =	1.11	1.00	1.29	2.28	4. ICP	F-pseudosigma =	0.07		
Maximum =	1.70	3.05	1.62		6. ICP/MS	Rating criterion =	0.08		
Median =	1.52	1.50	1.51		7. Ion chromatography	N =	73		
F-pseudosigma =	0.06	0.07	0.05			Uh =	1.55		
						Lh =	1.46		


Lab	Rating	Z-value	1	4	6	7	Lab	Rating	Z-value	1	4	6	7
1	4	-0.24	--	1.49	--	--	218	0	4.30	--	1.83	--	--
3	2	-1.03	--	1.43	--	--	220	4	-0.24	--	1.49	--	--
5	2	-1.30	--	1.41	--	--	227	4	-0.50	--	1.47	--	--
11	2	-1.43	--	1.40	--	--	234	2	1.22	--	1.60	--	--
12	0	3.87	--	1.80	--	--	236	2	-1.43	--	1.40	--	--
13	3	0.69	--	1.56	--	--	247	4	0.42	--	1.54	--	--
19	4	-0.37	--	1.48	--	--	254	4	0.16	--	1.52	--	--
23	4	-0.11	1.50	--	--	--	255	3	0.95	--	1.58	--	--
24	2	-1.43	--	1.40	--	--	259	3	-0.64	--	--	1.46	--
25	1	-1.83	--	1.37	--	--	265	3	0.56	--	1.55	--	--
26	4	-0.11	--	1.50	--	--	268	2	-1.03	--	--	1.43	--
32	2	1.49	--	--	1.62	--	270	0	20.45	--	3.05	--	--
42	3	-0.77	--	1.45	--	--	273	4	-0.11	--	1.50	--	--
46	4	-0.37	--	1.48	--	--	277	4	0.16	--	1.52	--	--
48	4	0.03	--	--	1.51	--	279	0	2.55	1.70	--	--	--
50	4	0.03	--	1.51	--	--	284	4	-0.24	--	1.49	--	--
57	3	0.56	--	1.55	--	--	296	4	0.03	--	--	1.51	--
59	0	-2.89	--	--	1.29	--	297	1	-1.83	--	1.37	--	--
69	NR	--	< 2	--	--	--	302	0	10.24	--	--	--	2.28
70	3	0.56	--	1.55	--	--	305	4	-0.12	--	--	1.50	--
76	3	0.54	--	--	1.55	--	307	4	-0.11	1.50	--	--	--
81	2	0.95	--	1.58	--	--	309	0	-6.74	--	1.00	--	--
83	3	-0.64	--	1.46	--	--	324	0	-5.28	1.11	--	--	--
84	4	0.16	1.52	--	--	--	330	0	-6.74	--	< 1	--	--
86	3	0.69	--	1.56	--	--	332	4	0.16	--	--	1.52	--
87	2	-1.43	1.40	--	--	--							
89	2	1.22	1.60	--	--	--							
93	4	0.03	--	1.51	--	--							
97	4	0.16	1.52	--	--	--							
105	4	0.16	--	1.52	--	--							
107	3	0.56	1.55	--	--	--							
109	3	0.95	1.58	--	--	--							
110	4	0.00	--	1.51	--	--							
113	4	-0.38	--	1.48	--	--							
121	4	-0.11	--	1.50	--	--							
127	3	-0.90	--	1.44	--	--							
131	0	-2.76	--	1.30	--	--							
134	4	-0.19	--	1.49	--	--							
138	4	0.03	--	1.51	--	--							
140	4	-0.37	1.48	--	--	--							
142	3	0.69	--	1.56	--	--							
145	4	-0.24	--	1.49	--	--							
146	4	0.16	--	1.52	--	--							
180	1	-1.83	--	1.37	--	--							
191	4	0.29	--	--	1.53	--							
193	4	0.16	--	1.52	--	--							
203	4	-0.24	--	1.49	--	--							
204	3	0.69	--	1.56	--	--							
209	0	-2.10	--	1.35	--	--							
212	3	0.69	--	1.56	--	--							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Mn (Manganese) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	10	3	40	18	1. AA: direct, air	MPV =	37.4		
Minimum =	32.0	29.8	13.0	31.3	3. AA: graphite furnace	F-pseudosigma =	2.3		
Maximum =	49.0	40.0	40.5	42.0	4. ICP	N =	71		
Median =	39.5		37.5	36.6	6. ICP/MS	Uh =	38.1		
F-pseudosigma =	4.4		2.2	2.1		Lh =	35.0		

Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	0.16	--	--	37.8	--	204	3	-0.80	--	--	--	35.6
3	2	-1.46	--	--	34.1	--	212	2	1.11	--	--	--	39.9
5	3	-0.62	--	--	36.0	--	220	4	-0.11	--	--	37.2	--
10	1	1.59	41.0	--	--	--	227	4	0.09	--	--	37.6	--
11	4	-0.18	--	--	37.0	--	234	4	0.09	--	--	37.6	--
13	2	1.37	--	--	40.5	--	236	2	-1.06	--	--	35.0	--
19	4	0.31	--	--	38.1	--	247	2	-1.42	--	--	--	34.2
23	4	0.09	--	--	37.6	--	254	4	0.22	--	--	37.9	--
24	2	1.15	--	--	40.0	--	255	4	-0.18	--	--	--	37.0
25	2	-1.06	--	--	35.0	--	259	3	0.66	--	--	--	38.9
26	4	0.44	--	--	38.4	--	265	4	0.27	--	--	38.0	--
32	4	0.04	--	--	--	37.5	268	4	-0.40	--	--	--	36.5
42	3	-0.53	--	--	36.2	--	273	3	-0.88	--	--	35.4	--
46	4	0.13	--	--	37.7	--	277	0	-3.27	--	--	30.0	--
48	2	-1.50	--	--	--	34.0	284	0	-10.79	--	--	13.0	--
50	4	0.18	--	--	--	37.8	296	2	1.11	--	--	--	39.9
57	3	-0.84	--	--	35.5	--	297	4	0.05	--	--	37.5	--
59	1	2.03	--	--	--	42.0	307	0	5.13	49.0	--	--	--
69	3	0.71	39.0	--	--	--	324	0	-2.39	32.0	--	--	--
70	1	-2.03	--	--	--	32.8	330	4	-0.35	--	--	--	36.6
81	4	0.27	--	--	38.0	--	332	0	-2.70	--	--	--	31.3
83	4	0.00	--	--	37.4	--							
84	0	-3.36	--	29.8	--	--							
86	4	0.18	--	--	37.8	--							
87	0	-7.17	--	--	21.2	--							
89	2	1.15	--	40.0	--	--							
93	2	-1.28	--	--	34.5	--							
96	1	2.03	42.0	--	--	--							
97	4	0.44	--	38.4	--	--							
105	3	-0.80	--	--	--	35.6							
107	2	1.15	40.0	--	--	--							
109	1	-1.83	33.3	--	--	--							
113	3	0.71	--	--	39.0	--							
114	3	-0.62	36.0	--	--	--							
121	4	0.27	--	--	38.0	--							
126	2	1.15	40.0	--	--	--							
127	2	-1.37	--	--	34.3	--							
131	2	-1.06	--	--	35.0	--							
134	4	0.46	--	--	38.4	--							
138	4	-0.27	--	--	36.8	--							
140	2	-1.06	35.0	--	--	--							
142	0	-2.83	--	--	31.0	--							
145	4	0.27	--	--	38.0	--							
146	4	0.27	--	--	38.0	--							
151	2	-1.06	--	--	--	35.0							
154	4	0.27	--	--	38.0	--							
180	2	-1.02	--	--	35.1	--							
191	4	0.00	--	--	--	37.4							
196	4	-0.44	--	--	--	36.4							
203	3	-0.62	--	--	36.0	--							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Mo (Molybdenum) Concentration Unit : mg/L



SUMMARY

N =	4	19	20
Minimum =	16.0	12.0	16.9
Maximum =	22.3	23.8	20.6
Median =		18.7	19.4
F-pseudosigma =		1.6	1.1


- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	18.9
F-pseudosigma =	1.6
N =	43
Uh =	19.6
Lh =	17.5

Lab	Rating	Z-value	3	4	6
1	4	0.33	--	--	19.4
3	4	-0.19	--	18.6	--
5	1	-1.54	--	16.5	--
11	2	-1.22	--	17.0	--
13	NR	--	--	< 50	--
23	4	0.19	--	19.2	--
24	3	0.70	--	20.0	--
26	4	0.00	--	18.9	--
32	4	-0.19	--	--	18.6
42	4	0.32	--	--	19.4
46	0	3.14	--	23.8	--
48	3	0.70	--	--	20.0
50	3	0.70	--	--	20.0
57	4	-0.13	--	18.7	--
59	4	-0.32	--	--	18.4
70	4	-0.06	--	--	18.8
81	4	0.06	--	19.0	--
86	2	-1.03	--	17.3	--
87	0	-8.27	--	< 6	--
97	0	2.18	22.3	--	--
105	3	0.58	--	--	19.8
109	1	-1.68	16.3	--	--
127	1	-1.54	16.5	--	--
131	0	-2.24	--	15.4	--
134	3	0.51	--	19.7	--
138	3	-0.90	--	--	17.5
142	4	0.26	--	--	19.3
145	4	0.06	--	19.0	--
146	4	0.26	--	19.3	--
149	1	-1.86	16.0	--	--
151	3	0.70	--	--	20.0
154	2	-1.35	--	16.8	--
180	NR	--	--	< 34.1	--
191	2	-1.28	--	--	16.9
196	4	0.32	--	--	19.4
212	2	1.09	--	--	20.6
220	4	0.28	--	19.3	--
234	4	-0.26	--	18.5	--
236	3	-0.58	--	18.0	--
247	4	0.32	--	--	19.4
259	3	-0.70	--	--	17.8
265	3	0.70	--	--	20.0
284	0	-4.42	--	12.0	--
296	2	1.03	--	--	20.5
305	3	-0.96	--	--	17.4
330	4	-0.26	--	--	18.5

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Na (Sodium) Concentration Unit : mg/L




SUMMARY

N =	13	47	9	1	3	1. AA: direct, air	MPV =	43.0
Minimum =	3.8	36.5	35.9	45.1	41.4	4. ICP	F-pseudosigma =	1.5
Maximum =	46.0	55.6	51.5		43.1	6. ICP/MS	Rating criterion =	2.2
Median =	43.2	43.0	43.2			7. Ion chromatography	N =	73
F-pseudosigma =	2.0	1.4	2.1			12. Flame emission	Uh =	44.0
							Lh =	42.0

Lab	Rating	Z-value	1	4	6	7	12	Lab	Rating	Z-value	1	4	6	7	12
1	4	-0.13	--	42.7	--	--	--	212	2	1.02	--	45.2	--	--	--
3	4	0.00	--	43.0	--	--	--	220	3	-0.55	--	41.8	--	--	--
5	4	-0.33	--	42.3	--	--	--	234	3	0.51	--	44.1	--	--	--
11	0	-3.02	--	36.5	--	--	--	236	4	-0.47	--	42.0	--	--	--
12	3	-0.93	--	41.0	--	--	--	247	4	0.14	--	43.3	--	--	--
13	4	0.28	--	43.6	--	--	--	254	4	0.19	--	43.4	--	--	--
19	4	0.00	--	43.0	--	--	--	259	4	0.09	--	--	43.2	--	--
23	2	-1.21	40.4	--	--	--	--	265	4	0.00	--	43.0	--	--	--
24	4	0.00	--	43.0	--	--	--	268	0	-2.14	--	--	38.4	--	--
25	1	-1.72	--	39.3	--	--	--	270	0	5.86	--	55.6	--	--	--
26	3	0.74	--	44.6	--	--	--	273	3	-0.65	--	41.6	--	--	--
32	3	0.84	--	--	44.8	--	--	277	4	-0.37	--	42.2	--	--	--
42	4	-0.19	--	42.6	--	--	--	279	2	-1.16	40.5	--	--	--	--
46	3	0.65	--	44.4	--	--	--	284	4	0.09	--	43.2	--	--	--
48	0	3.95	--	--	51.5	--	--	296	1	1.63	--	--	46.5	--	--
50	4	0.00	--	43.0	--	--	--	297	0	-2.05	--	38.6	--	--	--
57	4	-0.28	--	42.4	--	--	--	302	3	0.98	--	--	--	45.1	--
59	4	-0.19	42.6	--	--	--	--	305	0	-3.30	--	--	35.9	--	--
69	4	0.05	--	--	--	--	43.1	307	2	1.40	46.0	--	--	--	--
70	2	1.02	--	45.2	--	--	--	309	3	-0.93	--	41.0	--	--	--
76	4	0.36	--	--	43.8	--	--	324	2	1.02	45.2	--	--	--	--
81	2	1.02	--	45.2	--	--	--	330	0	5.12	--	54.0	--	--	--
83	3	-0.70	--	41.5	--	--	--	332	4	-0.47	--	--	42.0	--	--
84	4	-0.42	--	--	--	--	42.1								
86	4	0.47	--	44.0	--	--	--								
87	3	0.93	45.0	--	--	--	--								
89	3	-0.74	--	--	--	--	41.4								
93	1	1.53	--	46.3	--	--	--								
97	4	0.28	43.6	--	--	--	--								
105	4	0.19	--	43.4	--	--	--								
107	4	0.09	43.2	--	--	--	--								
109	4	-0.19	42.6	--	--	--	--								
110	0	-18.24	3.8	--	--	--	--								
113	4	-0.43	--	42.1	--	--	--								
121	4	-0.19	--	42.6	--	--	--								
126	2	1.40	46.0	--	--	--	--								
127	4	0.47	--	44.0	--	--	--								
131	4	-0.47	--	42.0	--	--	--								
134	4	-0.31	42.3	--	--	--	--								
138	4	0.14	--	43.3	--	--	--								
140	4	0.47	44.0	--	--	--	--								
142	1	1.77	--	46.8	--	--	--								
145	4	0.00	--	43.0	--	--	--								
146	4	-0.33	--	42.3	--	--	--								
180	1	-1.53	--	39.7	--	--	--								
191	4	-0.47	--	--	42.0	--	--								
193	3	-0.65	--	41.6	--	--	--								
203	4	-0.14	--	42.7	--	--	--								
204	4	0.42	--	43.9	--	--	--								
209	4	-0.34	--	42.3	--	--	--								

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Ni (Nickel) Concentration Unit : mg/L



SUMMARY

N =	3	10	30	22
Minimum =	23.0	25.7	10.0	25.1
Maximum =	46.0	32.2	45.3	31.3
Median =	30.0	28.3	29.2	28.9
F-pseudosigma =	0.0	1.9	2.3	1.3


- 1. AA: direct, air
- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	29.0
F-pseudosigma =	2.0
N =	65
Uh =	30.6
Lh =	28.0

Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	-0.05	--	--	--	28.9	212	3	0.87	--	--	--	30.7
3	2	-1.48	--	--	26.1	--	234	4	0.10	--	--	29.2	--
5	1	1.58	--	--	32.1	--	236	1	2.04	--	--	33.0	--
11	3	-0.51	--	--	28.0	--	247	4	0.20	--	--	--	29.4
13	1	1.69	--	--	32.3	--	254	4	0.10	--	--	29.2	--
19	4	0.10	--	--	29.2	--	255	4	-0.41	--	--	--	28.2
23	4	-0.15	--	--	28.7	--	259	3	0.51	--	--	--	30.0
24	0	2.55	--	--	34.0	--	265	4	0.26	--	--	--	29.5
25	2	-1.02	--	--	27.0	--	273	3	0.51	--	--	30.0	--
32	4	0.10	--	--	--	29.2	277	3	-0.51	--	--	28.0	--
42	4	0.05	--	--	--	29.1	284	0	-9.71	--	--	10.0	--
46	3	-0.56	--	--	27.9	--	296	2	1.07	--	--	--	31.1
48	2	-1.38	--	--	--	26.3	297	2	1.12	--	--	31.2	--
50	4	0.00	--	--	--	29.0	305	2	1.18	--	--	--	31.3
57	2	1.12	--	--	31.2	--	324	0	8.69	46.0	--	--	--
59	3	-0.61	--	--	--	27.8	330	4	-0.26	--	--	--	28.5
69	NR	--	< 50	--	--	--							
70	2	-1.38	--	--	--	26.3							
76	3	0.84	--	--	--	30.6							
81	1	-1.53	--	26.0	--	--							
83	4	-0.46	--	--	28.1	--							
86	4	0.05	--	--	29.1	--							
87	0	7.05	--	--	42.8	--							
89	4	0.20	--	29.4	--	--							
93	4	-0.26	--	--	28.5	--							
96	4	-0.36	--	28.3	--	--							
97	1	1.64	--	32.2	--	--							
105	3	-0.61	--	--	--	27.8							
114	0	-3.07	23.0	--	--	--							
121	3	-0.51	--	--	28.0	--							
126	3	-0.51	--	28.0	--	--							
127	4	-0.36	--	28.3	--	--							
131	0	8.33	--	--	45.3	--							
134	3	0.61	--	--	30.2	--							
138	4	0.36	--	--	29.7	--							
140	3	0.51	30.0	--	--	--							
142	4	-0.10	--	--	--	28.8							
144	3	-0.77	--	27.5	--	--							
145	2	1.02	--	--	31.0	--							
146	4	-0.15	--	--	28.7	--							
149	3	0.51	--	30.0	--	--							
151	2	-1.02	--	--	--	27.0							
154	4	0.00	--	--	29.0	--							
180	4	-0.15	--	--	28.7	--							
190	1	-1.69	--	25.7	--	--							
191	4	-0.26	--	--	--	28.5							
193	2	1.23	--	31.4	--	--							
196	2	-1.02	--	--	--	27.0							
203	1	1.53	--	--	32.0	--							
204	1	-1.99	--	--	--	25.1							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Pb (Lead) Concentration Unit : mg/L




SUMMARY									
N =	1	2	25	18	22	0. Other	MPV =	16.5	
Minimum =	17.0	20.0	14.0	8.0	15.0	1. AA: direct, air	F-pseudostigma =	1.2	
Maximum =		80.0	204.0	19.3	19.0	3. AA: graphite furnace	N =	68	
Median =			16.5	16.1	16.7	4. ICP	Uh =	17.5	
F-pseudostigma =			1.7	1.0	0.9	6. ICP/MS	Lh =	15.8	

Lab	Rating	Z-value	0	1	3	4	6	Lab	Rating	Z-value	0	1	3	4	6
1	3	0.55	--	--	--	--	17.2	193	3	0.57	--	--	17.2	--	--
3	3	0.90	--	--	--	17.6	--	196	3	-0.57	--	--	--	--	15.8
5	0	5.23	--	--	22.9	--	--	203	3	-0.57	--	--	15.8	--	--
10	4	-0.33	--	--	16.1	--	--	204	4	0.08	--	--	--	--	16.6
11	4	-0.41	--	--	--	16.0	--	212	1	2.04	--	--	--	--	19.0
12	0	153.30	--	--	204.0	--	--	227	4	-0.25	--	--	--	16.2	--
13	3	0.74	--	--	17.4	--	--	234	3	0.82	--	--	17.5	--	--
19	0	-3.60	--	--	--	12.1	--	236	0	-5.31	--	--	--	< 10	--
23	4	-0.25	--	--	--	16.2	--	247	4	0.49	--	--	--	--	17.1
25	4	-0.41	--	--	16.0	--	--	255	3	0.90	--	--	--	--	17.6
26	3	-0.83	--	--	--	15.5	--	259	4	-0.16	--	--	--	--	16.3
32	4	-0.08	--	--	--	--	16.4	265	4	-0.41	--	--	--	--	16.0
42	3	0.98	--	--	--	--	17.7	273	4	-0.33	--	--	--	16.1	--
46	3	0.74	--	--	--	17.4	--	277	4	-0.41	--	--	--	16.0	--
48	4	0.33	--	--	--	--	16.9	284	0	-6.95	--	--	--	8.0	--
50	3	-0.82	--	--	--	--	15.5	296	1	2.04	--	--	--	--	19.0
57	4	0.25	--	--	--	16.8	--	297	2	-1.37	--	--	--	14.8	--
59	1	1.55	--	--	--	--	18.4	305	2	-1.23	--	--	--	--	15.0
69	3	-0.98	--	--	15.3	--	--	307	3	-0.82	--	--	15.5	--	--
70	2	1.23	--	--	18.0	--	--	324	0	51.92	--	80.0	--	--	--
76	4	0.09	--	--	--	--	16.6	330	4	0.25	--	--	--	--	16.8
81	2	-1.23	--	--	15.0	--	--								
86	4	-0.41	--	--	16.0	--	--								
87	3	-0.57	--	--	--	15.8	--								
89	3	-0.90	--	--	15.4	--	--								
93	4	-0.41	--	--	--	16.0	--								
96	0	2.45	--	--	19.5	--	--								
97	3	0.57	--	--	--	17.2	--								
105	4	0.00	--	--	--	--	16.5								
109	1	1.64	--	--	18.5	--	--								
113	1	2.04	--	--	--	19.0	--								
114	0	-5.31	--	< 10	--	--	--								
126	0	2.37	--	--	19.4	--	--								
127	4	0.25	--	--	--	16.8	--								
131	4	0.33	--	--	--	16.9	--								
134	4	-0.08	--	--	--	16.4	--								
138	4	0.49	--	--	--	--	17.1								
140	0	2.86	--	20.0	--	--	--								
142	2	-1.06	--	--	--	--	15.2								
144	3	-0.65	--	--	15.7	--	--								
145	NR	--	--	--	--	< 35	--								
146	0	2.29	--	--	--	19.3	--								
147	4	0.25	--	--	--	--	16.8								
149	1	-2.04	--	--	14.0	--	--								
151	4	0.41	17.0	--	--	--	--								
154	4	0.00	--	--	16.5	--	--								
180	0	-2.45	--	--	--	13.5	--								
183	3	-0.65	--	--	15.7	--	--								
190	1	-2.04	--	--	14.0	--	--								
191	4	-0.41	--	--	--	--	16.0								
193	3	0.57	--	--	17.2	--	--								
196	3	-0.57	--	--	--	--	15.8								
203	3	-0.57	--	--	15.8	--	--								

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Sb (Antimony) Concentration Unit : mg/L



SUMMARY


N =	0	0	14	9	18	0. Other	MPV =	14.0
Minimum =			12.7	12.0	12.3	1. AA: direct, air	F-pseudostigma =	0.9
Maximum =			15.9	14.7	17.9	3. AA: graphite furnace	N =	41
Median =			14.0	13.5	14.5	4. ICP	Uh =	14.7
F-pseudostigma =			0.9	1.6	0.8	6. ICP/MS	Lh =	13.5

Lab	Rating	Z-value	0	1	3	4	6
1	4	0.34	--	--	14.3	--	--
3	3	0.79	--	--	--	14.7	--
5	NR	--	< 20	--	--	--	--
11	0	-2.25	--	--	--	12.0	--
13	2	1.01	--	--	14.9	--	--
23	4	0.45	--	--	--	14.4	--
25	0	-4.50	--	--	< 10	--	--
32	4	-0.22	--	--	--	--	13.8
42	2	1.01	--	--	--	--	14.9
46	0	2.14	--	--	15.9	--	--
48	4	-0.34	--	--	--	--	13.7
50	4	0.11	--	--	--	--	14.1
57	2	-1.01	--	--	--	13.1	--
59	1	2.02	--	--	--	--	15.8
69	2	-1.46	--	--	12.7	--	--
70	4	0.00	--	--	14.0	--	--
76	4	-0.27	--	--	--	--	13.8
81	2	-1.12	--	--	13.0	--	--
89	4	-0.22	--	--	13.8	--	--
96	4	0.34	--	--	14.3	--	--
97	4	-0.45	--	--	13.6	--	--
105	4	-0.11	--	--	--	--	13.9
113	3	-0.56	--	--	--	13.5	--
126	2	-1.12	--	--	13.0	--	--
127	2	-1.01	--	--	13.1	--	--
134	4	0.22	--	--	14.2	--	--
138	3	0.90	--	--	--	--	14.8
142	2	1.01	--	--	--	--	14.9
146	1	-1.91	--	--	--	12.3	--
151	4	0.00	--	--	--	--	14.0
154	4	0.00	--	--	14.0	--	--
180	3	0.67	--	--	--	14.6	--
191	1	-1.91	--	--	--	--	12.3
193	2	1.12	--	--	15.0	--	--
196	3	0.56	--	--	--	--	14.5
204	2	1.01	--	--	--	--	14.9
212	2	1.46	--	--	--	--	15.3
234	4	0.00	--	--	--	14.0	--
236	0	-4.50	--	--	--	< 10	--
247	3	0.56	--	--	--	--	14.5
265	2	-1.12	--	--	--	--	13.0
284	NR	--	--	< 200	--	--	--
296	0	4.38	--	--	--	--	17.9
297	0	-2.14	--	--	--	12.1	--
330	3	0.79	--	--	--	--	14.7

204 4 0.08 -- -- -- -- 16.6

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Se (Selenium) Concentration Unit : mg/L




SUMMARY									
N =	21	10	19	4	3. AA: graphite furnace	MPV =	9.58		
Minimum =	7.89	5.50	8.20	2.10	4. ICP	F-pseudosigma =	1.11		
Maximum =	13.10	14.00	12.80	8.00	6. ICP/MS	N =	54		
Median =	9.60	9.09	9.77		11. AA: hydride NaBH4	Uh =	10.50		
F-pseudosigma =	0.39	3.93	1.06			Lh =	9.00		

Lab	Rating	Z-value	3	4	6	11na	Lab	Rating	Z-value	3	4	6	11na
1	4	0.08	9.67	--	--	--	259	4	-0.25	--	--	9.30	--
3	3	0.92	--	10.60	--	--	265	3	0.83	--	--	10.50	--
5	0	3.17	13.10	--	--	--	284	0	3.98	--	14.00	--	--
10	2	-1.42	--	--	--	8.00	296	1	1.55	--	--	11.30	--
11	3	-0.52	--	9.00	--	--	297	0	-2.14	--	7.20	--	--
13	3	0.56	10.20	--	--	--	307	1	-1.52	7.89	--	--	--
19	0	3.35	--	13.30	--	--	330	4	0.02	--	--	9.60	--
23	2	-1.38	--	8.04	--	--							
25	4	0.02	9.60	--	--	--							
26	4	-0.36	--	9.18	--	--							
32	2	1.28	--	--	11.00	--							
42	2	1.10	--	--	10.80	--							
46	4	0.13	9.73	--	--	--							
48	2	-1.24	--	--	8.20	--							
50	4	0.02	--	--	9.60	--							
57	3	0.74	10.40	--	--	--							
59	2	1.01	--	--	10.70	--							
69	3	-0.97	8.50	--	--	--							
70	4	0.17	--	--	9.77	--							
81	NR	--	--	< 12	--	--							
86	1	-2.01	--	--	--	7.35							
87	0	-6.73	--	--	--	2.10							
89	0	-3.08	--	--	--	6.16							
93	0	2.63	--	12.50	--	--							
96	4	-0.07	9.50	--	--	--							
97	4	-0.26	9.29	--	--	--							
105	4	-0.43	--	--	9.10	--							
109	0	2.05	11.86	--	--	--							
113	4	-0.18	9.38	--	--	--							
126	4	0.02	9.60	--	--	--							
127	4	-0.21	9.35	--	--	--							
131	0	-3.67	--	5.50	--	--							
134	4	0.11	9.70	--	--	--							
138	4	-0.36	--	--	9.18	--							
142	0	2.90	--	--	12.80	--							
144	3	-0.52	9.00	--	--	--							
146	0	-2.30	--	7.02	--	--							
151	4	0.38	--	--	10.00	--							
154	1	1.73	11.50	--	--	--							
180	NR	--	--	< 21.4	--	--							
193	4	-0.16	9.40	--	--	--							
196	3	0.56	--	--	10.20	--							
203	4	-0.25	9.30	--	--	--							
204	3	-0.97	--	--	8.50	--							
212	4	-0.02	--	--	9.56	--							
220	4	-0.07	9.50	--	--	--							
234	4	--	9.88	--	--	--							
236	NR	--	--	< 100	--	--							
247	2	1.32	--	--	11.05	--							
255	4	-0.22	--	--	9.34	--							

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: SiO₂ (Silica) Concentration Unit : mg/L




SUMMARY

N =	29	6	1	9	4. ICP	MPV =	14.8
Minimum =	6.7	13.4	14.6	13.8	6. ICP/MS	F-pseudostandard deviation =	1.0
Maximum =	16.9	16.0		17.5	20. Titrate: colorimetric	N =	45
Median =	14.8			15.3	22mb. Color: molybdate blue	Uh =	15.5
F-pseudostandard deviation =	0.9			0.6		Lh =	14.1


Lab	Rating	Z-value	4	6	20	22mb
1	4	0.45	15.3	--	--	--
3	3	-0.81	14.0	--	--	--
5	4	-0.23	14.6	--	--	--
11	1	-1.98	12.8	--	--	--
13	4	0.25	15.1	--	--	--
24	2	1.21	16.1	--	--	--
25	3	-1.00	13.8	--	--	--
26	3	0.73	15.6	--	--	--
32	2	1.12	--	16.0	--	--
42	4	-0.04	14.8	--	--	--
50	3	0.73	--	15.6	--	--
57	4	0.35	15.2	--	--	--
70	3	0.64	--	--	--	15.5
76	4	-0.01	--	14.8	--	--
83	4	-0.33	14.5	--	--	--
87	4	0.15	--	--	--	15.0
89	4	0.15	--	--	--	15.0
97	3	-0.71	--	--	--	14.1
105	4	-0.04	14.8	--	--	--
110	4	-0.01	14.8	--	--	--
121	3	0.54	15.4	--	--	--
127	3	-0.71	14.1	--	--	--
131	1	-1.97	12.8	--	--	--
134	4	0.00	14.8	--	--	--
140	3	0.98	--	--	--	15.9
142	1	1.89	16.8	--	--	--
145	4	-0.33	14.5	--	--	--
151	4	-0.23	--	--	14.6	--
154	3	0.73	15.6	--	--	--
190	1	1.79	--	--	--	16.7
191	3	-0.81	--	14.0	--	--
203	0	2.56	--	--	--	17.5
212	3	0.64	15.5	--	--	--
218	1	1.97	16.9	--	--	--
227	3	-0.62	14.2	--	--	--
234	4	0.15	15.0	--	--	--
236	0	-2.16	12.6	--	--	--
254	4	0.41	15.3	--	--	--
259	4	-0.23	--	14.6	--	--
265	4	0.15	15.0	--	--	--
284	0	-7.62	6.9	--	--	--
297	2	-1.01	--	--	--	13.8
309	0	-7.84	6.7	--	--	--
330	4	0.44	--	--	--	15.3
332	2	-1.39	--	13.4	--	--

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Sr (Strontium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY

N =	1	1	28	9
Minimum =	72.6	36.7	50.3	52.5
Maximum =			58.9	58.2
Median =			54.1	56.0
F-pseudostigma =			2.3	2.3


- 1. AA: direct, air
- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	54.2
F-pseudostigma =	2.4
Rating criterion =	2.7
N =	39
Uh =	56.0
Lh =	52.8

Lab	Rating	Z-value	1	3	4	6
1	4	0.00	--	--	54.2	--
3	2	-1.23	--	--	50.9	--
5	3	-0.68	--	--	52.4	--
11	2	-1.19	--	--	51.0	--
24	3	0.65	--	--	56.0	--
25	4	-0.09	--	--	54.0	--
32	4	0.21	--	--	--	54.8
42	4	-0.09	--	--	54.0	--
50	2	1.46	--	--	--	58.2
57	4	-0.09	--	--	54.0	--
70	4	0.46	--	--	55.5	--
81	3	0.65	--	--	56.0	--
86	3	0.54	--	--	55.7	--
97	0	-6.47	--	36.7	--	--
105	3	0.65	--	--	56.0	--
109	0	6.77	72.6	--	--	--
113	4	-0.27	--	--	53.5	--
121	3	-0.83	--	--	52.0	--
127	3	-0.64	--	--	52.5	--
131	2	-1.45	--	--	50.3	--
134	4	0.28	--	--	55.0	--
138	4	-0.01	--	--	54.2	--
142	1	1.72	--	--	58.9	--
145	4	-0.46	--	--	53.0	--
151	3	0.65	--	--	--	56.0
154	4	0.17	--	--	54.7	--
191	3	-0.64	--	--	--	52.5
196	4	-0.20	--	--	--	53.7
212	2	1.35	--	--	57.9	--
234	4	0.43	--	--	55.4	--
236	3	-0.83	--	--	52.0	--
247	4	-0.09	--	--	--	54.0
254	3	0.76	--	--	56.3	--
259	3	0.76	--	--	--	56.3
265	3	-0.83	--	--	52.0	--
273	4	-0.16	--	--	53.8	--
284	4	0.28	--	--	55.0	--
330	2	1.05	--	--	--	57.1
332	2	1.09	--	--	--	57.2

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Tl (Thallium) Concentration Unit : mg/L




SUMMARY

N =	1	13	10	17	0. Other	MPV =	50.3
Minimum =	53.8	41.0	31.0	43.9	3. AA: graphite furnace	F-pseudosigma =	3.8
Maximum =		58.0	56.0	58.1	4. ICP	N =	41
Median =		50.2	48.8	52.0	6. ICP/MS	Uh =	52.8
F-pseudosigma =		2.7	5.6	3.2		Lh =	47.7

Lab	Rating	Z-value	0	3	4	6
1	4	0.44	--	--	--	51.9
3	4	-0.07	--	--	50.0	--
11	2	-1.39	--	--	45.0	--
13	3	-0.60	--	48.0	--	--
23	3	-0.99	--	--	46.5	--
25	3	-0.60	--	48.0	--	--
32	4	-0.12	--	--	--	49.8
42	4	0.46	--	--	--	52.0
46	3	0.67	--	52.8	--	--
48	4	-0.33	--	--	--	49.0
50	2	-1.12	--	--	--	46.0
57	4	-0.12	--	--	49.8	--
59	3	0.94	53.8	--	--	--
69	4	0.36	--	51.6	--	--
70	1	2.05	--	58.0	--	--
81	0	-2.45	--	41.0	--	--
89	2	1.42	--	55.6	--	--
97	1	-1.60	--	44.2	--	--
105	3	0.86	--	--	--	53.5
113	4	0.33	--	51.5	--	--
126	3	-0.60	--	48.0	--	--
127	4	-0.36	--	48.9	--	--
134	4	0.00	--	50.3	--	--
138	3	0.81	--	--	--	53.3
142	4	0.46	--	--	--	52.0
146	3	-0.67	--	--	47.7	--
151	4	0.46	--	--	--	52.0
154	4	-0.01	--	50.2	--	--
180	0	-3.77	--	--	36.0	--
196	1	-1.68	--	--	--	43.9
204	3	-0.78	--	--	--	47.3
212	3	0.67	--	--	--	52.8
234	3	0.60	--	--	52.5	--
247	3	0.94	--	--	--	53.8
265	3	-0.86	--	--	--	47.0
273	0	-5.09	--	--	31.0	--
284	1	1.52	--	--	56.0	--
296	0	2.08	--	--	--	58.1
297	3	1.00	--	--	54.0	--
305	1	1.76	--	--	--	56.9
330	3	0.52	--	--	--	52.2

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: U (Uranium) Concentration Unit : mg/L



SUMMARY

N = 0 9
 Minimum = 7.00
 Maximum = 9.49
 Median = 7.97
 F-pseudosigma = 0.44


4. ICP
 6. ICP/MS

MPV = 7.97
 F-pseudosigma = 0.44
 N = 9
 Uh = 8.30
 Lh = 7.70

Lab	Rating	Z-value	4	6
1	4	0.00	--	7.97
32	3	-0.56	--	7.72
70	0	3.42	--	9.49
127	NR	--	< 100	--
134	4	0.29	--	8.10
142	3	0.74	--	8.30
191	3	-0.61	--	7.70
196	2	-1.37	--	7.36
265	0	-2.18	--	7.00
330	2	1.19	--	8.50

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: V (Vanadium) Concentration Unit : mg/L



SUMMARY

N =	2	21	17
Minimum =	17.3	13.2	16.5
Maximum =	19.9	23.3	22.2
Median =		18.2	18.5
F-pseudostigma =		1.2	0.6


- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	18.4
F-pseudostigma =	1.2
N =	40
Uh =	19.0
Lh =	17.4

Lab	Rating	Z-value	3	4	6
1	4	0.12	--	--	18.5
3	3	-0.81	--	17.4	--
5	2	-1.15	--	17.0	--
11	2	-1.15	--	17.0	--
13	NR	--	--	< 50	--
24	2	1.38	--	20.0	--
25	0	-4.52	--	< 13	--
32	4	0.12	--	--	18.5
42	4	0.29	--	--	18.7
46	2	-1.40	--	16.7	--
48	1	-1.57	--	--	16.5
50	3	0.71	--	--	19.2
57	2	-1.15	--	17.0	--
59	4	0.37	--	--	18.8
70	3	-0.89	--	--	17.3
81	4	-0.30	--	18.0	--
86	0	2.48	--	21.3	--
89	2	1.30	19.9	--	--
93	3	-0.81	--	17.4	--
96	NR	--	< 40	--	--
97	3	-0.89	17.3	--	--
105	4	-0.47	--	--	17.8
127	0	-4.35	--	13.2	--
134	4	-0.05	--	18.3	--
138	4	0.12	--	18.5	--
142	4	-0.13	--	--	18.2
145	3	0.54	--	19.0	--
146	4	0.03	--	18.4	--
154	4	0.20	--	18.6	--
180	0	4.17	--	23.3	--
191	4	0.46	--	--	18.9
196	4	-0.22	--	--	18.1
212	2	-1.40	--	--	16.7
220	4	-0.41	--	17.9	--
234	4	-0.13	--	18.2	--
236	3	0.54	--	19.0	--
247	4	0.29	--	--	18.7
265	3	0.54	--	--	19.0
284	1	-1.99	--	16.0	--
296	0	3.24	--	--	22.2
297	4	-0.03	--	18.3	--
305	3	0.54	--	--	19.0
330	4	0.03	--	--	18.4

Table 11. Statistical summary of reported data for standard reference water sample T-161 (trace constituents)--Continued

Analyte: Zn (Zinc) Concentration Unit : mg/L



SUMMARY

N =	10	2	36	23	1. AA: direct, air	MPV =	40.6
Minimum =	35.0	4.5	9.0	28.4	3. AA: graphite furnace	F-pseudosigma =	2.4
Maximum =	47.0	36.2	50.0	48.7	4. ICP	N =	71
Median =	40.2		40.4	41.0	6. ICP/MS	Uh =	42.2
F-pseudosigma =	3.1		1.7	4.8	11. AA: hydride	Lh =	38.9


Lab	Rating	Z-value	1	3	4	6	11	Lab	Rating	Z-value	1	3	4	6	11
1	4	0.04	--	--	--	40.7	--	203	4	-0.25	--	--	40.0	--	--
3	4	0.33	--	--	41.4	--	--	204	2	-1.02	--	--	--	38.1	--
5	3	-0.78	--	--	38.7	--	--	212	3	0.94	--	--	--	42.9	--
10	4	-0.08	40.4	--	--	--	--	220	4	-0.42	--	--	39.6	--	--
11	3	-0.65	--	--	39.0	--	--	227	4	-0.25	--	--	40.0	--	--
12	0	-14.76	--	4.5	--	--	--	234	4	0.33	--	--	41.4	--	--
13	3	0.94	--	--	42.9	--	--	236	1	-1.88	--	--	36.0	--	--
19	4	0.12	--	--	40.9	--	--	247	0	2.21	--	--	--	46.0	--
23	4	0.08	--	--	40.8	--	--	254	4	0.12	--	--	40.9	--	--
24	2	1.39	--	--	44.0	--	--	255	3	-0.61	--	--	--	39.1	--
25	0	-12.92	--	--	9.0	--	--	259	4	0.00	--	--	--	40.6	--
26	3	-0.86	--	--	38.5	--	--	265	0	-2.29	--	--	--	35.0	--
32	3	-0.65	--	--	--	39.0	--	268	0	-4.99	--	--	--	28.4	--
42	0	2.25	--	--	--	46.1	--	273	0	-4.86	--	--	28.7	--	--
46	3	-0.53	--	--	39.3	--	--	277	3	-0.61	--	--	39.1	--	--
48	4	0.16	--	--	--	41.0	--	284	3	0.98	--	--	43.0	--	--
50	4	0.00	--	--	--	40.6	--	296	1	1.51	--	--	--	44.3	--
57	0	3.84	--	--	50.0	--	--	297	4	-0.41	--	--	39.6	--	--
59	2	1.06	--	--	--	43.2	--	305	2	1.16	--	--	--	43.4	--
69	NR	--	--	--	--	--	< 50	307	0	2.62	47.0	--	--	--	--
70	4	-0.08	--	--	--	40.4	--	324	0	-2.29	35.0	--	--	--	--
81	1	1.80	--	--	45.0	--	--	330	2	1.39	--	--	--	44.0	--
83	3	-0.65	--	--	39.0	--	--								
86	4	0.12	--	--	40.9	--	--								
87	4	0.08	--	--	40.8	--	--								
89	1	-1.80	--	36.2	--	--	--								
93	2	-1.43	--	--	37.1	--	--								
96	2	-1.06	38.0	--	--	--	--								
105	1	-1.88	--	--	--	36.0	--								
113	3	0.69	--	--	42.3	--	--								
114	2	1.19	43.5	--	--	--	--								
121	4	-0.25	--	--	40.0	--	--								
126	3	-0.65	39.0	--	--	--	--								
127	4	0.12	--	--	40.9	--	--								
131	2	-1.31	--	--	37.4	--	--								
134	4	0.16	--	--	41.0	--	--								
138	4	0.08	--	--	40.8	--	--								
140	3	0.57	42.0	--	--	--	--								
142	0	-3.31	--	--	--	32.5	--								
144	4	-0.25	40.0	--	--	--	--								
145	4	0.16	--	--	41.0	--	--								
146	0	-2.41	--	--	34.7	--	--								
147	4	0.41	--	--	--	41.6	--								
151	4	0.16	--	--	--	41.0	--								
154	3	0.57	--	--	42.0	--	--								
180	1	-2.04	--	--	35.6	--	--								
183	2	1.06	43.2	--	--	--	--								
191	0	3.31	--	--	--	48.7	--								
193	4	-0.25	40.0	--	--	--	--								
196	0	3.07	--	--	--	48.1	--								

Table 12. *Statistical summary of reported data for standard reference sample M-154 (major constituents)*


Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported		
1. AA: direct, air	=	atomic absorption: direct, air
2. AA: direct, N ₂ O	=	atomic absorption: direct, nitrous oxide
3. AA: graphite furnace	=	atomic absorption: graphite furnace
4. ICP	=	inductively coupled plasma
5. DCP	=	direct current plasma
6. ICP/MS	=	inductively coupled plasma / mass spectrometry
7. IC	=	ion chromatography
12. Flame emission	=	flame emission
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color:	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode
41. Electro	=	electrometric: (type meter specified)
50. Gravimetric	=	gravimetric: (precipitate specified)
51. Turbidimetric	=	turbidimetric: (precipitate specified)
<hr/>		
<u>Abbreviations and figure symbols</u>		
N	=	number of analyses--(excluding less than values)
MPV	=	most probable value ---
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit ----
Lwl	=	lower warning limit ----
Ucl	=	upper warning limit
Lcl	=	lower warning limit ----
µg/L	=	micrograms per liter
mg/L	=	milligrams per liter
µS/cm	=	microsiemens per centimeter at 25 ⁰ C
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported
<hr/>		
<u>Constituent</u>		<u>page</u>
Alk	Alkalinity as CaCO ₃	65
B	Boron	66
Ca	Calcium	67
Cl	Chloride	68
DSRD	Dissolved solids	69
F	Fluoride	70
K	Potassium	71
Mg	Magnesium	72
Na	Sodium	73
total P	Phosphorus	74
pH		75
SiO ₂	Silica	76
SO ₄	Sulfate	77
Sp Cond	Specific Conductance	78
Sr	Strontium	79
V	Vanadium	80

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: Alkalinity (as CaCO₃) Concentration Unit : mg/L




ANALYTICAL METHOD CODE



SUMMARY																
N =	4	4	61	3	3	0. Other					MPV =	112				
Minimum =	101	34	11	98	109	20. Titrate: colorimetric					F-pseudsigma =	4				
Maximum =	115	115	219	117	122	21. Titrate: electrometric					Rating criterion =	6				
Median =			112			22. Colorimetric					N =	75				
F-pseudsigma =			4			41. Direct reading					Uh =	114				
											Lh =	109				
Lab	Rating	Z-value	0	20	21	22	41	Lab	Rating	Z-value	0	20	21	22	41	
1	4	0.18	--	--	113	--	--	204	3	-0.89	--	--	107	--	--	
3	4	0.36	--	--	114	--	--	212	3	-0.54	--	--	109	--	--	
5	1	-1.96	101	--	--	--	--	213	4	0.00	--	--	112	--	--	
10	4	0.16	--	--	113	--	--	220	1	-1.68	--	--	103	--	--	
11	1	1.61	--	--	121	--	--	227	4	0.50	115	--	--	--	--	
12	2	-1.25	--	--	105	--	--	234	3	0.89	--	--	--	117	--	
13	2	1.43	--	--	120	--	--	236	4	0.00	--	--	112	--	--	
19	4	-0.36	--	--	110	--	--	244	3	0.80	--	--	117	--	--	
23	3	-0.57	--	--	109	--	--	247	4	-0.18	--	--	111	--	--	
24	4	0.21	--	--	113	--	--	257	3	0.71	--	--	116	--	--	
25	2	1.43	--	--	120	--	--	259	4	0.00	--	--	112	--	--	
26	1	1.59	--	--	121	--	--	267	4	0.00	--	--	112	--	--	
38	4	0.04	--	--	112	--	--	268	3	-0.57	--	--	109	--	--	
42	0	-18.00	--	--	11	--	--	269	4	0.36	--	--	114	--	--	
43	1	1.96	--	--	123	--	--	273	4	0.36	114	--	--	--	--	
46	3	-0.54	--	--	--	--	109	277	0	19.11	--	--	219	--	--	
48	2	-1.07	--	--	--	106	--	292	4	0.36	--	--	114	--	--	
50	4	0.36	--	--	114	--	--	297	4	0.18	--	--	113	--	--	
57	0	-13.93	--	34	--	--	--	302	3	-0.89	--	107	--	--	--	
59	3	0.63	--	--	116	--	--	305	3	0.54	--	--	115	--	--	
69	4	0.00	--	--	112	--	--	307	3	-0.54	--	109	--	--	--	
70	4	-0.18	--	--	111	--	--	309	3	0.54	--	115	--	--	--	
81	4	0.18	--	--	113	--	--	323	1	-1.96	--	--	101	--	--	
83	4	-0.37	--	--	110	--	--	324	0	3.04	--	--	129	--	--	
84	4	0.00	--	--	112	--	--	330	4	-0.18	--	--	111	--	--	
85	3	-0.89	--	--	107	--	--									
87	4	0.18	--	--	113	--	--									
89	4	0.18	--	--	113	--	--									
93	4	-0.27	111	--	--	--	--									
96	4	-0.36	--	--	110	--	--									
97	4	0.11	--	--	113	--	--									
105	4	0.00	--	--	112	--	--									
107	3	-0.54	--	--	109	--	--									
109	1	1.68	--	--	121	--	--									
113	4	-0.20	--	--	111	--	--									
114	3	-0.71	--	--	108	--	--									
127	3	-0.71	--	--	108	--	--									
134	4	0.15	--	--	--	--	113									
138	4	0.00	--	--	112	--	--									
142	0	-2.14	--	--	100	--	--									
145	0	-2.50	--	--	--	98	--									
146	0	-2.82	--	--	96	--	--									
149	4	0.00	--	--	112	--	--									
151	4	-0.18	--	--	111	--	--									
154	0	-2.95	--	--	96	--	--									
180	1	1.79	--	--	--	--	122									
183	4	0.18	--	--	113	--	--									
190	3	-0.54	--	--	109	--	--									
193	4	0.00	--	--	112	--	--									
203	2	-1.18	--	--	105	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: B (Boron) Concentration Unit : mg/L



SUMMARY

N = 31 3
 Minimum = 225 276
 Maximum = 370 435
 Median = 283
 F-pseudosigma = 17


4. ICP
 6. ICP/MS

MPV = 284
 F-pseudosigma = 18
 N = 34
 Uh = 297
 Lh = 272

Lab	Rating	Z-value	4	6
1	4	-0.39	276	--
3	4	-0.19	280	--
5	3	-0.62	272	--
11	3	-0.73	270	--
24	4	0.24	288	--
25	0	-3.17	225	--
26	3	-0.79	269	--
42	2	1.16	--	305
46	3	-0.57	273	--
48	4	-0.41	--	276
50	4	-0.49	274	--
57	0	4.69	370	--
70	4	0.14	286	--
86	4	0.03	284	--
105	2	-1.06	264	--
127	3	0.62	295	--
131	3	-0.95	266	--
134	4	0.35	290	--
138	4	-0.08	282	--
142	3	0.95	301	--
145	4	0.46	292	--
154	4	0.41	291	--
180	3	-0.73	270	--
212	3	-0.57	273	--
220	3	0.73	297	--
234	3	0.73	297	--
236	3	-0.89	267	--
247	4	-0.03	283	--
255	3	0.79	298	--
259	2	1.49	311	--
265	3	0.89	300	--
273	2	-1.27	260	--
323	4	0.14	286	--
330	0	8.21	--	435

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: Ca (Calcium) Concentration Unit : mg/L




SUMMARY																
N =	2	12	2	52	8	4	0. Other	MPV =	67.5							
Minimum =	60.9	6.6	66.7	61.7	60.2	66.4	1. AA: direct, air	F-pseudosigma =	3.1							
Maximum =	64.3	69.0	67.4	74.5	75.4	76.8	2. AA: direct, nitrous oxide	Rating criterion =	3.4							
Median =		64.3		67.5	69.4		4. ICP	N =	80							
F-pseudosigma =		4.6		2.5	3.7		6. ICP/MS	Uh =	69.4							
							20. Titrate: colorimetric	Lh =	65.2							


Lab	Rating	Z-value	0	1	2	4	6	20	Lab	Rating	Z-value	0	1	2	4	6	20
1	4	-0.06	--	--	--	67.3	--	--	212	2	1.10	--	--	--	71.2	--	--
3	3	-0.56	--	--	--	65.6	--	--	218	1	1.68	--	--	--	73.2	--	--
5	2	-1.07	--	--	--	63.9	--	--	220	4	0.33	--	--	--	68.6	--	--
10	4	0.15	--	68.0	--	--	--	--	227	0	2.07	--	--	--	74.5	--	--
11	4	-0.44	--	--	--	66.0	--	--	234	4	-0.03	--	--	--	67.4	--	--
12	4	0.15	--	--	--	68.0	--	--	236	4	0.21	--	--	--	68.2	--	--
13	2	1.30	--	--	--	71.9	--	--	247	4	0.27	--	--	--	68.4	--	--
19	3	-0.59	--	--	--	65.5	--	--	254	4	0.39	--	--	--	68.8	--	--
23	0	2.76	--	--	--	--	--	76.8	255	2	1.21	--	--	--	71.6	--	--
24	4	0.15	--	--	--	68.0	--	--	257	4	0.44	--	--	--	--	--	69.0
25	3	0.68	--	--	--	69.8	--	--	259	3	0.53	--	--	--	--	69.3	--
26	2	1.21	--	--	--	71.6	--	--	265	3	0.68	--	--	--	69.8	--	--
38	4	-0.03	--	--	67.4	--	--	--	267	4	-0.33	--	--	--	--	--	66.4
42	4	-0.24	--	--	--	66.7	--	--	268	4	0.00	--	67.5	--	--	--	--
46	4	-0.18	--	--	--	66.9	--	--	269	3	0.74	--	--	--	--	--	70.0
48	0	2.34	--	--	--	--	75.4	--	270	3	0.74	--	--	--	70.0	--	--
50	4	-0.39	--	--	--	66.2	--	--	273	4	-0.12	--	--	--	67.1	--	--
57	4	0.06	--	--	--	67.7	--	--	277	1	1.99	--	--	--	74.2	--	--
59	3	0.59	--	--	--	--	69.5	--	279	3	-0.95	--	64.3	--	--	--	--
69	3	-0.98	--	64.2	--	--	--	--	292	0	-2.07	--	60.5	--	--	--	--
70	3	0.74	--	--	--	70.0	--	--	296	2	1.42	--	--	--	--	72.3	--
76	4	0.26	--	--	--	--	68.4	--	297	1	-1.69	--	--	--	61.8	--	--
81	4	-0.24	--	--	--	66.7	--	--	302	3	-0.95	64.3	--	--	--	--	--
83	4	-0.15	--	--	--	67.0	--	--	305	3	-0.78	--	--	--	--	64.9	--
84	2	-1.33	--	63.0	--	--	--	--	307	2	-1.04	--	64.0	--	--	--	--
85	0	-18.06	--	6.6	--	--	--	--	309	2	-1.04	--	--	--	64.0	--	--
86	4	0.50	--	--	--	69.2	--	--	323	3	-0.89	--	--	--	64.5	--	--
87	4	-0.24	--	--	66.7	--	--	--	324	0	-8.36	--	39.3	--	--	--	--
89	1	-1.96	60.9	--	--	--	--	--	330	1	1.66	--	--	--	73.1	--	--
93	4	-0.21	--	--	--	66.8	--	--	332	0	-2.16	--	--	--	60.2	--	--
97	4	0.27	--	68.4	--	--	--	--									
102	4	0.44	--	--	--	69.0	--	--									
105	4	-0.06	--	--	--	67.3	--	--									
109	4	0.12	--	67.9	--	--	--	--									
113	4	-0.19	--	--	--	66.9	--	--									
121	4	0.00	--	--	--	67.5	--	--									
127	4	0.00	--	--	--	67.5	--	--									
131	1	-1.63	--	--	--	62.0	--	--									
134	4	-0.13	--	--	--	67.1	--	--									
138	4	0.15	--	--	--	68.0	--	--									
140	4	0.44	--	69.0	--	--	--	--									
142	1	1.93	--	--	--	74.0	--	--									
145	4	-0.24	--	--	--	66.7	--	--									
146	4	0.00	--	--	--	67.5	--	--									
154	2	1.19	--	--	--	71.5	--	--									
180	1	-1.54	--	--	--	62.3	--	--									
191	3	0.98	--	--	--	--	70.8	--									
193	2	-1.27	--	--	--	63.2	--	--									
203	1	-1.72	--	--	--	61.7	--	--									
209	2	-1.33	--	--	--	63.0	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: Cl (Chloride) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
		N =	2	40	5	9	18	3	0. Other	MPV =	141						
		Minimum =	136	74	135	132	136	140	7. Ion chromatography	F-pseudosigma =	4						
		Maximum =	138	167	142	184	173	142	20. Titrate: colorimetric	Rating criterion =	7						
		Median =	140		142	142			21. Titrate: electrometric	N =	77						
		F-pseudosigma =	3		3	4			22. Colorimetric	Uh =	143						
									40. Ion selective electrode	Lh =	138						
Lab	Rating	Z-value	0	7	20	21	22	40	Lab	Rating	Z-value	0	7	20	21	22	40
1	3	-0.51	--	137	--	--	--	--	208	3	-0.78	--	136	--	--	--	--
3	1	1.99	--	--	--	--	155	--	209	0	-9.56	--	74	--	--	--	--
4	4	-0.14	--	140	--	--	--	--	212	3	-0.85	--	135	--	--	--	--
5	4	-0.28	--	139	--	--	--	--	213	4	-0.28	--	--	139	--	--	--
10	4	-0.24	--	--	--	--	139	--	220	1	1.66	--	153	--	--	--	--
11	3	0.57	--	--	--	--	145	--	227	1	-1.55	--	130	--	--	--	--
12	4	0.43	--	--	--	--	144	--	234	4	0.14	--	142	--	--	--	--
13	4	-0.43	--	138	--	--	--	--	236	0	2.41	--	158	--	--	--	--
19	3	-0.85	--	--	135	--	--	--	247	4	-0.14	--	140	--	--	--	--
23	3	0.71	--	146	--	--	--	--	254	4	-0.28	--	139	--	--	--	--
24	4	-0.14	--	--	--	--	140	--	257	4	0.35	--	--	--	144	--	--
25	4	-0.28	--	139	--	--	--	--	259	4	0.00	--	--	--	141	--	--
43	4	0.14	--	--	--	--	--	142	265	4	0.00	--	141	--	--	--	--
46	4	0.43	--	--	--	--	144	--	267	4	-0.20	--	--	--	140	--	--
48	2	1.14	--	--	--	--	149	--	268	3	-0.99	--	134	--	--	--	--
50	3	-0.71	--	--	--	--	136	--	269	4	-0.14	--	--	--	--	--	140
57	4	-0.14	--	--	140	--	--	--	270	0	3.69	--	--	167	--	--	--
59	4	0.00	--	141	--	--	--	--	273	4	-0.48	138	--	--	--	--	--
69	4	-0.28	--	139	--	--	--	--	277	4	-0.34	--	139	--	--	--	--
70	1	-1.99	--	127	--	--	--	--	292	4	0.48	--	144	--	--	--	--
76	4	0.13	--	142	--	--	--	--	297	3	-0.71	136	--	--	--	--	--
81	4	0.14	--	--	--	142	--	--	302	4	-0.43	--	138	--	--	--	--
85	4	-0.28	--	139	--	--	--	--	307	4	0.14	--	--	142	--	--	--
87	4	0.28	--	--	--	--	143	--	309	3	-0.71	--	--	136	--	--	--
89	4	-0.28	--	139	--	--	--	--	323	4	0.14	--	--	--	--	142	--
93	3	-0.85	--	135	--	--	--	--	324	0	6.15	--	--	--	184	--	--
96	4	-0.43	--	--	--	--	138	--	330	3	0.99	--	--	--	--	148	--
97	4	0.00	--	--	--	--	141	--									
102	4	0.13	--	142	--	--	--	--									
105	3	0.71	--	146	--	--	--	--									
107	2	-1.28	--	--	--	132	--	--									
109	3	0.64	--	--	--	146	--	--									
113	3	-0.57	--	137	--	--	--	--									
114	4	-0.14	--	--	--	--	--	140									
127	4	0.28	--	143	--	--	--	--									
131	1	-1.70	--	129	--	--	--	--									
134	4	-0.02	--	141	--	--	--	--									
138	4	0.00	--	141	--	--	--	--									
140	4	-0.02	--	--	--	--	141	--									
142	4	0.14	--	142	--	--	--	--									
143	0	4.54	--	--	--	--	173	--									
145	4	0.17	--	142	--	--	--	--									
146	4	-0.14	--	--	--	--	140	--									
149	4	0.14	--	142	--	--	--	--									
154	4	0.00	--	--	--	--	141	--									
180	3	0.99	--	148	--	--	--	--									
183	3	-0.67	--	--	--	--	136	--									
191	4	0.28	--	143	--	--	--	--									
203	4	0.28	--	--	--	143	--	--									
204	4	-0.43	--	--	--	138	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued
 Analyte: DSRD (Dissolved solids) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY					
N =	3	50	0. Other		MPV = 701
Minimum =	694	450	50. Gravimetric		F-pseudosigma = 14
Maximum =	699	757			Rating criterion = 35
Median =		703			N = 53
F-pseudosigma =		15			Uh = 711
					Lh = 692

Lab	Rating	Z-value	0	50	Lab	Rating	Z-value	0	50
1	4	0.24	--	710	277	1	1.60	--	757
3	4	0.26	--	710	305	4	0.26	--	710
5	4	-0.06	699	--	323	3	-0.60	--	680
10	4	0.14	--	706					
11	1	-1.74	--	640					
12	3	-0.66	--	678					
13	4	-0.14	--	696					
19	4	0.29	--	711					
23	4	-0.49	--	684					
25	4	-0.26	--	692					
26	4	-0.29	--	691					
38	4	-0.31	--	690					
43	4	0.43	--	716					
46	4	0.14	--	706					
48	3	0.54	--	720					
50	4	0.17	--	707					
57	4	0.31	--	712					
59	4	0.03	--	702					
69	4	-0.20	--	694					
70	4	0.00	--	701					
76	2	1.32	--	747					
81	4	-0.14	--	696					
85	4	-0.20	--	694					
87	4	0.26	--	710					
89	4	0.11	--	705					
96	3	0.63	--	723					
97	4	0.29	--	711					
105	4	-0.20	--	694					
109	3	0.68	--	725					
113	4	-0.20	--	694					
114	0	-7.16	--	450					
127	4	0.34	--	713					
134	4	0.06	--	703					
138	3	-0.66	--	678					
140	4	0.31	--	712					
142	4	-0.09	--	698					
143	4	0.29	--	711					
146	3	-0.86	--	671					
149	4	-0.09	--	698					
151	4	0.26	--	710					
154	2	-1.46	--	650					
190	4	-0.14	696	--					
212	4	0.46	--	717					
227	4	0.43	--	716					
234	4	-0.31	--	690					
236	3	-0.54	--	682					
247	0	-2.23	--	623					
257	4	-0.20	--	694					
259	4	-0.20	694	--					
273	4	0.20	--	708					

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: F (Fluoride) Concentration Unit : mg/L



SUMMARY

N =	1	25	2	32
Minimum =	2.18	0.18	1.90	1.80
Maximum =		2.90	2.26	21.40
Median =		2.12		2.20
F-pseudosigma =		0.16		0.13


- 0. Other
- 7. Ion chromatography
- 22. Colorimetric
- 40. Ion selective electrode

MPV =	2.18
F-pseudosigma =	0.13
N =	60
Uh =	2.28
Lh =	2.10

Lab	Rating	Z-value	0	7	22	40
1	4	0.07	--	--	--	2.19
3	0	-2.32	--	--	--	1.87
5	0	-14.99	--	0.18	--	--
10	4	0.15	--	--	--	2.20
11	0	13.64	--	--	--	4.00
12	0	-2.85	--	--	--	1.80
13	2	-1.42	--	1.99	--	--
23	3	-0.60	--	2.10	--	--
25	3	0.75	--	2.28	--	--
46	3	0.67	--	--	--	2.27
50	4	0.30	--	--	--	2.22
57	3	-0.60	--	--	--	2.10
59	3	-0.90	--	--	--	2.06
69	3	0.75	--	2.28	--	--
70	4	0.15	--	--	--	2.20
76	4	-0.39	--	--	--	2.13
81	2	1.50	--	2.38	--	--
84	4	-0.37	--	2.13	--	--
85	3	-0.60	--	--	--	2.10
89	3	0.90	--	--	--	2.30
93	3	-0.60	--	2.10	--	--
96	3	-0.97	--	--	--	2.05
97	4	0.00	--	--	--	2.18
102	3	-0.60	--	2.10	--	--
105	2	-1.35	--	2.00	--	--
107	4	0.00	--	--	--	2.18
109	3	0.52	--	--	--	2.25
113	3	0.60	--	--	--	2.26
127	1	1.87	--	2.43	--	--
131	3	-0.52	--	2.11	--	--
134	3	-0.52	--	--	--	2.11
138	1	1.65	--	--	--	2.40
140	2	1.50	--	--	--	2.38
142	3	0.75	--	--	--	2.28
145	3	0.97	--	2.31	--	--
146	4	-0.07	--	--	--	2.17
149	0	-2.10	--	1.90	--	--
154	3	0.60	--	--	--	2.26
180	0	2.40	--	2.50	--	--
183	3	-0.60	--	--	--	2.10
190	0	144.04	--	--	--	21.40
191	0	-2.25	--	1.88	--	--
212	4	-0.22	--	--	--	2.15
220	3	0.60	--	2.26	--	--
227	4	0.00	2.18	--	--	--
234	3	0.82	--	2.29	--	--
236	0	-4.65	--	1.56	--	--
247	4	-0.45	--	2.12	--	--
255	1	1.65	--	--	--	2.40
259	4	-0.22	--	--	--	2.15

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: K (Potassium) Concentration Unit : mg/L




SUMMARY

N =	12	42	7	1	5	1. AA: direct, air	MPV =	9.08
Minimum =	8.35	3.83	8.71	8.84	7.90	4. ICP	F-pseudosigma =	0.62
Maximum =	11.30	15.00	11.88		13.00	6. ICP/MS	N =	67
Median =	8.91	9.10	9.33			7. Ion chromatography	Uh =	9.53
F-pseudosigma =	0.64	0.67	0.59			12. Flame emission	Lh =	8.69

Lab	Rating	Z-value	1	4	6	7	12	Lab	Rating	Z-value	1	4	6	7	12
1	4	-0.26	8.92	--	--	--	--	257	0	6.33	--	--	--	--	13.00
3	3	0.61	--	9.46	--	--	--	259	4	-0.13	--	--	9.00	--	--
5	3	-0.65	--	8.68	--	--	--	265	4	-0.37	--	8.85	--	--	--
10	4	-0.29	8.90	--	--	--	--	270	3	0.68	--	9.50	--	--	--
11	1	1.65	--	10.10	--	--	--	273	4	-0.45	--	8.80	--	--	--
13	4	0.10	--	9.14	--	--	--	277	4	0.36	--	9.30	--	--	--
19	3	0.63	--	9.47	--	--	--	279	0	3.59	11.30	--	--	--	--
23	1	-1.91	--	--	--	--	7.90	292	4	-0.45	8.80	--	--	--	--
24	4	0.03	--	9.10	--	--	--	296	4	0.40	--	--	9.33	--	--
25	4	-0.26	--	8.92	--	--	--	297	3	-0.57	--	8.73	--	--	--
26	4	0.10	--	9.14	--	--	--	302	4	-0.39	--	--	--	8.84	--
38	4	-0.19	8.96	--	--	--	--	305	0	4.52	--	--	11.88	--	--
42	4	0.23	--	9.22	--	--	--	309	3	0.84	--	9.60	--	--	--
46	4	-0.23	--	8.94	--	--	--	323	2	1.18	--	9.81	--	--	--
48	2	1.05	--	--	9.73	--	--	324	2	-1.18	8.35	--	--	--	--
50	2	-1.10	--	8.40	--	--	--	330	4	0.19	--	9.20	--	--	--
57	0	9.56	--	15.00	--	--	--	332	2	1.39	--	--	9.94	--	--
59	3	-0.60	--	--	8.71	--	--								
69	4	0.06	--	--	--	--	9.12								
70	4	0.03	--	9.10	--	--	--								
81	0	3.59	--	11.30	--	--	--								
83	3	-0.78	--	8.60	--	--	--								
85	3	0.71	9.52	--	--	--	--								
86	4	0.45	--	9.36	--	--	--								
87	3	-0.76	8.61	--	--	--	--								
89	3	-0.89	--	--	--	--	8.53								
93	2	-1.15	--	8.37	--	--	--								
102	1	-1.74	--	8.00	--	--	--								
105	2	-1.50	--	8.15	--	--	--								
109	3	0.73	9.53	--	--	--	--								
113	3	-0.84	--	8.56	--	--	--								
127	1	1.65	--	10.10	--	--	--								
131	0	2.13	--	10.40	--	--	--								
134	3	-0.61	8.70	--	--	--	--								
138	3	-0.53	--	8.75	--	--	--								
140	2	-1.02	8.45	--	--	--	--								
142	3	0.84	--	9.60	--	--	--								
145	3	-0.86	--	8.55	--	--	--								
146	2	1.18	--	9.81	--	--	--								
180	1	-1.89	--	7.91	--	--	--								
191	4	0.00	--	--	9.08	--	--								
193	2	1.26	9.86	--	--	--	--								
203	4	0.03	--	--	--	--	9.10								
209	0	4.15	--	11.65	--	--	--								
212	4	-0.11	--	9.01	--	--	--								
220	0	-8.48	--	3.83	--	--	--								
234	3	-0.82	--	8.57	--	--	--								
236	2	-1.26	--	8.30	--	--	--								
247	4	-0.08	--	9.03	--	--	--								
254	4	0.15	--	9.17	--	--	--								

Table 19. Statistical summary of reported data for standard reference water sample GWM-4 (ground-water major constituents)--Continue

Mg (Magnesium) mg/L



0. Other	4. ICP
1. AA: direct, air	6. ICP/MS
2. AA: direct, nitrous oxide	20. Titrate: colorimetric
N =	6 12 2 34 2 5
Minimum =	18.3 16.8 19.3 17.1 19.1 17.4
Maximum =	21.3 25.5 19.6 21.8 19.4 20.1
Median =	18.7
F-pseudostigma =	0.6 1.0


MPV = 18.8
 F-pseudostigma = 0.9
 N = 61
 Hu = 19.5
 Hl = 18.3

Lab	Rating	Z-value	0	1	2	4	6	20
1	4	-0.45			18.4			
12	0	3.37			21.8			
13	4	-0.34			18.5			
18	2	1.35					20.0	
23	3	0.79	19.5					
24	3	0.56			19.3			
25	2	1.01			19.7			
30.1	1	1.57			20.2			
33	0	2.81	21.3					
45	4	0.11	18.9					
46	2	-1.12			17.8			
48	3	0.67				19.4		
59	4	0.34				19.1		
64	3	-0.56			18.3			
69	3	-0.79	18.1					
76	4	-0.27	18.6					
81	2	1.01			19.7			
83	3	-0.79			18.1			
87	0	-2.25	16.8					
89	3	-0.56	18.3					
93	3	0.90		19.6				
102	4	0.00			18.8			
109	4	0.18	19.0					
131	3	-0.90			18.0			
134	3	-0.73			18.2			
138	3	-0.90			18.0			
140	4	-0.11	18.7					
141	1	-1.91			17.1			
142	2	1.24			19.9			
145	4	0.17			19.0			
146	2	1.24			19.9			
151	4	-0.11	18.7					
180	4	0.00			18.8			
190	3	0.67	19.4					
196	4	-0.22	18.6					
212	3	-0.90			18.0			
215	4	0.34			19.1			
219	3	-0.56			18.3			
220	2	-1.35			17.6			
227	0	2.47			21.0			
234	3	-0.56			18.3			
235	3	0.67			19.4			
236	3	-0.83			18.1			
247	4	-0.22			18.6			
254	4	-0.11			18.7			
255	4	0.45			19.2			
256	2	1.42					20.1	
257	2	1.35					20.0	
258	4	-0.06	18.8					
261	0	7.52	25.5					

Lab	Rating	Z-value	0	1	2	4	6	20
262	4	-0.36	18.5					
263	3	0.79	19.5					
265	4	0.22			19.0			
268	1	-2.02	17.0					
270	3	-0.79			18.1			
273	1	1.80			20.4			
274	3	0.57					19.3	
276	3	0.56		19.3				
279	1	-1.57						17.4
289	3	-0.79				18.1		
292	4	0.45	19.2					

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: Na (Sodium) Concentration Unit : mg/L




SUMMARY

N =	15	47	7	1	4	1. AA: direct, air	MPV =	139
Minimum =	117	109	107	135	131	4. ICP	F-pseudosigma =	6
Maximum =	144	179	171		139	6. ICP/MS	Rating criterion =	7
Median =	139	138	140			7. Ion chromatography	N =	74
F-pseudosigma =	5	6	14			12. Flame emission	Uh =	142
							Lh =	134

Lab	Rating	Z-value	1	4	6	7	12	Lab	Rating	Z-value	1	4	6	7	12
1	3	-0.48	--	135	--	--	--	212	3	0.65	--	143	--	--	--
3	4	-0.07	--	138	--	--	--	220	4	-0.22	--	137	--	--	--
5	3	-0.65	--	134	--	--	--	234	4	-0.22	--	137	--	--	--
10	4	0.03	139	--	--	--	--	236	2	-0.87	--	133	--	--	--
11	0	-4.26	--	109	--	--	--	247	4	0.22	--	140	--	--	--
12	3	-0.51	--	135	--	--	--	254	4	0.14	--	140	--	--	--
13	3	0.51	--	142	--	--	--	257	2	-1.08	--	--	--	--	131
19	3	0.65	--	143	--	--	--	259	3	0.43	--	--	142	--	--
23	3	0.51	142	--	--	--	--	265	4	0.36	--	141	--	--	--
24	3	0.65	--	143	--	--	--	268	3	0.51	142	--	--	--	--
25	3	-0.65	--	134	--	--	--	273	2	-1.08	--	131	--	--	--
26	3	0.59	--	143	--	--	--	277	4	-0.07	--	138	--	--	--
38	4	0.38	141	--	--	--	--	279	0	-3.10	117	--	--	--	--
42	3	0.51	--	142	--	--	--	292	2	-0.87	133	--	--	--	--
46	3	0.65	--	143	--	--	--	296	0	2.96	--	--	159	--	--
48	0	4.69	--	--	171	--	--	297	4	-0.01	--	138	--	--	--
50	4	-0.07	--	138	--	--	--	302	3	-0.51	--	--	--	135	--
57	2	-1.23	--	130	--	--	--	305	0	-4.61	--	--	107	--	--
59	4	-0.22	137	--	--	--	--	307	4	-0.22	137	--	--	--	--
69	4	-0.36	--	--	--	--	136	309	2	1.23	--	147	--	--	--
70	2	0.94	--	145	--	--	--	323	2	-1.08	--	131	--	--	--
76	4	0.17	--	--	140	--	--	324	3	0.82	144	--	--	--	--
81	4	0.36	--	141	--	--	--	330	0	5.85	--	179	--	--	--
83	2	-1.00	--	132	--	--	--	332	1	-1.46	--	--	128	--	--
84	4	-0.36	--	--	--	--	136								
85	3	0.51	142	--	--	--	--								
86	3	0.79	--	144	--	--	--								
87	3	0.51	142	--	--	--	--								
89	4	0.07	--	--	--	--	139								
93	4	0.07	--	139	--	--	--								
97	4	0.22	140	--	--	--	--								
102	0	-4.12	--	110	--	--	--								
105	4	0.36	--	141	--	--	--								
109	4	0.01	139	--	--	--	--								
113	4	-0.14	--	138	--	--	--								
121	4	-0.07	--	138	--	--	--								
127	4	0.36	--	141	--	--	--								
131	3	0.51	--	142	--	--	--								
134	3	-0.65	134	--	--	--	--								
138	4	0.07	--	139	--	--	--								
140	2	-0.94	132	--	--	--	--								
142	1	1.66	--	150	--	--	--								
145	4	-0.07	--	138	--	--	--								
146	3	0.51	--	142	--	--	--								
154	3	-0.79	--	133	--	--	--								
180	0	-2.09	--	124	--	--	--								
191	3	-0.51	--	--	135	--	--								
193	3	-0.65	--	134	--	--	--								
203	2	-0.94	--	132	--	--	--								
209	0	3.00	--	159	--	--	--								

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: pH



ANALYTICAL METHOD CODE

—○— 41

SUMMARY

N = 79
 Minimum = 7.29
 Maximum = 8.60
 Median = 8.33
 F-pseudostigma = 0.13


41. Direct reading

MPV = 8.33
 F-pseudostigma = 0.13
 Rating criterion = 0.42
 N = 79
 Uh = 8.41
 Lh = 8.24


Lab	Rating	Z-value	41	Lab	Rating	Z-value	41
1	4	0.10	8.37	203	0	-0.86	7.97
3	3	0.17	8.40	209	1	0.46	8.52
5	0	-2.50	7.29	212	3	0.19	8.41
10	3	0.22	8.42	213	4	-0.10	8.29
11	2	-0.41	8.16	220	4	0.00	8.33
12	2	-0.31	8.20	227	2	-0.41	8.16
13	4	0.14	8.39	234	0	0.62	8.59
19	4	0.12	8.38	236	3	-0.19	8.25
23	3	0.17	8.40	243	4	-0.02	8.32
24	3	0.17	8.40	244	2	0.36	8.48
25	3	0.19	8.41	247	3	0.26	8.44
26	2	0.34	8.47	257	3	-0.19	8.25
38	3	0.17	8.40	259	3	0.17	8.40
42	4	-0.14	8.27	267	3	0.17	8.40
43	4	-0.10	8.29	268	2	-0.41	8.16
46	4	-0.10	8.29	269	3	0.17	8.40
48	2	-0.31	8.20	270	0	-0.79	8.00
50	3	-0.29	8.21	273	4	0.10	8.37
57	4	-0.07	8.30	277	2	-0.31	8.20
59	3	-0.19	8.25	279	1	-0.50	8.12
69	4	-0.07	8.30	291	1	-0.46	8.14
70	2	0.31	8.46	292	3	0.19	8.41
76	2	0.39	8.49	302	4	-0.07	8.30
81	2	0.36	8.48	305	3	0.17	8.40
84	3	0.17	8.40	307	4	-0.02	8.32
85	4	-0.07	8.30	309	4	0.00	8.33
86	4	0.02	8.34	323	3	0.17	8.40
87	4	0.00	8.33	324	0	-1.90	7.54
89	2	0.36	8.48	330	2	-0.38	8.17
93	4	0.00	8.33				
96	1	-0.46	8.14				
97	2	0.36	8.48				
105	4	-0.07	8.30				
109	2	-0.36	8.18				
113	3	-0.24	8.23				
114	0	-0.72	8.03				
127	3	0.17	8.40				
134	4	0.05	8.35				
138	2	0.36	8.48				
140	0	-2.28	7.38				
142	3	0.19	8.41				
143	2	0.36	8.48				
145	4	-0.07	8.30				
146	4	0.10	8.37				
149	0	0.65	8.60				
151	3	0.22	8.42				
154	3	0.24	8.43				
180	2	0.41	8.50				
183	4	-0.05	8.31				
190	2	-0.31	8.20				

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: SiO₂ (Silica) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	27	3	1	22	4. ICP	MPV =	10.8		
Minimum =	4.7	9.0	11.6	5.5	6. ICP/MS	F-pseudosigma =	0.8		
Maximum =	11.9	11.2		12.6	22. Colorimetric	N =	53		
Median =	10.5			11.0	22mb. Color: molybdate blue	Uh =	11.3		
F-pseudosigma =	0.8			0.7		Lh =	10.2		

Lab	Rating	Z-value	4	6	22	22mb	Lab	Rating	Z-value	4	6	22	22mb
1	4	0.13	--	--	--	10.9	323	4	0.00	10.8	--	--	--
3	2	-1.29	9.8	--	--	--	330	4	-0.25	--	--	--	10.6
10	4	-0.37	--	--	--	10.5	332	0	-2.22	--	9.0	--	--
11	1	-1.88	9.3	--	--	--							
13	4	0.12	10.9	--	--	--							
23	3	0.74	--	--	--	11.4							
24	3	0.74	11.4	--	--	--							
25	2	-1.48	9.6	--	--	--							
26	3	0.69	11.4	--	--	--							
38	4	0.09	--	--	--	10.9							
42	4	-0.25	10.6	--	--	--							
50	4	-0.39	10.5	--	--	--							
57	4	0.12	10.9	--	--	--							
59	3	0.61	--	--	--	11.3							
70	1	1.59	--	--	--	12.1							
83	3	-0.74	10.2	--	--	--							
87	0	-6.55	--	--	--	5.5							
89	4	0.25	--	--	--	11.0							
93	2	-1.40	--	--	--	9.7							
97	3	-0.74	--	--	--	10.2							
102	0	2.16	--	--	--	12.6							
105	2	-1.07	9.9	--	--	--							
107	4	0.49	--	--	--	11.2							
113	4	0.05	--	--	--	10.8							
121	4	-0.12	10.7	--	--	--							
127	2	-1.28	9.8	--	--	--							
131	1	-1.74	9.4	--	--	--							
134	3	-0.52	10.4	--	--	--							
138	4	0.00	--	--	--	10.8							
140	3	0.76	--	--	--	11.4							
142	2	1.35	11.9	--	--	--							
145	4	-0.37	10.5	--	--	--							
151	4	-0.49	--	--	--	10.4							
154	4	0.49	11.2	--	--	--							
190	2	1.35	--	--	--	11.9							
191	4	0.49	--	11.2	--	--							
193	3	0.98	--	--	11.6	--							
203	2	1.39	--	--	--	11.9							
204	1	1.59	--	--	--	12.1							
212	4	0.00	10.8	--	--	--							
218	2	1.17	11.8	--	--	--							
227	4	0.37	11.1	--	--	--							
234	3	-0.61	10.3	--	--	--							
236	0	-2.45	8.8	--	--	--							
247	2	1.10	--	--	--	11.7							
254	4	0.05	10.8	--	--	--							
259	4	0.00	--	10.8	--	--							
265	4	-0.37	10.5	--	--	--							
297	1	-1.88	--	--	--	9.3							
309	0	-7.48	4.7	--	--	--							

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: SO₄ (Sulfate) Concentration Unit : mg/L



SUMMARY


N =	1	1	46	12	3	11	0. Other	MPV =	164
Minimum =	156	162	120	152	163	150	4. ICP	F-pseudostigma =	5
Maximum =			178	175	167	190	7. Ion chromatography	Rating criterion =	8
Median =			164	168		165	22. Colorimetric	N =	74
F-pseudostigma =			4	6		9	50. Gravimetric	Uh =	168
							51. Turbidimetric	Lh =	161

Lab	Rating	Z-value	0	4	7	22	50	51	Lab	Rating	Z-value	0	4	7	22	50	51
1	4	-0.32	--	--	161	--	--	--	209	3	-0.98	--	--	156	--	--	--
3	4	0.49	--	--	--	168	--	--	212	4	-0.24	--	--	162	--	--	--
4	4	-0.49	--	--	160	--	--	--	220	4	0.02	--	--	164	--	--	--
5	3	0.98	--	--	172	--	--	--	227	3	-0.63	--	--	159	--	--	--
10	2	-1.34	--	--	--	--	--	153	234	4	0.12	--	--	165	--	--	--
11	0	-2.20	--	--	146	--	--	--	236	4	0.00	--	--	164	--	--	--
12	4	-0.24	--	--	--	162	--	--	247	2	1.10	--	--	173	--	--	--
13	4	0.49	--	--	168	--	--	--	254	4	0.00	--	--	164	--	--	--
23	1	1.71	--	--	178	--	--	--	255	3	-0.98	--	--	--	156	--	--
24	4	0.49	--	--	--	168	--	--	257	3	-0.78	--	--	158	--	--	--
25	4	0.12	--	--	165	--	--	--	259	4	0.00	--	--	164	--	--	--
43	4	0.37	--	--	--	--	167	--	265	4	0.24	--	--	166	--	--	--
46	2	1.34	--	--	--	175	--	--	268	4	0.02	--	--	164	--	--	--
48	2	1.10	--	--	--	--	--	173	270	0	-5.37	--	--	120	--	--	--
50	3	-0.61	--	--	--	159	--	--	273	3	-0.93	156	--	--	--	--	--
57	3	0.73	--	--	--	170	--	--	277	4	0.41	--	--	167	--	--	--
59	4	0.00	--	--	164	--	--	--	292	4	0.35	--	--	167	--	--	--
69	4	0.37	--	--	167	--	--	--	297	2	1.02	--	--	--	172	--	--
70	4	-0.37	--	--	161	--	--	--	302	1	-1.95	--	--	148	--	--	--
81	4	0.49	--	--	--	168	--	--	307	4	0.37	--	--	--	--	--	167
83	4	-0.24	--	162	--	--	--	--	309	4	0.00	--	--	--	--	--	164
84	4	0.00	--	--	164	--	--	--	323	4	-0.49	--	--	160	--	--	--
85	4	0.24	--	--	166	--	--	--	324	4	-0.06	--	--	--	--	164	--
86	4	-0.24	--	--	162	--	--	--	330	4	-0.24	--	--	--	162	--	--
87	4	0.12	--	--	--	--	--	165									
89	4	0.00	--	--	164	--	--	--									
93	3	0.61	--	--	169	--	--	--									
96	3	-0.61	--	--	--	--	--	159									
97	2	-1.46	--	--	--	152	--	--									
102	3	-0.55	--	--	160	--	--	--									
105	4	-0.24	--	--	162	--	--	--									
109	4	-0.13	--	--	--	--	163	--									
113	3	-0.51	--	--	160	--	--	--									
114	3	-0.85	--	--	--	--	--	157									
127	3	0.85	--	--	171	--	--	--									
131	4	-0.12	--	--	163	--	--	--									
134	4	-0.09	--	--	163	--	--	--									
138	4	0.49	--	--	168	--	--	--									
140	0	3.17	--	--	--	--	--	190									
142	1	1.59	--	--	177	--	--	--									
145	4	-0.37	--	--	161	--	--	--									
146	4	0.37	--	--	--	--	--	167									
149	4	-0.12	--	--	163	--	--	--									
154	2	1.34	--	--	--	--	--	175									
180	2	1.22	--	--	174	--	--	--									
183	1	-1.74	--	--	--	--	--	150									
191	4	0.12	--	--	165	--	--	--									
203	4	0.00	--	--	164	--	--	--									
204	4	0.49	--	--	--	168	--	--									
208	4	-0.02	--	--	164	--	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: Sp Cond (Specific Conductance)

Concentration Unit : mS/cm



SUMMARY

N = 75
 Minimum = 1
 Maximum = 1350
 Median = 1124
 F-pseudostigma = 23


41. Direct reading

MPV = 1124
 F-pseudostigma = 23
 Rating criterion = 56
 N = 75
 Uh = 1140
 Lh = 1110


Lab	Rating	Z-value	41	Lab	Rating	Z-value	41
1	4	0.27	1139	193	4	0.46	1150
3	4	0.28	1140	203	4	0.16	1133
5	4	0.39	1146	212	4	-0.07	1120
10	4	0.16	1133	220	3	-0.71	1084
11	3	-0.60	1090	227	4	0.07	1128
12	4	0.11	1130	234	0	-6.92	735
13	4	0.36	1144	236	0	4.02	1350
19	3	0.82	1170	243	4	0.11	1130
23	4	-0.34	1105	244	4	-0.09	1119
24	4	0.28	1140	247	4	0.32	1142
25	4	-0.12	1117	257	2	1.21	1192
26	4	-0.21	1112	259	4	0.28	1140
38	4	0.11	1130	267	4	0.14	1132
42	4	-0.39	1102	268	4	-0.07	1120
43	4	0.11	1130	269	4	0.21	1136
46	4	-0.43	1100	273	4	0.28	1140
48	3	0.94	1177	277	4	-0.09	1119
50	4	0.11	1130	279	0	-19.98	1
57	4	-0.43	1100	292	4	-0.18	1114
59	4	0.07	1128	297	4	-0.07	1120
70	3	-0.96	1070	302	4	0.28	1140
76	4	-0.20	1113	307	4	0.41	1147
81	4	-0.25	1110	323	4	-0.20	1113
84	4	0.02	1125	324	0	-11.23	493
85	3	-1.00	1068	330	4	0.21	1136
86	3	0.68	1162				
87	0	-14.73	296				
89	3	0.64	1160				
93	4	-0.46	1098				
96	4	-0.07	1120				
97	4	0.00	1124				
102	3	0.80	1169				
105	3	0.53	1154				
109	4	-0.34	1105				
113	4	0.12	1131				
114	4	-0.27	1109				
127	4	-0.25	1110				
134	4	-0.03	1123				
138	4	-0.07	1120				
140	3	-0.78	1080				
142	4	0.27	1139				
143	4	-0.04	1122				
145	4	-0.20	1113				
146	4	-0.43	1100				
149	4	0.28	1140				
151	3	-0.84	1077				
154	4	0.41	1147				
180	4	-0.07	1120				
183	4	-0.27	1109				
190	4	0.28	1140				

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: Sr (Strontium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY					
N =	29	6	4. ICP	MPV =	596
Minimum =	496	1	6. ICP/MS	F-pseudosigma =	29
Maximum =	684	792		Rating criterion =	30
Median =	595			N =	35
F-pseudosigma =	27			Uh =	617
				Lh =	578


Lab	Rating	Z-value	4	6
1	4	0.19	602	--
3	1	-1.54	550	--
5	4	-0.03	595	--
11	3	-0.84	571	--
24	2	1.21	632	--
25	4	0.30	605	--
42	3	0.54	612	--
57	3	-0.57	579	--
70	3	0.54	612	--
81	4	-0.03	595	--
86	3	0.70	617	--
102	0	-19.98	--	1
105	1	-1.78	543	--
113	2	-1.21	560	--
121	3	-0.54	580	--
127	4	0.00	596	--
131	0	-2.92	509	--
134	4	-0.10	593	--
138	4	-0.44	583	--
142	2	1.04	627	--
145	4	-0.34	586	--
151	0	6.58	--	792
154	4	0.00	596	--
191	3	0.70	--	617
212	1	1.68	646	--
234	3	-0.67	576	--
236	2	-1.01	566	--
247	2	1.01	626	--
254	3	0.54	612	--
259	4	0.47	--	610
265	3	0.87	622	--
273	0	-3.36	496	--
323	0	2.95	684	--
330	0	2.35	--	666
332	4	0.10	--	599

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: total P as P (total Phosphorus as Phosphorus) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	2	7	3	50	0. Other	MPV =	1.31		
Minimum =	1.27	1.26	0.80	0.73	4. ICP	F-pseudostigma =	0.06		
Maximum =	1.28	1.36	1.30	1.68	7. Ion chromatography	Rating criterion =	0.07		
Median =		1.31		1.31	22m. Color:phosphomolybdate	N =	62		
F-pseudostigma =		0.01		0.07		Uh =	1.34		
						Lh =	1.26		

Lab	Rating	Z-value	0	4	7	22m	Lab	Rating	Z-value	0	4	7	22m
3	1	1.83	--	--	--	1.43	255	0	-0.31	--	--	--	1.29
5	4	-0.41	1.28	--	--	--	257	0	-8.89	--	--	--	0.73
11	4	0.31	--	--	--	1.33	259	3	0.46	--	--	--	1.34
12	1	-0.15	--	--	--	1.30	265	4	-0.15	--	--	1.30	--
13	2	5.62	--	--	--	1.68	267	4	-1.37	--	--	--	1.22
23	1	-0.15	--	--	--	1.30	273	0	-3.66	--	--	--	1.07
25	3	-1.22	--	--	--	1.23	292	2	0.92	--	--	--	1.37
38	2	1.04	--	--	--	1.38	297	0	-0.46	--	--	--	1.28
42	4	0.00	--	1.31	--	--	305	4	0.46	--	--	--	1.34
46	3	0.92	--	--	--	1.37	307	2	-0.11	--	--	--	1.30
48	1	-0.31	--	--	--	1.29	323	3	0.61	--	--	--	1.35
57	1	1.37	--	--	--	1.40	330	4	0.00	--	1.31	--	--
59	4	0.46	--	--	--	1.34							
70	3	0.46	--	--	--	1.34							
81	3	0.61	--	--	--	1.35							
83	4	0.00	--	1.31	--	--							
85	4	-1.37	--	--	--	1.22							
86	4	0.15	--	1.32	--	--							
87	3	-1.22	--	--	--	1.23							
89	4	-1.68	--	--	--	1.20							
97	3	-0.92	--	--	--	1.25							
102	4	0.15	--	--	--	1.32							
105	4	0.46	--	--	--	1.34							
113	4	0.15	--	--	--	1.32							
114	3	0.00	--	--	--	1.31							
127	3	0.76	--	1.36	--	--							
131	0	-3.97	--	--	1.05	--							
134	2	1.22	--	--	--	1.39							
138	2	1.22	--	--	--	1.39							
140	2	-3.05	--	--	--	1.11							
142	4	-2.44	--	--	--	1.15							
143	4	-1.37	--	--	--	1.22							
145	1	1.53	--	--	--	1.41							
146	4	-2.44	--	--	--	1.15							
149	0	-4.73	--	--	--	1.00							
154	0	2.44	--	--	--	1.47							
180	3	0.00	--	--	--	1.31							
183	2	1.01	--	--	--	1.38							
190	3	0.00	--	--	--	1.31							
191	3	-0.61	1.27	--	--	--							
203	4	0.15	--	--	--	1.32							
208	0	-7.79	--	--	0.80	--							
212	1	-0.15	--	--	--	1.30							
213	1	1.37	--	--	--	1.40							
220	0	-0.61	--	--	--	1.27							
227	3	-0.76	--	1.26	--	--							
234	0	-0.32	--	--	--	1.29							
236	4	-0.15	--	1.30	--	--							
243	2	-0.76	--	--	--	1.26							
247	4	0.46	--	--	--	1.34							

Table 12. Statistical summary of reported data for standard reference water sample M-154 (major constituents)--Continued

Analyte: V (Vanadium) Concentration Unit : mg/L



SUMMARY

N =	2	20	11
Minimum =	17.6	14.4	14.6
Maximum =	25.7	21.0	24.5
Median =		16.6	18.2
F-pseudosigma =		1.1	2.8

- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	16.9
F-pseudosigma =	1.8
N =	33
Uh =	18.2
Lh =	15.8

Lab	Rating	Z-value	3	4	6
1	4	0.00	--	--	16.9
3	3	-0.62	--	15.8	--
5	4	-0.17	--	16.6	--
11	2	-1.07	--	15.0	--
13	NR	--	--	< 50	--
24	3	-0.51	--	16.0	--
25	0	-2.17	--	< 13	--
42	3	0.73	--	--	18.2
46	2	-1.41	--	14.4	--
48	2	-1.18	--	--	14.8
57	4	-0.34	--	16.3	--
59	2	-1.29	--	--	14.6
70	NR	--	--	< 50	--
81	2	-1.07	--	15.0	--
86	0	4.27	--	--	24.5
89	0	4.95	25.7	--	--
105	3	-0.62	--	--	15.8
127	4	0.39	17.6	--	--
134	3	-0.76	--	15.6	--
138	4	-0.17	--	16.6	--
142	2	1.12	--	--	18.9
145	0	2.30	--	21.0	--
146	4	0.11	--	17.1	--
154	4	0.06	--	17.0	--
180	4	-0.34	--	16.3	--
212	4	-0.11	--	16.7	--
220	4	0.24	--	17.3	--
234	4	0.22	--	17.3	--
236	1	1.74	--	20.0	--
247	1	1.57	--	--	19.7
265	4	0.34	--	17.5	--
296	0	2.36	--	--	21.1
297	4	-0.28	--	16.4	--
305	3	-0.67	--	--	15.7
323	2	1.07	--	18.8	--
330	2	1.35	--	--	19.3

Table 13. *Statistical summary of reported data for standard reference sample N-65 (nutrient constituents)*


Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported		
4. ICP	=	inductively coupled plasma
5. DCP	=	direct coupled plasma
7. IC	=	ion chromatography
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode

<u>Abbreviations and figure symbols</u>		
	N =	number of analyses--(excluding less than values)
	MPV =	most probable value ---
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit - - - -
Lwl	=	lower warning limit - - - -
Ucl	=	upper warning limit
Lcl	=	lower warning limit - - - -
mg/L	=	milligrams per liter
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported


<u>Constituent</u>		<u>page</u>
NH ₃ as N	Ammonia as nitrogen	82
NH ₃ +Org N as N	Ammonia plus organic nitrogen as nitrogen	83
NO ₃ as N	Nitrate as nitrogen	84
Total P as P	Total Phosphorus as phosphorus	85
PO ₄ as P	Orthophosphate as phosphorus	86

Table 13. Statistical summary of reported data for standard reference water sample N-65 (nutrient constituents)--Continued

Analyte: NH₃ as N (Ammonia as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =	2	0	9	1	50	7	0. Other	MPV =	0.124								
Minimum =	0.143	0.060	0.069	0.060	0.115	0.20	Titrate: colorimetric	F-pseudosigma =	0.023								
Maximum =	0.180	0.480	0.625	0.243	0.22n	Color: Nesslerization	N =	69									
Median =		0.131	0.121	0.140	0.22p	Color: phenate	Hu =	0.142									
F-pseudosigma =		0.034	0.019	0.018	0.40	Ion selective electrode	HI =	0.111									
Lab	Rating	Z-value	0	20	22	22N	22p	40	Lab	Rating	Z-value	0	20	22	22N	22p	40
1	4	-0.22	--	--	--	--	0.119	--	213	NR	--	--	< 1	--	--	--	--
3	0	-2.38	--	--	--	--	0.070	--	220	0	15.69	--	--	0.480	--	--	--
10	4	-0.18	--	--	--	--	0.120	--	224	0	2.60	--	--	--	--	0.183	--
11	0	-2.82	--	--	--	--	0.060	--	227	4	-0.18	--	--	0.120	--	--	--
13	0	2.25	--	--	--	--	0.175	--	234	4	-0.40	--	--	--	--	--	0.115
21	4	0.00	--	--	--	--	0.124	--	243	3	-0.62	--	--	--	--	0.110	--
23	1	-1.94	--	--	--	--	0.080	--	247	3	-0.75	--	--	--	--	0.107	--
25	3	-0.62	--	--	0.110	--	--	--	253	0	3.84	--	--	0.211	--	--	--
28	0	9.52	--	--	--	--	0.340	--	255	2	1.15	--	--	--	--	0.150	--
31	4	0.00	--	--	--	--	0.124	--	287	0	5.25	--	--	--	--	--	0.243
38	4	0.18	--	--	--	--	0.128	--	292	3	0.71	--	--	--	--	--	0.140
46	3	-0.57	--	--	--	--	0.111	--	297	3	0.66	--	--	--	--	0.139	--
51	4	0.26	--	--	--	--	0.130	--	305	2	1.50	--	--	--	--	--	0.158
59	2	-1.06	--	--	--	--	0.100	--	313	2	1.50	--	--	--	--	0.158	--
70	3	-0.97	--	--	--	--	0.102	--	314	3	0.84	0.143	--	--	--	--	--
72	1	-1.94	--	--	--	--	0.080	--	316	4	0.31	--	--	--	--	0.131	--
76	4	-0.30	--	--	--	--	0.117	--	318	4	-0.18	--	--	--	--	0.120	--
81	3	-0.66	--	--	--	--	0.109	--	319	0	2.47	0.180	--	--	--	--	--
83	4	-0.18	--	--	--	--	0.120	--	320	3	0.97	--	--	--	--	0.146	--
85	3	-0.57	--	--	--	--	0.111	--	321	4	-0.18	--	--	--	--	0.120	--
86	3	0.71	--	--	--	--	0.140	--									
87	0	5.55	--	--	--	--	0.250	--									
89	4	-0.40	--	--	--	--	0.115	--									
90	3	-0.53	--	--	--	--	0.112	--									
93	3	-0.62	--	--	--	--	0.110	--									
96	4	0.48	--	--	--	--	0.135	--									
97	4	0.31	--	--	0.131	--	--	--									
102	2	1.15	--	--	--	--	0.150	--									
105	0	-2.82	--	--	0.060	--	--	--									
110	3	0.78	--	--	0.142	--	--	--									
113	4	0.26	--	--	--	--	0.130	--									
114	3	0.71	--	--	--	--	0.140	--									
127	4	0.44	--	--	--	--	0.134	--									
129	0	-2.42	--	--	--	0.069	--	--									
134	4	-0.31	--	--	--	--	0.117	--									
138	2	1.15	--	--	--	--	0.150	--									
140	4	-0.18	--	--	0.120	--	--	--									
142	1	1.85	--	--	0.166	--	--	--									
143	4	0.26	--	--	--	--	0.130	--									
145	3	-0.62	--	--	--	--	0.110	--									
146	4	0.40	--	--	--	--	0.133	--									
154	4	-0.18	--	--	--	--	0.120	--									
158	4	0.09	--	--	--	--	0.126	--									
180	4	0.04	--	--	--	--	0.125	--									
190	0	22.09	--	--	--	--	0.625	--									
193	4	-0.26	--	--	--	--	0.118	--									
198	4	-0.09	--	--	--	--	0.122	--									
203	4	0.26	--	--	--	--	0.130	--									
205	3	-0.84	--	--	--	--	0.105	--									
212	0	-2.34	--	--	--	--	0.071	--									

Table 13. Statistical summary of reported data for standard reference water sample N-65 (nutrient constituents)--Continued
 Analyte: NH₃ + Organic N as N (Ammonia + organic nitrogen as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =	2	15	2	23	3	1	0. Other	MPV =	0.159								
Minimum =	0.114	0.107	0.120	0.059	0.106	0.160	22. Colorimetric	F-pseudostigma =	0.067								
Maximum =	0.230	0.592	0.839	0.517	0.299		22n. Color: Nesslerization	N =	46								
Median =		0.177		0.152			22p. Color: phenate	Hu =	0.220								
F-pseudostigma =		0.058		0.050			22H2SO4. Color: Sulfuric acid	HI =	0.130								
							40. Ion selective electrode										
Lab	Rating	Z-value	0	22	22N	22p	22H2SO4	40	Lab	Rating	Z-value	0	22	22N	22p	22H2SO4	40
1	4	-0.16	--	--	--	0.148	--	--	318	4	0.16	--	--	--	0.170	--	--
3	NR	--	< 1	--	--	--	--	--	319	0	4.21	--	--	--	0.440	--	--
10	4	0.01	--	--	--	0.160	--	--	320	3	0.76	--	0.210	--	--	--	--
11	0	3.61	--	0.400	--	--	--	--									
21	4	0.27	--	0.177	--	--	--	--									
23	NR	--	--	--	--	< 0.1	--	--									
25	2	1.06	0.230	--	--	--	--	--									
31	4	0.27	--	0.177	--	--	--	--									
38	3	-0.58	--	--	0.120	--	--	--									
46	2	1.14	--	--	--	--	0.235	--									
51	4	0.01	--	--	--	--	--	0.160									
59	3	0.61	--	0.200	--	--	--	--									
70	3	-0.57	--	--	--	0.121	--	--									
72	4	-0.06	--	--	--	0.155	--	--									
81	2	-1.50	--	--	--	0.059	--	--									
87	1	1.66	--	--	--	0.270	--	--									
89	3	-0.52	--	--	--	0.124	--	--									
90	4	-0.24	--	0.143	--	--	--	--									
96	NR	--	--	--	--	< 0.15	--	--									
97	4	-0.13	--	0.150	--	--	--	--									
102	3	0.91	--	0.220	--	--	--	--									
105	NR	--	--	< 1	--	--	--	--									
113	NR	--	--	--	--	< 0.5	--	--									
127	4	-0.42	--	--	--	0.131	--	--									
129	0	10.19	--	--	0.839	--	--	--									
134	4	-0.01	--	--	--	0.158	--	--									
138	4	0.46	--	--	--	0.190	--	--									
140	4	-0.43	--	0.130	--	--	--	--									
142	3	-0.67	0.114	--	--	--	--	--									
143	4	0.01	--	--	--	0.160	--	--									
145	3	-0.73	--	--	--	0.110	--	--									
146	4	-0.22	--	--	--	0.144	--	--									
154	4	-0.28	--	--	--	0.140	--	--									
158	4	-0.45	--	--	--	0.129	--	--									
180	4	-0.10	--	--	--	0.152	--	--									
193	3	-0.79	--	--	--	0.106	--	--									
203	4	-0.43	--	0.130	--	--	--	--									
209	3	0.91	--	0.220	--	--	--	--									
212	2	1.32	--	--	--	0.247	--	--									
213	NR	--	--	< 1	--	--	--	--									
220	0	6.49	--	0.592	--	--	--	--									
224	0	5.37	--	--	--	0.517	--	--									
227	4	0.39	--	0.185	--	--	--	--									
247	3	-0.63	--	0.117	--	--	--	--									
253	3	-0.78	--	0.107	--	--	--	--									
297	0	2.10	--	--	--	0.299	--	--									
305	NR	--	--	< 1	--	--	--	--									
313	3	-0.82	--	--	--	0.104	--	--									
314	1	1.56	--	--	--	0.263	--	--									
316	4	-0.40	--	--	--	0.132	--	--									

Table 13. Statistical summary of reported data for standard reference water sample N-65 (nutrient constituents)--Continued
 Analyte: NO₃ as N (Nitrate as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																		
										MPV = 0.037								
										F-pseudosigma = 0.009								
										N = 63								
										Hu = 0.045								
										HI = 0.033								
Lab	Rating	Z-value	0	7	22	22cd	22sulf	40		Lab	Rating	Z-value	0	7	22	22cd	22sulf	40
1	3	-0.76	--	--	--	0.030	--	--		212	0	19.99	--	0.220	--	--	--	--
3	NR	--	--	--	--	< 0.05	--	--		220	0	7.32	--	--	0.104	--	--	--
10	3	0.76	--	--	--	0.044	--	--		224	3	0.87	--	0.045	--	--	--	--
11	0	2.51	--	--	--	--	0.060	--		227	NR	--	--	< 0.199	--	--	--	--
12	3	-0.76	--	--	--	0.030	--	--		234	4	-0.33	--	0.034	--	--	--	--
13	NR	--	--	< 0.05	--	--	--	--		243	1	-1.86	--	--	--	0.020	--	--
21	4	-0.04	--	--	0.037	--	--	--		247	1	-1.86	--	0.020	--	--	--	--
23	4	0.33	--	0.040	--	--	--	--		253	0	-4.04	--	--	0.000	--	--	--
25	0	15.73	--	0.181	--	--	--	--		255	NR	--	--	--	--	< 0.1	--	--
28	0	12.34	--	--	--	--	0.150	--		292	0	10.16	--	0.130	--	--	--	--
31	4	-0.04	--	--	0.037	--	--	--		297	0	3.39	--	--	--	0.068	--	--
38	3	-0.55	--	--	--	0.032	--	--		305	0	23.27	--	--	--	--	--	0.250
42	2	1.20	--	0.048	--	--	--	--		313	3	-0.55	--	--	--	0.032	--	--
45	4	0.00	--	0.037	--	--	--	--		314	4	-0.33	--	--	--	0.034	--	--
46	4	-0.22	--	--	--	0.035	--	--		316	4	-0.09	--	--	--	0.036	--	--
51	1	-1.86	--	0.020	--	--	--	--		318	4	-0.16	--	--	--	0.036	--	--
53	3	-0.76	0.030	--	--	--	--	--		319	4	0.33	--	--	--	0.040	--	--
59	2	1.42	--	--	--	0.050	--	--		320	4	-0.33	--	--	0.034	--	--	--
70	1	-1.97	--	--	--	0.019	--	--		321	3	-0.76	--	0.030	--	--	--	--
72	0	15.62	--	--	--	0.180	--	--										
81	3	0.58	--	--	0.042	--	--	--										
85	4	0.00	--	--	--	0.037	--	--										
87	4	0.11	--	--	--	0.038	--	--										
89	3	0.76	--	--	--	0.044	--	--										
90	4	0.33	--	--	--	0.040	--	--										
96	NR	--	--	--	--	< 0.05	--	--										
97	1	-1.53	--	--	0.023	--	--	--										
102	0	-4.04	--	0.000	--	--	--	--										
105	NR	--	--	--	< 0.04	--	--	--										
113	4	-0.11	--	--	--	0.036	--	--										
127	0	8.41	--	0.114	--	--	--	--										
129	4	-0.33	--	0.034	--	--	--	--										
134	4	-0.11	--	--	--	0.036	--	--										
138	4	0.31	--	--	--	--	--	--										
140	1	-1.86	--	--	0.020	--	--	--										
142	3	0.95	--	--	0.046	--	--	--										
143	4	0.44	--	--	--	0.041	--	--										
145	4	0.33	--	--	--	0.040	--	--										
146	4	-0.07	--	--	--	0.036	--	--										
154	4	-0.11	--	--	--	--	0.036	--										
158	1	-1.64	--	--	0.022	--	--	--										
180	4	0.33	--	0.040	--	--	--	--										
183	0	159.80	--	--	--	--	--	1.500										
190	0	15.95	--	--	--	0.183	--	--										
191	4	-0.33	--	0.034	--	--	--	--										
193	4	0.00	--	--	0.037	--	--	--										
198	3	-0.66	--	--	0.031	--	--	--										
203	4	0.33	--	--	--	0.040	--	--										
205	0	9.18	--	--	--	0.121	--	--										
209	4	0.11	--	0.038	--	--	--	--										

Table 13. Statistical summary of reported data for standard reference water sample N-65 (nutrient constituents)--Continued

Analyte: total P as P (total Phosphorus as phosphorus) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY						
N =	2	2	60	0. Other	MPV =	0.119
Minimum =	0.127	0.110	0.037	4. ICP	F-pseudosigma =	0.011
Maximum =	0.196	0.115	0.530	22m. Color:phosphomolybdate	N =	64
Median =			0.119		Hu =	0.125
F-pseudosigma =			0.010		HI =	0.111

Lab	Rating	Z-value	0	4	22m	Lab	Rating	Z-value	0	4	22m
1	4	0.42	--	--	0.123	243	4	0.14	--	--	0.120
3	0	38.28	--	--	0.530	247	3	-0.51	--	--	0.113
10	4	0.42	--	--	0.123	253	0	4.61	--	--	0.168
11	1	-1.72	--	--	0.100	255	4	0.14	--	--	0.120
13	4	0.33	--	--	0.122	292	4	0.14	--	--	0.120
21	4	-0.05	--	--	0.118	297	1	-1.72	--	--	0.100
23	0	-5.44	--	--	0.060	305	0	12.23	--	--	0.250
25	3	-0.79	--	--	0.110	313	3	-0.88	--	--	0.109
31	4	-0.05	--	--	0.118	314	0	7.21	0.196	--	--
38	4	-0.23	--	--	0.116	316	4	-0.07	--	--	0.118
45	4	-0.33	--	--	0.115	318	4	-0.30	--	--	0.115
46	4	-0.14	--	--	0.117	319	3	0.79	0.127	--	--
51	3	-0.51	--	--	0.113	320	3	0.60	--	--	0.125
59	0	7.58	--	--	0.200	321	3	0.60	--	--	0.125
70	0	3.49	--	--	0.156						
72	0	-2.65	--	--	0.090						
81	1	-2.00	--	--	0.097						
85	0	2.37	--	--	0.144						
86	4	-0.33	--	0.115	--						
87	4	0.23	--	--	0.121						
89	4	0.14	--	--	0.120						
96	3	-0.88	--	--	0.109						
97	1	-1.72	--	--	0.100						
102	4	-0.05	--	--	0.118						
105	4	0.14	--	--	0.120						
113	4	-0.05	--	--	0.118						
114	2	1.07	--	--	0.130						
127	2	1.26	--	--	0.132						
129	3	-0.60	--	--	0.112						
134	4	0.14	--	--	0.120						
138	4	0.42	--	--	0.123						
140	0	-5.44	--	--	0.060						
142	4	0.14	--	--	0.120						
143	3	0.98	--	--	0.129						
145	3	-0.79	--	--	0.110						
146	3	-0.70	--	--	0.111						
154	4	0.05	--	--	0.119						
158	3	0.79	--	--	0.127						
180	3	0.51	--	--	0.124						
183	4	-0.14	--	--	0.117						
190	4	-0.14	--	--	0.117						
193	4	0.14	--	--	0.120						
198	3	-0.79	--	--	0.110						
203	2	1.07	--	--	0.130						
212	2	-1.16	--	--	0.106						
213	2	1.07	--	--	0.130						
220	0	-7.58	--	--	0.037						
224	3	-0.88	--	--	0.109						
227	3	-0.79	--	0.110	--						
234	1	2.00	--	--	0.140						


Table 13. Statistical summary of reported data for standard reference water sample N-65 (nutrient constituents)--Continued

Analyte : PO₄ as P (Orthophosphate as phosphorus)

Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY						
N =	2	8	57	0. Other	MPV =	0.112
Minimum =	0.110	0.090	0.050	7. Ion chromatography	F-pseudosigma =	0.007
Maximum =	0.178	0.290	0.600	22m. Color:phosphomolybdate	N =	67
Median =		0.120	0.111		Hu =	0.120
F-pseudosigma =		0.042	0.006		HI =	0.110

Lab	Rating	Z-value	0	7	22m	Lab	Rating	Z-value	0	7	22m
1	4	0.27	--	--	0.114	227	4	0.00	--	--	0.112
3	0	65.83	--	--	0.600	234	4	-0.27	--	0.110	--
10	2	1.35	--	--	0.122	247	0	-2.97	--	0.090	--
11	2	1.08	--	--	0.120	253	1	1.75	--	--	0.125
12	4	-0.27	--	--	0.110	255	4	-0.27	--	--	0.110
13	1	-2.02	--	0.097	--	287	0	10.52	--	0.190	--
21	4	0.00	--	--	0.112	292	4	0.27	--	--	0.114
23	0	-8.36	--	--	0.050	297	4	0.00	--	--	0.112
25	0	27.11	--	--	0.313	305	4	-0.13	--	--	0.111
31	4	0.00	--	--	0.112	313	0	-2.48	--	--	0.094
38	4	-0.27	--	--	0.110	314	0	8.90	0.178	--	--
42	1	1.75	--	0.125	--	316	4	-0.28	--	--	0.110
45	4	0.40	--	0.115	--	318	4	-0.13	--	--	0.111
46	4	0.27	--	--	0.114	319	4	-0.27	0.110	--	--
51	4	-0.13	--	--	0.111	320	4	-0.27	--	--	0.110
53	4	-0.27	--	--	0.110	321	2	1.35	--	--	0.122
59	4	-0.27	--	--	0.110	322	4	-0.40	--	--	0.109
70	3	-0.67	--	--	0.107						
72	2	1.08	--	--	0.120						
81	4	0.40	--	--	0.115						
83	4	0.27	--	--	0.114						
85	2	1.21	--	--	0.121						
87	3	-0.81	--	--	0.106						
89	4	-0.13	--	--	0.111						
96	2	-1.35	--	--	0.102						
97	3	-0.94	--	--	0.105						
102	4	-0.13	--	--	0.111						
105	4	0.27	--	--	0.114						
113	3	0.81	--	--	0.118						
127	0	2.43	--	--	0.130						
129	3	0.81	--	--	0.118						
134	4	-0.13	--	--	0.111						
138	4	-0.40	--	--	0.109						
140	0	-7.01	--	--	0.060						
142	4	-0.13	--	--	0.111						
143	4	0.13	--	--	0.113						
145	0	2.43	--	--	0.130						
146	0	3.51	--	--	0.138						
154	2	-1.21	--	--	0.103						
158	3	-0.54	--	--	0.108						
180	4	0.27	--	--	0.114						
183	2	-1.08	--	--	0.104						
190	4	-0.40	--	--	0.109						
191	0	2.43	--	0.130	--						
198	4	-0.27	--	--	0.110						
203	2	1.08	--	--	0.120						
212	0	24.01	--	0.290	--						
213	4	-0.27	--	--	0.110						
220	0	2.43	--	--	0.130						
224	4	0.00	--	--	0.112						

Table 14. *Statistical summary of reported data for standard reference sample N-66 (nutrient constituents)*


Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported		
4. ICP	=	inductively coupled plasma
5. DCP	=	direct coupled plasma
7. IC	=	ion chromatography
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode

<u>Abbreviations and figure symbols</u>		
	N =	number of analyses--(excluding less than values)
	MPV =	most probable value ---
F-pseudostigma =		nonparametric statistic deviation
Uh =		upper hinge value
Lh =		lower hinge value
Uwl =		upper warning limit - - - -
Lwl =		lower warning limit - - - -
Ucl =		upper warning limit
Lcl =		lower warning limit - - - -
mg/L =		milligrams per liter
Lab =		laboratory code number
NR =		not rated, less than value reported or insufficient data
< =		less than
-- =		not reported


<u>Constituent</u>		<u>page</u>
NH ₃ as N	Ammonia as nitrogen	88
NH ₃ +Org N as N	Ammonia plus organic nitrogen as nitrogen	89
NO ₃ as N	Nitrate as nitrogen	90
Total P as P	Total Phosphorus as phosphorus	91
PO ₄ as P	Orthophosphate as phosphorus	92

Table 14. Statistical summary of reported data for standard reference water sample N-66 (nutrient constituents)--Continued

Analyte : NH₃ as N (Ammonia as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =		2	12	1	1	47	7	0. Other									MPV = 0.770
Minimum =		0.720	0.442	0.805	0.806	0.017	0.060	22. Colorimetric									F-pseudosigma = 0.048
Maximum =		0.750	1.010			1.000	1.000	22n. Color: Nesslerization									N = 70
Median =			0.785			0.766	0.794	22n. Color: Nesslerization									Uh = 0.810
F-pseudosigma =			0.109			0.048	0.082	22p. Color: phenate									Lh = 0.745
									40. Ion selective electrode								
Lab	Rating	Z-value	0	22	22N	22nitro	22p	40	Lab	Rating	Z-value	0	22	22N	22nitro	22p	40
1	4	-0.08	--	--	--	--	0.766	--	205	4	-0.15	--	--	--	--	0.763	--
3	0	-9.34	--	--	--	--	0.320	--	212	0	-6.81	--	0.442	--	--	--	--
11	4	0.42	--	--	--	--	0.790	--	213	NR	--	< 1	--	--	--	--	--
12	0	-2.91	--	--	--	--	0.630	--	220	4	0.00	--	0.770	--	--	--	--
13	0	4.77	--	--	--	--	1.000	--	224	4	0.39	--	--	--	--	0.789	--
19	2	-1.45	--	--	--	--	0.700	--	227	2	-1.04	0.720	--	--	--	--	--
23	1	-1.66	--	--	--	--	0.690	--	234	4	-0.21	--	--	--	--	--	0.760
25	1	-1.87	--	0.680	--	--	--	--	243	3	0.83	--	--	--	--	0.810	--
28	0	4.36	--	--	--	--	0.980	--	247	2	1.43	--	--	--	--	0.839	--
38	3	0.98	--	--	--	--	0.817	--	253	2	1.02	--	0.819	--	--	--	--
39	3	0.81	--	--	--	--	0.809	--	255	3	0.62	--	0.800	--	--	--	--
46	4	-0.33	--	--	--	--	0.754	--	287	0	4.17	--	--	--	--	--	0.971
50	3	0.75	--	--	--	0.806	--	--	292	3	0.62	--	--	--	--	--	0.800
59	3	-0.62	--	--	--	--	0.740	--	297	3	0.60	--	--	--	--	0.799	--
70	0	-2.57	--	--	--	--	0.646	--	305	4	0.50	--	--	--	--	--	0.794
72	4	-0.21	--	--	--	--	0.760	--	307	0	-14.74	--	--	--	--	--	0.060
81	1	1.76	--	--	--	--	0.855	--	308	2	-1.25	--	--	--	--	0.710	--
83	2	-1.04	--	--	--	--	0.720	--	313	4	-0.02	--	--	--	--	0.769	--
84	0	4.77	--	--	--	--	--	1.000	316	3	0.82	--	--	--	--	0.810	--
85	1	-1.76	--	--	--	--	0.685	--	319	4	-0.42	0.750	--	--	--	--	--
86	3	-0.56	--	--	--	--	0.743	--	320	0	3.11	--	0.920	--	--	--	--
87	1	1.87	--	--	--	--	0.860	--									
89	3	-0.52	--	0.745	--	--	--	--									
90	4	-0.10	--	--	--	--	0.765	--									
93	3	0.83	--	--	--	--	0.810	--									
96	4	-0.44	--	--	--	--	0.749	--									
97	3	0.87	--	0.812	--	--	--	--									
102	4	0.21	--	--	--	--	0.780	--									
105	2	-1.45	--	0.700	--	--	--	--									
107	3	0.52	--	--	--	--	0.795	--									
113	4	-0.15	--	--	--	--	0.763	--									
114	4	0.42	--	--	--	--	--	0.790									
127	2	-1.25	--	--	--	--	0.710	--									
129	3	0.73	--	--	0.805	--	--	--									
134	1	1.91	--	--	--	--	0.862	--									
138	4	-0.23	--	--	--	--	0.759	--									
140	0	4.98	--	1.010	--	--	--	--									
142	0	3.98	--	0.962	--	--	--	--									
143	4	0.00	--	--	--	--	0.770	--									
145	4	-0.42	--	--	--	--	0.750	--									
146	1	1.60	--	--	--	--	0.847	--									
154	4	-0.42	--	--	--	--	0.750	--									
158	4	-0.29	--	--	--	--	0.756	--									
180	4	0.25	--	--	--	--	0.782	--									
183	4	0.00	--	0.770	--	--	--	--									
190	0	-15.63	--	--	--	--	0.017	--									
193	4	-0.46	--	--	--	--	0.748	--									
198	4	0.37	--	--	--	--	0.788	--									
203	2	1.25	--	--	--	--	0.830	--									
204	2	1.12	--	--	--	--	0.824	--									

Table 14. Statistical summary of reported data for standard reference water sample N-66 (nutrient constituents)--Continued

Analyte : NH₃ + Organic N as N (Ammonia + organic nitrogen as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY									
N =	4	3	19	2	22	0. Other	MPV = 0.925		
Minimum =	0.858	1.000	0.726	0.900	0.564	20. Titrate: colorimetric	F-pseudostigma = 0.105		
Maximum =	1.360	1.800	4.000	1.020	1.170	22. Colorimetric	N = 50		
Median =			0.917	0.911		22n. Color: Nesslerization	Uh = 1.000		
F-pseudostigma =			0.078	0.085		22p. Color: phenate	Lh = 0.858		

Lab	Rating	Z-value	0	20	22	22N	22p	Lab	Rating	Z-value	0	20	22	22N	22p
1	4	-0.02	--	--	--	--	0.923	320	0	9.49	--	--	1.924	--	--
3	NR	--	--	< 1	--	--	--								
11	3	0.71	--	--	1.000	--	--								
12	0	-3.09	--	--	--	--	0.600								
23	4	0.33	--	--	--	--	0.960								
25	0	4.13	1.360	--	--	--	--								
38	4	-0.24	--	--	--	0.900	--								
39	1	1.85	--	1.120	--	--	--								
46	4	0.34	--	--	0.961	--	--								
50	3	0.90	--	--	1.020	--	--								
59	3	0.71	--	--	1.000	--	--								
70	3	-0.94	--	--	--	--	0.826								
72	4	0.26	--	--	--	--	0.952								
81	0	-2.19	--	--	--	--	0.695								
87	3	0.62	--	--	--	--	0.990								
89	4	0.24	--	--	--	--	0.950								
90	3	-0.74	--	--	--	--	0.847								
96	4	-0.25	--	--	--	--	0.899								
97	4	0.43	--	--	0.970	--	--								
102	3	-0.71	--	--	--	--	0.850								
105	0	29.21	--	--	4.000	--	--								
113	2	1.15	--	--	--	--	1.046								
127	3	-0.86	--	--	--	--	0.834								
129	3	0.90	--	--	--	1.020	--								
134	3	-0.75	--	--	--	--	0.846								
138	4	0.25	--	--	--	--	0.951								
140	3	1.00	--	--	1.030	--	--								
142	4	-0.44	--	--	0.879	--	--								
143	4	0.43	0.970	--	--	--	--								
145	3	-0.62	--	--	--	--	0.860								
146	4	0.02	--	--	--	--	0.927								
154	3	0.81	--	--	--	--	1.010								
158	3	-0.65	--	--	--	--	0.857								
180	4	0.47	--	--	--	--	0.974								
193	3	-0.95	--	--	0.825	--	--								
203	4	-0.14	--	--	0.910	--	--								
204	4	-0.03	--	--	0.922	--	--								
209	4	-0.23	--	--	0.901	--	--								
212	0	-3.43	--	--	--	--	0.564								
213	0	8.31	--	1.800	--	--	--								
220	4	-0.08	--	--	0.917	--	--								
224	0	2.33	--	--	--	--	1.170								
227	3	-0.64	0.858	--	--	--	--								
247	3	-0.55	--	--	0.867	--	--								
253	4	-0.34	--	--	0.889	--	--								
297	1	-1.89	--	--	0.726	--	--								
305	3	0.71	--	1.000	--	--	--								
313	1	-1.53	--	--	0.764	--	--								
316	4	-0.40	--	--	0.883	--	--								
319	0	2.71	1.210	--	--	--	--								

Table 14. Statistical summary of reported data for standard reference water sample N-66 (nutrient constituents)--Continued
 Analyte: NO₃ as N (Nitrate as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
Lab	Rating	Z-value	0	7	22	22cd	22h	22sulf	Lab	Rating	Z-value	0	7	22	22cd	22h	22sulf
			N = 4 20 10 35 1 4 0. Other				MPV = 0.931										
			Minimum = 0.870 0.290 0.900 0.182 0.940 0.910 7. Ion chromatography				F-pseudosigma = 0.033										
			Maximum = 1.100 1.160 0.983 1.070 0.970 22. Colorimetric				Rating criterion = 0.047										
			Median = 0.930 0.935 0.930 22cd. Cd diazotization				N = 74										
			F-pseudosigma = 0.036 0.047 0.036 22h. Color: hydrazine diazotization				Uh = 0.961										
							22sulf. Color: sulfanilamide				Lh = 0.916						
1	4	0.41	--	--	--	0.950	--	--	193	4	0.11	--	--	0.936	--	--	--
3	0	-16.09	--	--	--	0.182	--	--	198	4	0.06	--	--	0.934	--	--	--
11	4	-0.45	--	--	--	--	0.910	--	203	4	-0.24	--	--	--	0.920	--	--
12	0	-15.92	--	--	--	0.190	--	--	204	3	0.75	--	--	--	0.966	--	--
13	3	0.84	--	0.970	--	--	--	--	205	3	0.95	--	--	--	0.975	--	--
19	3	0.84	--	--	--	0.970	--	--	208	4	-0.02	--	0.930	--	--	--	--
23	0	-15.70	--	--	--	0.200	--	--	209	4	0.49	--	0.954	--	--	--	--
25	4	-0.09	--	0.927	--	--	--	--	212	0	-13.77	--	0.290	--	--	--	--
28	3	0.62	--	--	--	--	0.960	--	220	3	0.86	--	--	0.971	--	--	--
38	4	-0.15	--	--	--	0.924	--	--	224	2	1.33	--	0.993	--	--	--	--
39	4	0.02	--	0.932	--	--	--	--	227	3	-0.67	--	0.900	--	--	--	--
42	4	-0.28	--	0.918	--	--	--	--	234	4	-0.21	--	0.921	--	--	--	--
45	3	0.67	--	0.962	--	--	--	--	243	4	-0.02	--	--	--	0.930	--	--
46	4	0.37	--	--	--	0.948	--	--	247	0	-2.81	--	0.800	--	--	--	--
50	4	0.30	--	--	--	--	0.945	--	253	3	-0.54	--	--	0.906	--	--	--
53	2	-1.31	0.870	--	--	--	--	--	255	4	-0.24	--	--	--	0.920	--	--
59	2	1.48	--	--	--	1.000	--	--	292	4	-0.24	--	0.920	--	--	--	--
70	3	-0.67	--	--	--	0.900	--	--	297	3	-0.64	--	--	--	0.901	--	--
72	2	-1.10	--	--	--	0.880	--	--	305	3	0.62	--	--	0.960	--	--	--
81	4	0.19	--	--	--	--	0.940	--	307	0	-19.70	--	--	--	< 0.014	--	--
83	4	-0.02	--	--	--	0.930	--	--	308	4	0.19	--	--	--	0.940	--	--
84	0	-5.09	--	0.694	--	--	--	--	313	4	-0.02	--	--	--	0.930	--	--
85	4	0.49	--	--	--	0.954	--	--	316	3	0.51	--	--	--	0.955	--	--
86	2	-1.31	--	--	--	0.870	--	--	319	4	0.19	0.940	--	--	--	--	--
87	0	-2.58	--	--	--	0.811	--	--	320	3	0.52	--	--	--	0.955	--	--
89	4	-0.32	--	--	--	0.916	--	--									
90	2	1.01	--	--	--	0.978	--	--									
96	4	-0.24	--	--	--	0.920	--	--									
97	2	1.12	--	--	0.983	--	--	--									
102	0	4.92	--	1.160	--	--	--	--									
105	3	-0.67	--	--	0.900	--	--	--									
107	4	-0.15	--	--	--	0.924	--	--									
113	3	0.64	--	--	--	0.961	--	--									
114	0	2.99	--	--	--	1.070	--	--									
126	4	0.19	0.940	--	--	--	--	--									
127	4	-0.28	--	0.918	--	--	--	--									
129	3	0.92	--	0.974	--	--	--	--									
134	4	0.45	--	--	--	0.952	--	--									
138	4	-0.02	--	0.930	--	--	--	--									
140	2	1.03	--	--	0.979	--	--	--									
142	3	-0.52	--	--	0.907	--	--	--									
143	4	-0.13	--	--	--	0.925	--	--									
145	2	-1.10	--	--	--	0.880	--	--									
146	2	1.16	--	--	--	0.985	--	--									
154	3	0.84	--	--	--	--	0.970	--									
158	3	0.56	--	--	--	0.957	--	--									
180	1	1.70	--	1.010	--	--	--	--									
183	0	3.63	1.100	--	--	--	--	--									
190	4	-0.02	--	--	0.930	--	--	--									
191	4	-0.02	--	0.930	--	--	--	--									

Table 14. Statistical summary of reported data for standard reference water sample N-66 (nutrient constituents)--Continued

Analyte: total P as P (total Phosphorus as phosphorus) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY													
				N = 1 2 4 57				0. Other				MPV = 0.856	
Minimum = 0.903 0.830 0.780 0.069				4. ICP				F-pseudosigma = 0.040				Rating criterion = 0.043	
Maximum = 0.000 0.897 0.870 1.190				22. Colorimetric				22m. Color:phosphomolybdate				N = 64	
Median = 0.856												Uh = 0.880	
F-pseudosigma = 0.040												Lh = 0.825	
Lab	Rating	Z-value	0	4	22	22m	Lab	Rating	Z-value	0	4	22	22m
1	4	0.40	--	--	--	0.872	234	3	-0.89	--	--	--	0.820
3	0	-6.56	--	--	--	0.591	243	0	-2.62	--	--	--	0.750
11	4	0.35	--	--	--	0.870	247	2	-1.04	--	--	--	0.814
12	0	-6.83	--	--	--	0.580	253	4	0.15	--	--	--	0.862
13	3	0.94	--	--	--	0.894	255	4	0.35	--	--	--	0.870
23	0	-2.13	--	--	--	0.770	292	4	0.10	--	--	--	0.860
25	1	-1.88	--	--	0.780	--	297	4	-0.42	--	--	--	0.839
38	2	1.14	--	--	--	0.902	305	4	0.02	--	--	--	0.857
39	3	-0.59	--	--	--	0.832	307	0	-19.48	--	--	--	0.069
45	2	1.44	--	--	--	0.914	308	4	0.35	--	--	--	0.870
46	4	0.00	--	--	--	0.856	313	4	0.02	--	--	--	0.857
50	4	-0.30	--	--	--	0.844	316	3	0.85	--	--	--	0.890
59	0	8.27	--	--	--	1.190	319	2	1.16	0.903	--	--	--
70	4	-0.25	--	--	--	0.846	320	4	0.17	--	--	0.863	--
72	3	0.84	--	--	--	0.890							
81	3	-0.57	--	--	--	0.833							
85	0	4.55	--	--	--	1.040							
86	2	1.01	--	0.897	--	--							
87	3	-0.89	--	--	--	0.820							
89	4	-0.47	--	--	--	0.837							
96	4	0.50	--	--	--	0.876							
97	4	0.35	--	--	0.87	--							
102	4	0.10	--	--	--	0.860							
105	3	-0.64	--	--	0.830	--							
107	4	-0.12	--	--	--	0.851							
113	4	0.00	--	--	--	0.856							
114	4	-0.40	--	--	--	0.840							
127	4	-0.50	--	--	--	0.836							
129	3	0.57	--	--	--	0.879							
134	3	0.67	--	--	--	0.883							
138	3	-0.59	--	--	--	0.832							
140	3	-0.89	--	--	--	0.820							
142	3	-0.99	--	--	--	0.816							
143	3	0.64	--	--	--	0.882							
145	0	2.08	--	--	--	0.940							
146	3	0.89	--	--	--	0.892							
154	4	-0.30	--	--	--	0.844							
158	0	2.50	--	--	--	0.957							
180	3	0.59	--	--	--	0.880							
183	3	-0.94	--	--	--	0.818							
190	4	0.00	--	--	--	0.856							
193	3	-0.92	--	--	--	0.819							
198	4	-0.30	--	--	--	0.844							
203	4	0.35	--	--	--	0.870							
204	1	-1.86	--	--	--	0.781							
212	0	-7.20	--	--	--	0.565							
213	4	-0.15	--	--	--	0.850							
220	0	-2.23	--	--	--	0.766							
224	3	0.64	--	--	--	0.882							
227	3	-0.64	--	0.830	--	--							


Table 14. Statistical summary of reported data for standard reference water sample N-66 (nutrient constituents)--Continued

Analyte : PO₄ as P (Orthophosphate as phosphorus)

Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY					
N =	3	9	53	0. Other	MPV = 0.811
Minimum =	0.786	0.590	0.540	7. Ion chromatography	F-pseudosigma = 0.030
Maximum =	0.877	0.920	0.942	22m. Color:phosphomolybdate	Rating criterion = 0.041
Median =		0.800	0.811		N = 65
F-pseudosigma =		0.061	0.027		Uh = 0.836
					Lh = 0.795

Lab	Rating	Z-value	0	7	22m	Lab	Rating	Z-value	0	7	22m
1	3	-0.59	--	--	0.787	224	3	0.57	--	--	0.834
3	0	-5.47	--	--	0.589	227	3	0.64	--	--	0.837
11	4	-0.27	--	--	0.800	234	1	-1.55	--	0.748	--
12	0	-6.04	--	--	0.566	247	2	-1.01	--	0.770	--
13	4	0.37	--	0.826	--	253	4	0.20	--	--	0.819
23	0	-6.44	--	--	0.550	255	4	-0.02	--	--	0.810
25	1	1.63	0.877	--	--	287	0	-5.45	--	0.590	--
38	4	0.37	--	--	0.826	292	4	0.35	--	--	0.825
39	0	3.23	--	--	0.942	297	4	-0.20	--	--	0.803
42	2	1.23	--	0.861	--	305	3	0.52	--	--	0.832
45	4	0.47	--	0.830	--	308	2	-1.50	--	--	0.750
46	3	0.81	--	--	0.844	313	3	0.59	--	--	0.835
50	4	-0.37	--	--	0.796	316	3	0.62	--	--	0.836
53	1	1.82	--	--	0.885	319	4	0.22	0.820	--	--
59	4	0.22	--	--	0.820	320	3	-0.62	0.786	--	--
70	2	-1.09	--	--	0.767						
72	4	-0.27	--	--	0.800						
81	4	0.10	--	--	0.815						
83	3	0.74	--	--	0.841						
85	1	1.87	--	--	0.887						
87	4	-0.10	--	--	0.807						
89	4	-0.22	--	--	0.802						
96	4	-0.39	--	--	0.795						
97	3	-0.52	--	--	0.790						
102	0	-6.68	--	--	0.540						
105	4	-0.02	--	--	0.810						
107	4	0.00	--	--	0.811						
113	4	0.47	--	--	0.830						
127	2	1.26	--	--	0.862						
129	4	-0.07	--	--	0.808						
134	3	0.72	--	--	0.840						
138	2	-1.16	--	--	0.764						
140	3	-0.52	--	--	0.790						
142	4	0.35	--	--	0.825						
143	2	1.26	--	--	0.862						
145	2	1.21	--	--	0.860						
146	3	0.86	--	--	0.846						
154	4	-0.15	--	--	0.805						
158	4	0.42	--	--	0.828						
180	4	-0.15	--	--	0.805						
183	3	-0.57	--	--	0.788						
190	4	-0.27	--	--	0.800						
191	0	2.69	--	0.920	--						
198	4	-0.07	--	--	0.808						
203	4	0.47	--	--	0.830						
204	2	1.04	--	--	0.853						
208	4	-0.27	--	0.800	--						
212	0	-4.96	--	0.610	--						
213	4	-0.02	--	--	0.810						
220	2	1.18	--	--	0.859						

Table 15. *Statistical summary of reported data for standard reference sample P-34 (low ionic strength constituents)*

Definition of analytical methods, abbreviations, and symbols

Analytical methods

0. Other/Not reported	=	
1. AA: direct, air	=	atomic absorption: direct, air
2. AA: direct, N ₂ O	=	atomic absorption: direct, nitrous oxide
3. AA: graphite furnace	=	atomic absorption: graphite furnace
4. ICP	=	inductively coupled plasma
5. DCP	=	direct current plasma
6. ICP/MS	=	inductively coupled plasma / mass spectrometry
7. IC	=	ion chromatography
12. Flame emission	=	flame emission
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color:	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode
41. Electro	=	electrometric: (type meter specified)
50. Gravimetric	=	gravimetric: (precipitate specified)
51. Turbidimetric	=	turbidimetric: (precipitate specified)


Abbreviations and figure symbols

N	=	number of analyses--(excluding less than values)
MPV	=	most probable value ---
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit - - - -
Lwl	=	lower warning limit - - - -
Ucl	=	upper warning limit
Lcl	=	lower warning limit - - - -
mg/L	=	milligrams per liter
μS/cm	=	microsiemens per centimeter at 25 ^o C
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported

<u>Constituent</u>	<u>page</u>
Acid Acidity as CaCO ₃	94
Ca Calcium	95
Cl Chloride	96
F Fluoride	97
K Potassium	98
Mg Magnesium	99
Na Sodium	100
pH	101
PO ₄ as P Orthophosphate as Phosphorus	102
SO ₄ Sulfate	103
Sp Cond Specific Conductance	104

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: Acidity (as CaCO₃) Concentration Unit : mg/L



— 0 — 21

SUMMARY

N = 1 8
 Minimum = 3.2 1.8
 Maximum = 10.8
 Median = 5.8
 F-pseudosigma = 4.3


0. Other
 21. Titrate: electrometric

MPV = 4.6
 F-pseudosigma = 3.1
 N = 9
 Uh = 7.42
 Lh = 3.24

Lab	Rating	Z-value	0	21
3	1	1.74	--	10.0
25	NR	--	--	< 8
81	4	-0.40	--	3.4
89	4	0.00	--	4.6
105	1	2.00	--	10.8
127	3	0.80	--	7.1
146	3	0.91	--	7.4
220	3	-0.92	--	1.8
237	3	-0.68	--	2.5
273	4	-0.44	3.2	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: Ca (Calcium) Concentration Unit : mg/L




SUMMARY									
N =	4	4	30	1	4	1	0. Other	MPV =	1.63
Minimum =	0.65	1.59	0.90	1.65	1.43	1.62	1. AA: direct, air	F-pseudsigma =	0.07
Maximum =	2.40	1.81	1.80		1.74		4. ICP	Rating Criterion =	0.08
Median =			1.62				5. DCP	N =	44
F-pseudsigma =			0.08				6. ICP/MS	Uh =	1.67
							12. Flame emission	Lh =	1.57


Lab	Rating	Z-value	0	1	4	5	6	12
1	4	-0.12	--	--	1.62	--	--	--
2	1	1.73	1.77	--	--	--	--	--
3	4	-0.25	--	--	1.61	--	--	--
5	2	-1.23	--	--	1.53	--	--	--
11	4	-0.37	--	--	1.60	--	--	--
23	0	9.45	2.40	--	--	--	--	--
25	0	-4.29	--	--	1.28	--	--	--
38	3	0.61	1.68	--	--	--	--	--
46	4	0.37	--	--	1.66	--	--	--
48	4	0.25	--	--	--	--	1.65	--
59	0	-12.02	0.65	--	--	--	--	--
81	4	0.00	--	--	1.63	--	--	--
83	4	0.12	--	--	1.64	--	--	--
85	4	-0.37	--	1.60	--	--	--	--
86	4	0.37	--	--	1.66	--	--	--
89	4	-0.12	--	--	--	--	--	1.62
93	3	-0.74	--	--	1.57	--	--	--
105	4	0.25	--	--	--	1.65	--	--
110	3	-0.67	--	--	1.58	--	--	--
113	0	-3.51	--	--	1.34	--	--	--
127	2	1.47	--	--	1.75	--	--	--
134	4	0.37	--	--	1.66	--	--	--
138	3	0.49	--	--	1.67	--	--	--
140	4	-0.49	--	1.59	--	--	--	--
145	4	0.00	--	--	1.63	--	--	--
146	4	0.37	--	--	1.66	--	--	--
180	3	-0.98	--	--	1.55	--	--	--
191	3	0.49	--	--	--	--	1.67	--
193	4	-0.49	--	--	1.59	--	--	--
203	1	-1.96	--	--	1.47	--	--	--
209	0	-3.44	--	--	1.35	--	--	--
220	3	-0.61	--	--	1.58	--	--	--
227	3	0.74	--	--	1.69	--	--	--
237	3	-0.98	--	--	1.55	--	--	--
247	0	2.09	--	--	1.80	--	--	--
255	3	0.86	--	--	1.70	--	--	--
265	4	0.25	--	--	1.65	--	--	--
268	0	2.21	--	1.81	--	--	--	--
270	0	-8.96	--	--	0.90	--	--	--
273	3	0.61	--	--	1.68	--	--	--
277	4	-0.37	--	--	1.60	--	--	--
279	4	0.00	--	1.63	--	--	--	--
296	2	1.35	--	--	--	--	1.74	--
332	0	-2.45	--	--	--	--	1.43	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: Cl (Chloride) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	31	1	3	4	1	0. Other	MPV =	4.18
Minimum =	4.15	1.32	4.05	3.96	3.51	5.16	7. Ion chromatography	F-pseudosigma =	0.19
Maximum =		4.66		4.20	4.17		21. Titrate: electrometric	Rating Criterion =	0.21
Median =		4.20					22. Colorimetric	N =	41
F-pseudosigma =		0.18					22fe: Color: ferricyanide	Uh =	4.25
							22h. Color: hydrazine diazotization	Lh =	4.00

Lab	Rating	Z-value	0	7	21	22	22fe	22hg
1	1	1.91	--	4.58	--	--	--	--
2	0	2.15	--	4.63	--	--	--	--
3	0	-3.21	--	--	--	--	3.51	--
5	4	0.00	--	4.18	--	--	--	--
11	3	-0.86	--	--	--	--	4.00	--
23	4	0.24	--	4.23	--	--	--	--
25	4	0.10	--	4.20	--	--	--	--
46	4	0.10	--	--	--	4.20	--	--
48	3	-0.86	--	--	--	4.00	--	--
59	0	-13.68	--	1.32	--	--	--	--
81	3	-0.62	--	--	4.05	--	--	--
85	4	0.10	--	4.20	--	--	--	--
86	4	-0.14	--	4.15	--	--	--	--
89	4	-0.29	--	4.12	--	--	--	--
93	3	0.81	--	4.35	--	--	--	--
105	3	-0.86	--	4.00	--	--	--	--
110	4	0.17	--	4.22	--	--	--	--
113	2	1.44	--	4.48	--	--	--	--
127	3	0.81	--	4.35	--	--	--	--
134	4	0.05	--	4.19	--	--	--	--
138	3	-0.91	--	3.99	--	--	--	--
140	2	-1.05	--	--	--	3.96	--	--
143	0	-2.97	--	--	--	--	3.56	--
145	4	0.05	--	4.19	--	--	--	--
146	4	-0.05	--	--	--	--	4.17	--
180	4	0.33	--	4.25	--	--	--	--
183	0	4.69	--	--	--	--	--	5.16
191	4	-0.43	--	4.09	--	--	--	--
203	2	-1.34	--	3.90	--	--	--	--
208	2	-1.05	--	3.96	--	--	--	--
209	3	0.67	--	4.32	--	--	--	--
220	0	2.32	--	4.66	--	--	--	--
227	4	0.21	--	4.22	--	--	--	--
247	0	-2.01	--	3.76	--	--	--	--
265	0	-2.01	--	3.76	--	--	--	--
268	2	-1.34	--	3.90	--	--	--	--
270	4	0.10	--	4.20	--	--	--	--
273	4	-0.14	4.15	--	--	--	--	--
277	3	0.62	--	4.31	--	--	--	--
301	4	0.37	--	4.26	--	--	--	--
322	4	-0.14	--	4.15	--	--	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: F (Fluoride) Concentration Unit : mg/L



SUMMARY

N =	1	15	1	13
Minimum =	0.160	0.018	0.300	0.100
Maximum =		0.212		0.214
Median =		0.164		0.160
F-pseudosigma =		0.029		0.010


- 0. Other
- 7. Ion chromatography
- 22. Colorimetric
- 40. Ion selective electrode

MPV =	0.161
F-pseudosigma =	0.022
N =	30
Uh =	0.180
Lh =	0.150

Lab	Rating	Z-value	0	7	22	40
1	3	-0.94	--	--	--	0.140
2	4	0.12	--	0.164	--	--
3	4	0.27	--	--	--	0.167
5	3	0.85	--	0.180	--	--
11	0	-2.74	--	--	--	0.100
23	0	-6.43	--	0.018	--	--
25	4	0.40	--	0.170	--	--
59	4	-0.04	--	0.160	--	--
81	0	2.29	--	0.212	--	--
85	4	-0.04	--	--	--	0.160
89	4	-0.04	--	--	--	0.160
93	4	0.40	--	0.170	--	--
105	NR	--	--	< 0.2	--	--
113	4	0.04	--	--	--	0.162
127	3	-0.90	--	0.141	--	--
134	4	-0.04	--	--	--	0.160
138	0	2.38	--	--	--	0.214
140	0	2.20	--	--	--	0.210
145	3	-0.94	--	0.140	--	--
146	3	0.58	--	--	--	0.174
180	2	1.44	--	0.193	--	--
183	4	0.40	--	--	--	0.170
190	4	-0.22	--	--	--	0.156
191	4	-0.04	--	0.160	--	--
220	0	2.16	--	0.209	--	--
227	4	-0.04	0.160	--	--	--
247	4	-0.49	--	0.150	--	--
255	NR	--	--	--	--	< 0.5
265	0	-2.74	--	< 0.1	--	--
270	0	-2.74	--	0.100	--	--
273	4	-0.49	--	--	--	0.150
277	2	1.30	--	0.190	--	--
279	0	6.25	--	--	0.300	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: K (Potassium) Concentration Unit : mg/L



SUMMARY

N =	9	16	4	3	2
Minimum =	0.228	0.150	0.140	0.212	0.200
Maximum =	0.330	1.160	0.940	0.350	0.230
Median =	0.250	0.233			
F-pseudosigma =	0.030	0.042			


- 1. AA: direct, air
- 4. ICP
- 6. ICP/MS
- 7. Ion chromatography
- 12. Flame emission

MPV =	0.238
F-pseudosigma =	0.036
N =	34
Uh =	0.261
Lh =	0.212

Lab	Rating	Z-value	1	4	6	7	12
1	4	-0.15	0.232	--	--	--	--
2	3	-0.70	--	--	--	0.212	--
3	0	25.61	--	1.160	--	--	--
5	NR	--	--	< 1	--	--	--
11	0	-2.43	--	0.150	--	--	--
23	2	-1.04	--	--	--	--	0.200
25	2	-1.35	--	0.189	--	--	--
38	4	0.35	0.250	--	--	--	--
48	3	0.57	--	--	0.258	--	--
59	0	3.12	--	--	--	0.350	--
81	3	-0.99	--	0.202	--	--	--
85	1	1.90	0.306	--	--	--	--
86	3	0.65	--	0.261	--	--	--
89	4	-0.21	--	--	--	--	0.230
93	4	-0.21	--	0.230	--	--	--
105	NR	--	--	< 1	--	--	--
110	4	-0.21	0.230	--	--	--	--
113	1	-1.54	--	0.182	--	--	--
127	4	-0.26	0.228	--	--	--	--
134	4	0.07	0.240	--	--	--	--
138	4	-0.07	--	0.235	--	--	--
140	3	0.99	0.273	--	--	--	--
145	NR	--	--	< 0.71	--	--	--
146	2	-1.07	--	0.199	--	--	--
180	NR	--	--	< 0.62	--	--	--
191	4	-0.49	--	--	0.220	--	--
193	3	0.54	0.257	--	--	--	--
203	4	-0.49	--	0.220	--	--	--
209	0	17.83	--	0.880	--	--	--
237	4	0.15	--	--	--	0.243	--
247	1	1.57	--	0.294	--	--	--
265	4	0.07	--	0.240	--	--	--
270	4	0.35	--	0.250	--	--	--
273	4	0.07	--	0.240	--	--	--
277	2	-1.04	--	0.200	--	--	--
279	0	2.57	0.330	--	--	--	--
296	0	-2.71	--	--	0.140	--	--
332	0	19.50	--	--	0.940	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: Mg (Magnesium) Concentration Unit : mg/L




SUMMARY

N =	7	28	1	4	2	1. AA: direct, air	MPV = 0.592
Minimum =	0.570	0.510	0.623	0.500	0.624	4. ICP	F-pseudosigma = 0.037
Maximum =	0.661	0.662		0.660	1.010	5. DCP	N = 42
Median =	0.600	0.589				4. ICP	Uh = 0.620
F-pseudosigma =	0.032	0.029				5. DCP	Lh = 0.570


Lab	Rating	Z-value	1	4	5	6	7
1	4	-0.11	--	0.588	--	--	--
2	3	0.87	--	--	--	--	0.624
3	4	0.05	--	0.594	--	--	--
5	2	-1.08	--	0.552	--	--	--
11	2	-1.40	--	0.540	--	--	--
23	4	0.22	0.600	--	--	--	--
25	1	-1.97	--	0.519	--	--	--
38	3	0.89	0.625	--	--	--	--
46	4	0.00	--	0.592	--	--	--
48	1	1.83	--	--	--	0.660	--
59	0	11.28	--	--	--	--	1.010
81	3	0.67	--	0.617	--	--	--
83	4	-0.32	--	--	--	--	--
85	1	1.86	0.661	--	--	--	--
86	3	0.84	--	--	0.623	--	--
89	3	0.76	0.620	--	--	--	--
93	3	-0.59	--	0.570	--	--	--
105	4	0.49	--	0.610	--	--	--
110	4	-0.41	--	0.577	--	--	--
113	1	-1.62	--	0.532	--	--	--
127	1	1.89	--	0.662	--	--	--
134	4	-0.27	--	0.582	--	--	--
138	4	0.00	--	0.592	--	--	--
140	3	-0.59	0.570	--	--	--	--
145	4	-0.05	--	0.590	--	--	--
146	4	0.19	--	0.599	--	--	--
180	2	-1.46	--	0.538	--	--	--
191	2	1.03	--	--	--	0.630	--
193	4	-0.38	--	0.578	--	--	--
203	0	-2.21	--	0.510	--	--	--
209	1	-1.67	--	0.530	--	--	--
220	3	-0.59	--	0.570	--	--	--
227	4	0.22	--	0.600	--	--	--
247	4	0.08	--	0.595	--	--	--
255	3	0.92	--	0.626	--	--	--
265	4	0.22	--	0.600	--	--	--
268	4	-0.32	0.580	--	--	--	--
273	3	0.76	--	0.620	--	--	--
277	4	0.49	--	0.610	--	--	--
279	4	-0.32	0.580	--	--	--	--
296	0	-2.48	--	--	--	0.500	--
332	4	-0.05	--	--	--	0.590	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: Na (Sodium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	8	25	4	3	1	1. AA: direct, air	MPV =	0.450	
Minimum =	0.385	0.370	0.440	0.455	0.470	4. ICP	F-pseudostandard deviation =	0.037	
Maximum =	0.477	1.400	0.700	0.480		6. ICP/MS	N =	41	
Median =	0.432	0.450				7. Ion chromatography	Uh =	0.480	
F-pseudostandard deviation =	0.019	0.039				12. Flame emission	Lh =	0.430	

Lab	Rating	Z-value	1	4	6	7	12
1	4	-0.24	--	0.441	--	--	--
2	3	0.64	--	--	--	0.474	--
3	0	8.09	--	0.750	--	--	--
5	3	-0.89	--	0.417	--	--	--
11	3	-0.54	--	0.430	--	--	--
23	3	-0.81	0.420	--	--	--	--
25	4	-0.49	--	0.432	--	--	--
38	4	-0.27	0.440	--	--	--	--
46	4	0.05	--	0.452	--	--	--
48	3	0.89	--	--	0.483	--	--
59	3	0.81	--	--	--	0.480	--
81	0	3.08	--	0.564	--	--	--
83	3	-0.54	--	0.430	--	--	--
85	3	0.73	0.477	--	--	--	--
86	3	-0.51	--	0.431	--	--	--
89	3	0.54	--	--	--	--	0.470
93	3	-0.81	--	0.420	--	--	--
105	4	0.00	--	0.450	--	--	--
110	4	-0.30	0.439	--	--	--	--
113	0	-6.75	--	< 0.2	--	--	--
127	4	0.08	--	0.453	--	--	--
134	1	-1.75	0.385	--	--	--	--
138	4	-0.22	--	0.442	--	--	--
140	3	-0.70	0.424	--	--	--	--
145	0	-2.16	--	0.370	--	--	--
146	4	0.40	--	0.465	--	--	--
180	3	0.89	--	0.483	--	--	--
191	4	-0.27	--	--	0.440	--	--
193	4	0.00	--	0.450	--	--	--
203	2	-1.35	--	0.400	--	--	--
209	2	-1.08	--	0.410	--	--	--
220	0	2.16	--	0.530	--	--	--
237	4	0.13	--	--	--	0.455	--
247	0	2.99	--	0.561	--	--	--
265	4	0.00	--	0.450	--	--	--
268	4	-0.27	0.440	--	--	--	--
270	0	25.63	--	1.400	--	--	--
273	2	1.35	--	0.500	--	--	--
277	2	-1.35	--	0.400	--	--	--
279	2	-1.08	0.410	--	--	--	--
296	0	6.75	--	--	0.700	--	--
332	1	1.89	--	--	0.520	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: pH



—○— 41

SUMMARY


N = 45
 Minimum = 3.31
 Maximum = 6.74
 Median = 4.32
 F-pseudosigma = 0.10

41. Direct reading

MPV = 4.32
 F-pseudosigma = 0.10
 Rating Criterion = 0.22
 N = 45
 Uh = 4.36
 Lh = 4.23

Lab	Rating	Z-value	41
1	0	11.20	6.74
2	4	0.17	4.36
3	3	-0.23	4.27
5	0	-4.68	3.31
11	4	0.00	4.32
12	1	-0.88	4.13
23	4	-0.05	4.31
25	3	0.42	4.41
38	3	0.37	4.40
46	4	0.09	4.34
48	0	-1.48	4.00
59	3	-0.42	4.23
81	3	-0.32	4.25
85	4	0.00	4.32
86	4	0.00	4.32
89	0	-1.16	4.07
93	3	0.37	4.40
105	1	0.83	4.50
107	3	0.28	4.38
110	4	0.00	4.32
113	1	0.83	4.50
127	4	0.19	4.36
134	4	0.14	4.35
138	4	0.00	4.32
140	0	-1.30	4.04
143	4	0.19	4.36
145	3	-0.32	4.25
146	4	0.19	4.36
180	4	-0.09	4.30
183	3	0.37	4.40
190	3	-0.28	4.26
203	0	-3.38	3.59
204	4	0.14	4.35
209	0	-1.57	3.98
227	3	-0.37	4.24
237	4	-0.19	4.28
243	4	-0.14	4.29
244	4	0.00	4.32
247	3	0.28	4.38
268	0	-2.41	3.80
270	0	-1.20	4.06
273	4	-0.05	4.31
277	0	-1.02	4.10
279	0	-2.50	3.78
301	3	0.42	4.41

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu
 Analyte: PO₄ as P (Orthophospate as phosphorus) Concentration Unit : mg/L




N =	1	5	26	4. ICP	MPV = 0.108
Minimum =	0.107	0.095	0.070	7. Ion chromatography	F-pseudosigma = 0.011
Maximum =		0.220	0.150	22m. Color:phosphomolybdate	N = 32
Median =			0.108		Uh = 0.113
F-pseudosigma =			0.011		Lh = 0.098

Lab	Rating	Z-value	4	7	22m
3	0	3.74	--	--	0.150
5	3	-0.83	--	--	0.098
11	3	-0.66	--	--	0.100
23	1	-1.54	--	--	0.090
25	4	0.48	--	--	0.113
38	4	-0.04	--	--	0.107
46	4	0.22	--	--	0.110
48	3	0.57	--	--	0.114
59	2	-1.01	--	0.096	--
81	4	0.13	--	0.109	--
85	4	0.48	--	--	0.113
89	4	-0.04	--	--	0.107
105	3	-0.66	--	--	0.100
113	4	0.22	--	--	0.110
127	1	1.54	--	--	0.125
134	4	-0.04	--	--	0.107
138	3	-0.90	--	--	0.097
140	1	-1.54	--	--	0.090
143	3	0.57	--	--	0.114
145	4	0.22	--	--	0.110
146	1	-1.98	--	--	0.085
180	4	0.22	--	--	0.110
183	4	0.22	--	--	0.110
190	4	0.04	--	--	0.108
191	0	9.89	--	0.220	--
203	4	-0.48	--	--	0.102
220	0	3.74	--	0.150	--
227	4	-0.04	0.107	--	--
247	0	-3.30	--	--	0.070
255	0	2.15	--	--	0.132
265	2	-1.10	--	0.095	--
273	0	-2.42	--	--	0.080

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: SO₄ (Sulfate) Concentration Unit : mg/L




SUMMARY

N =	1	1	24	1	0. Other	MPV = 0.437
Minimum =	0.37	0.37	0.078	1.00	4. ICP	F-pseudosigma = 0.141
Maximum =			14.700		7. Ion chromatography	N = 27
Median =			0.439		22mtb. Color: methyl thymol blue	Uh = 0.560
F-pseudosigma =			0.140		51. Turbidimetric	Lh = 0.370

Lab	Rating	Z-value	0	4	7	22mtb	51
1	3	-0.55	--	--	0.359	--	--
2	4	0.15	--	--	0.459	--	--
3	NR	--	--	--	--	< 10	--
5	0	3.36	--	--	0.910	--	--
11	NR	--	--	--	< 1	--	--
23	NR	--	--	--	< 1	--	--
25	NR	--	--	--	< 5	--	--
48	NR	--	--	--	--	--	< 1
59	0	101.27	--	--	14.700	--	--
81	NR	--	--	--	--	< 5	--
83	4	-0.48	--	0.370	--	--	--
85	2	-1.04	--	--	0.290	--	--
89	4	0.29	--	--	0.478	--	--
93	4	-0.26	--	--	0.400	--	--
105	NR	--	--	--	< 1	--	--
110	4	0.00	--	--	0.437	--	--
113	3	-0.83	--	--	0.320	--	--
127	4	0.18	--	--	0.462	--	--
134	4	0.09	--	--	0.450	--	--
138	4	-0.45	--	--	0.373	--	--
140	0	4.00	--	--	--	--	1.000
145	2	1.37	--	--	0.630	--	--
146	NR	--	--	--	--	--	< 5
180	4	-0.18	--	--	0.411	--	--
191	4	0.38	--	--	0.490	--	--
203	0	2.58	--	--	0.800	--	--
209	4	-0.05	--	--	0.430	--	--
220	1	1.54	--	--	0.654	--	--
227	NR	--	--	--	< 0.780	--	--
247	NR	--	--	--	< 1	--	--
255	NR	--	--	--	--	< 15	--
265	3	-0.62	--	--	0.350	--	--
268	4	-0.12	--	--	0.420	--	--
270	0	8.97	--	--	1.700	--	--
273	4	-0.48	0.370	--	--	--	--
277	4	-0.48	--	--	0.370	--	--
301	0	-2.55	--	--	0.078	--	--
322	4	0.02	--	--	0.440	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-34 (low ionic strength constituents)--Continu

Analyte: Sp Cond (Specific Conductance) Concentration Unit : mS/cm



SUMMARY

N = 43
 Minimum = 0.0
 Maximum = 51.0
 Median = 41.8
 F-pseudsigma = 1.7

41. Direct reading

MPV = 41.8
 F-pseudsigma = 1.7
 Rating Criterion = 2.1
 N = 43
 Uh = 42.5
 Lh = 40.2

Lab	Rating	Z-value	41
1	0	-5.06	31.2
3	3	0.59	43.0
5	4	0.30	42.4
11	4	-0.37	41.0
23	2	-1.28	39.1
25	0	4.42	51.0
38	4	0.49	42.8
46	3	0.59	43.0
48	4	0.35	42.5
59	3	0.64	43.1
81	4	-0.18	41.4
85	1	-1.81	38.0
86	4	0.21	42.2
89	4	0.25	42.3
93	3	-0.56	40.6
105	3	-0.61	40.5
107	4	0.25	42.3
110	2	-1.16	39.4
113	3	-0.61	40.5
127	3	0.73	43.3
134	4	0.00	41.8
138	3	-0.99	39.7
140	4	-0.37	41.0
143	4	-0.37	41.0
145	3	-0.90	39.9
146	4	0.11	42.0
180	0	2.03	46.0
183	0	2.89	47.8
190	4	-0.18	41.4
193	2	-1.42	38.8
203	4	0.06	41.9
204	4	0.25	42.3
220	0	-2.76	36.0
227	4	-0.46	40.8
243	4	-0.08	41.6
244	3	0.54	42.9
247	4	0.30	42.4
268	4	0.11	42.0
273	4	0.40	42.6
277	2	-1.33	39.0
279	0	-19.98	0.0
301	4	0.35	42.5
308	1	-1.52	38.6

Table 16. *Statistical summary of reported data for standard reference sample Hg-30 (mercury)*

Definition of analytical methods, abbreviations, and symbols

Analytical methods

- 0. Other/Not reported
 - 4. ICP = inductively coupled plasma
 - 6. ICP/MS = inductively coupled plasma / mass spectrometry
 - 8. AA: cold vapor = atomic absorption: cold vapor
 - 9. Atomic fluorescence
-


Abbreviations and figure symbols

- N = number of analyses--(excluding less than values)
 - MPV = most probable value ---
 - F-pseudostigma = nonparametric statistic deviation
 - Uh = upper hinge value
 - Lh = lower hinge value
 - Uwl = upper warning limit ---
 - Lwl = lower warning limit ---
 - Ucl = upper warning limit
 - Lcl = lower warning limit ---
 - µg/L = micrograms per liter
 - Lab = laboratory code number
 - NR = not rated, less than value reported or insufficient data
 - < = less than
 - = not reported
-


<u>Constituent</u>	<u>page</u>
Hg Mercury	106

Table 16. Statistical summary of reported data for standard reference water sample Hg-30 (mercury)--Continued

Analyte: Hg (Mercury) Concentration Unit : mg/L



ANALITICAL METHOD CODE



SUMMARY						
N =	1	3	37	4	0. Other	MPV = 1.92
Minimum =	1.95	1.91	1.07	1.34	6. ICP/MS	F-pseudosigma = 0.16
Maximum =		2.10	2.51	1.96	8. AA: cold vapor	N = 45
Median =			1.90		9. Atomic fluorescence	Uh = 2.01
F-pseudosigma =			0.21			Lh = 1.80

Lab	Rating	Z-value	0	6	8	9
1	3	0.51	--	--	2.00	--
3	4	-0.13	--	--	1.90	--
10	2	1.14	--	--	2.10	--
11	2	-1.08	--	--	1.75	--
12	2	1.14	--	--	2.10	--
13	4	-0.06	--	--	1.91	--
32	2	1.14	--	2.10	--	--
46	4	-0.32	--	--	1.87	--
47	4	0.19	--	--	--	1.95
50	4	-0.06	--	1.91	--	--
54	4	-0.13	--	--	1.90	--
59	4	0.06	--	--	1.93	--
70	2	1.33	--	--	2.13	--
81	3	-0.82	--	--	1.79	--
86	0	-3.67	--	--	--	1.34
87	3	0.51	--	--	2.00	--
89	2	-1.01	--	--	1.76	--
96	2	1.08	--	--	2.09	--
97	0	3.74	--	--	2.51	--
105	0	2.28	--	--	2.28	--
109	4	0.25	--	--	1.96	--
127	2	1.08	--	--	2.09	--
134	3	0.59	--	--	2.01	--
142	4	0.00	--	--	1.92	--
144	4	-0.32	--	--	1.87	--
145	0	-5.38	--	--	1.07	--
146	1	-1.58	--	--	1.67	--
147	4	0.19	--	--	--	1.95
149	4	-0.13	--	--	1.90	--
154	4	-0.13	--	--	1.90	--
193	3	-0.51	--	--	1.84	--
212	1	-1.58	--	--	1.67	--
213	0	-2.66	--	--	1.50	--
234	1	-1.71	--	--	1.65	--
245	4	-0.19	--	--	1.89	--
247	1	1.77	--	--	2.20	--
259	4	0.00	--	--	1.92	--
265	4	0.19	--	1.95	--	--
277	2	-1.20	--	--	1.73	--
284	3	0.95	--	--	2.07	--
292	1	-2.03	--	--	1.60	--
298	4	0.19	1.95	--	--	--
304	4	0.25	--	--	--	1.96
307	0	2.60	--	--	2.33	--
330	3	-0.76	--	--	1.80	--

Table 17. Most probable values for constituents and properties in standard reference samples distributed in March 2000
 [MPV, most probable value; N, number of samples; µg/L, microgram per liter; mg/L, milligram per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius.]

T-161 (trace constituents)

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Ag	17.3 µg/L	1.3	61	Mg	1.51 mg/L	0.07	73
Al	32.4 µg/L	6.3	48	Mn	37.4 µg/L	2.3	71
As	26.1 µg/L	1.3	63	Mo	18.9 µg/L	1.6	43
B	37.5 µg/L	4.9	33	Na	43.0 mg/L	1.5	73
Ba	70.4 µg/L	3.3	63	Ni	29.0 µg/L	2.0	65
Be	12.9 µg/L	0.5	56	Pb	16.5 µg/L	1.2	68
Ca	7.17 mg/L	0.28	74	Sb	14.0 µg/L	0.9	41
Cd	17.5 µg/L	0.9	71	Se	9.58 µg/L	1.11	54
Co	12.5 µg/L	0.6	48	SiO ₂	14.8 mg/L	1.0	45
Cr	34.6 µg/L	1.7	71	Sr	54.2 µg/L	2.4	39
Cu	22.0 µg/L	1.7	70	Tl	50.3 µg/L	3.8	41
Fe	61.7 µg/L	10.2	70	U	7.97 µg/L	0.44	9
K	1.26 mg/L	0.11	63	V	18.4 µg/L	1.2	40
Li	17.8 µg/L	1.6	28	Zn	40.6 µg/L	2.4	71

M-154 (major constituents)

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Alkalinity as CaCO ₃	112 mg/L	4	75	Na	139 mg/L	6	74
B	284 µg/L	18	34	total P as P	1.31 mg/L	0.06	62
Ca	67.5 mg/L	3.1	80	pH	8.33 units	0.13	79
Cl	141 mg/L	4	77	SiO ₂	10.8 mg/L	0.8	53
DSRD	701 mg/L	14	53	SO ₄	164.0 mg/L	5.0	74
F	2.18 mg/L	0.13	60	Sp Cond	1124 µS/cm	23	75
K	9.08 mg/L	0.62	67	Sr	596 µg/L	29	35
Mg	15.1 mg/L	0.6	80	V	16.9 µg/L	1.8	33

N-65 (nutrient constituents)

Analyte	MPV	F-pseudosigma	N
NH ₃ as N	0.124 mg/L	0.023	69
NH ₃ +OrgN as N	0.159 mg/L	0.067	46
NO ₃ as N	0.037 mg/L	0.009	63
total P as P	0.119 mg/L	0.011	64
PO ₄ as P	0.112 mg/L	0.007	67

N-66 (nutrient constituents)

Analyte	MPV	F-pseudosigma	N
NH ₃ as N	0.770 mg/L	0.048	70
NH ₃ +OrgN as N	0.925 mg/L	0.105	50
NO ₃ as N	0.931 mg/L	0.033	74
total P as P	0.856 mg/L	0.040	64
PO ₄ as P	0.811 mg/L	0.030	65

P-34 (low ionic strength constituents)

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Acidity	4.6 mg/L	3.1	9	Na	0.450 mg/L	0.037	41
Ca	1.63 mg/L	0.07	44	pH	4.32 units	0.10	45
Cl	4.18 mg/L	0.19	41	PO ₄ as P	0.108 mg/L	0.011	32
F	0.161 mg/L	0.022	30	SO ₄	0.44 mg/L	0.14	27
K	0.238 mg/L	0.036	34	Sp Cond	41.8 µS/cm	1.7	43
Mg	0.592 mg/L	0.037	42				

Hg-30 (mercury)

Analyte	MPV	F-pseudosigma	N
Hg	1.92 µg/L	0.16	45